

SAFEGUARDS MEASUREMENT EVALUATION PROGRAM

URANIUM AND PLUTONIUM
SAMPLES EXCHANGE ANNUAL REPORT



JANUARY-DECEMBER 2007

B. Srinivasan, Kattathu Mathew, Joseph Waggoner,
Usha Narayanan and Jon Neuhoff



U.S. DEPARTMENT OF ENERGY

SAFEGUARDS MEASUREMENT EVALUATION PROGRAM

**URANIUM AND PLUTONIUM
SAMPLES EXCHANGE ANNUAL REPORT**
January 2007 through December 2007

B. Srinivasan, Kattathu Mathew, Joseph Waggoner,
Usha Narayanan and Jon Neuhoff

NEW BRUNSWICK LABORATORY
9800 South Cass Avenue
Argonne, Illinois 60439

JULY 2008

NBL Home Page: www.nbl.doe.gov
NBL Internet Address: USDOE.NBL@CH.DOE.GOV
NBL Telephone Number: 630-252-2442
NBL Facsimile Number: 630-252-6256

[This page intentionally left blank.]

TABLE OF CONTENTS

NBL: HISTORY AND MISSION	iii
ACKNOWLEDGEMENTS	v
SUMMARY	vii
A. INTRODUCTION	1
B. SAFEGUARDS MEASUREMENT EVALUATION PROGRAM	1
C. CY 2007 SAFEGUARDS MEASUREMENT EVALUATION PROGRAM	2
C.1. SME Program Participants	6
C.2. Materials and Measurement Methods	6
C.3. Test Materials, Shipping, and Analysis	8
C.4. Database and Statistical Evaluation Programs	8
C.5. Statistical Evaluation of Measurement Results	8
C.6. Examples of Statistical Evaluation Reports	10
C.7. SMES Design Features and Availability to Users	16
D. ANALYSES RESULTS AND REPORTING FORMAT	17
E. PERFORMANCE EVALUATION: MATERIAL BY MATERIAL	19
E.1. Uranyl Nitrate Solutions	19
E.2. Enriched Uranium Dioxide (UO_2) Pellet	24
E.3. Uranium Hexafluoride (UF_6)	28
E.4. Uranium Oxide (UO_3) Powder	31
E.5. ^{235}U Enrichment	35
E.6. Plutonium Assay and Isotopic Abundance	51
F. SPECIAL STUDY: MOISTURE IN UO_3 TEST SAMPLES	52
G. LONG TERM EVALUATION OF URANIUM MEASUREMENTS	53
APPENDICES	111
Appendix A: Uranium Assay Results.....	113
Appendix B: Uranium Isotopic Results	143
Appendix C: NBL Measurement Evaluation Program: New Database Development	159
Appendix D: New Brunswick Laboratory Safeguards Measurement Evaluation Program: Operational Features	165

[This page intentionally left blank.]

NBL: HISTORY AND MISSION

The New Brunswick Laboratory (NBL) is owned and operated by the United States Department of Energy through the Office of Science (SC). NBL was established in 1949 as an analytical chemistry laboratory in New Brunswick, New Jersey to provide support to the United States Atomic Energy Commission. At that time, it was staffed by scientists from the National Bureau of Standards who had contributed significantly to nuclear material measurement programs in the Manhattan Project. At NBL, they provided the technical expertise and skills to solve problems related to quantitative analyses of uranium-bearing materials. Over the years, these scientists and others following them have expanded the capabilities of the laboratory to include chemical and mass spectrometric analyses of plutonium and other trans-uranium elements, research and development activities in chemical analyses techniques, preparation of certified reference materials, and operation of the nuclear safeguards measurement evaluation program. In 1977, the laboratory moved from New Jersey to its present location at the Argonne National Laboratory site in Illinois.

NBL's major mission is to provide technical assistance to the Department of Energy in the following areas: measurement evaluation program operation, certified (nuclear) reference materials preparation, measurement techniques development, and measurement services to domestic and international customers. In addition to fulfilling these tasks, the laboratory helps the Department in three other areas: conducting technical audits, resolving shipper/receiver differences in material transfers, and assisting in nuclear nonproliferation programs.

[This page intentionally left blank.]

ACKNOWLEDGEMENTS

The Safeguards Measurement Evaluation program is administered by the United States Department of Energy, Office of Science (SC). The authors of this annual report thank the following individuals for their contributions to the CY 2007 program: Richard Palczynski (DOE-CH), Ron Kuziel, Kevin Atto, Joel Catausan, Miguelito Domingo, Mark Jilek, Stephen Smith (all from Chickasaw Nation Industries), William Guthrie and Thomas Vetter (both from NIST), and Kimberly Johnson-Miller, Peter Mason, Frank Orlowicz, Glennda Orlowicz, Michael Soriano, Gary Sowell, Alma Stiffin, Albert Thomas, Stephan Vogt, Heidi Williams (all from NBL).

[This page intentionally left blank.]

SUMMARY

The New Brunswick Laboratory (NBL) has been tasked by the United States Department of Energy, Office of Science (SC) to evaluate the quality of measurement results in nuclear materials accounting at the Department of Energy laboratories. Both destructive and non-destructive methods of analysis come under this purview. The destructive analyses results are evaluated in the Safeguards Measurement Evaluation (SME) Program, and the non-destructive analyses results in the Calorimetric Exchange (CALEX) Program. This report describes the activities in the SME program during calendar year (CY) 2007. A separate report has been prepared and issued for the CY 2007 CALEX program.

Participants: The following DOE laboratories participated in the CY 2007 SME program: Idaho National Laboratory, Los Alamos National Laboratory, New Brunswick Laboratory, Savannah River Site, and the Y-12 plant. The ABACC network laboratories (five laboratories in Argentina and four in Brazil), Nuclear Material Control Center in Japan, and URENCO Laboratory in England also participated.

Test samples: Four different types of uranium test samples were used in the CY 2007 program; uranyl nitrate solution (UNH), UO_2 pellets, UO_3 powder and UF_6 . Note that plutonium test samples could not be sent this year since NBL plutonium laboratories were in “stand down” mode. The test samples were shipped to the participants and an analysis schedule was provided that specified the frequency and number of measurements to be made on each sample type.

All uranium test samples have been characterized earlier at NBL, except UF_6 , for uranium concentration and uranium isotopes abundance. The UF_6 samples were made and characterized at the Portsmouth Gaseous Diffusion Plant. The measurement results were evaluated in comparison to the characterized values.

Performance evaluation: The SME program participants analyzed the test samples at the prescribed frequencies and as a part of their routine measurements. The analyses methods are shown in Table A. The measurement results were evaluated for accuracy, precision, bias and day-to-day variations. Performance evaluation reports were issued based on these evaluations.

Table S1: CY 2007 SME program: Test samples analyzed and analyses techniques.

Lab Code	UNH	UO ₂	UO ₃	UF ₆
<u>Uranium concentration analysis</u>				
A	IDMS, XRF			
B	D&G, IDMS			
F	D&G	D&G	D&G	
G	D&G			
T		D&G		
AB	D&G			D&G
AC		D&G		
AD	D&G	D&G		
AE		D&G	D&G	D&G
AW	IDMS			
BA	D&G	D&G	D&G	
BC	D&G	D&G		
BF		D&G	D&G	
EA				D&G
<u>Uranium isotope abundance analysis*</u>				
A	TIMS			
B	TIMS			
F	TIMS	TIMS	TIMS	
T	TIMS	TIMS		
AA	TIMS	TIMS		TIMS
AW	TIMS			
BC		TIMS		GSMS
BE	ICPMS	ICPMS		
EA	ICPMS			ICPMS

*Isotopic abundance samples were natural uranium, low enriched or high enriched uranium.

Safeguards Measurement Evaluation System (SMES): The uranium measurement results were evaluated using the newly developed Safeguards Measurement Evaluation System (SMES) – a database for storing the measurement results and an evaluation system for statistical analysis of the results. Quality assurance to SMES was provided partly through comparison of the outputs of statistical evaluation reports from the new SMES as well as the old FoxPro based system.

Measurement Evaluation Program Annual Meeting: The Measurement Evaluation Program Annual Meeting was conducted on July 7, 2007 in Tucson, Arizona, a day prior to the start of INMM annual meeting at the same location. About thirty scientists/technicians attended the NBL meeting and presented seventeen technical papers. The meeting minutes was prepared and mailed to the attendees.

ABACC workshops: Two back-to-back workshops were held in August 2007, one in Argentina, and the other in Brazil. The workshops were of one week duration each and attended by about 20 scientists/technicians from each country. Subject matter experts from NBL were instructors. Hands-on training was provided in NBL modified method of Davies-Gray titration for uranium determination. Instructions needed for measurement uncertainty evaluations in Davies-Gray titration and thermal ionization mass spectrometry results were also provided. These instructions included hands-on training in using “Kragten” spread sheets for uncertainty estimations. The spread sheet - a simple and elegant software tool – based on an original work by Kragten was developed further by NIST statisticians to calculate uncertainties according to ISO guidelines. The spread sheets were given to the attendees free of charge. At these workshops, the attendees demonstrated the procedures of uranium analyses currently in use at their facilities using auto titrators and other semi-automated procedures.

Technical presentations: The NBL Standards and Evaluation Division staff attended scientific meetings and presented papers related to the measurement evaluation program:

- a) Plutonium Metal Exchange Meeting, Los Alamos, New Mexico, June 2007. Two papers were presented; i) “Introducing the New Safeguards Measurement (SME) Program Database”, and ii) “New ICPMS Instrumentation at NBL”.
- b) 48th INMM Annual Meeting, Tucson, Arizona, July 2007. One paper was presented. “New Brunswick Laboratory Measurement Evaluation Program: New Database Development”.
- c) American Chemical Society Meeting, Boston, Massachusetts, August 2007. One presentation on “Measurement Evaluation in Chronometry and Preparation of Chronometry Test Material Standards”.
- d) Conference on Nuclear safety and Nuclear Education, Obninsk, Russia, October 2007. One presentation on “Nuclear Material Sample Measurement Comparison and Performance Evaluation in Support of Nonproliferation Programs”.

Papers published: Two technical papers were published in CY 2007.

- a) "New Brunswick Laboratory Measurement Evaluation Program: New Database Development" in the proceedings of the INMM 48th annual meeting (appendix C).
- b) "New Brunswick Laboratory Safeguards Measurement Evaluation Program: Operational Features" in the International Journal of Nuclear Knowledge Management (appendix D).

NBL participation in other measurement evaluation programs: NBL not only conducted its own SME program but also participated or planning to participate in measurement evaluation programs of other facilities.

- a) Los Alamos Laboratory Plutonium Metal Exchange Program: A NBL staff attended the Plutonium Metal Exchange Workshop held in June 2007 in Los Alamos.
- b) Brazilian Safeguards Measurement Evaluation Program: NBL will provide help to LASAL (Laboratorio de Salvaguardas in Rio de Janeiro) to set up their Measurement evaluation program for Brazilian laboratories. This work involves characterizing uranium oxide (UO_2) pellets of Brazilian origin for use as test samples in that program. Sampling and analyses plans were made and the pellets were analyzed for uranium concentration and uranium isotopic abundances.
- c) Atomic Weapons Establishment Uranium Metal Exchange Program: NBL has agreed to participate in AWE uranium metal exchange program in CY 2008. Uranium metal samples of different levels of ^{235}U enrichments will be analyzed.

A. INTRODUCTION

The New Brunswick Laboratory (NBL) is a nuclear material measurement laboratory of the U.S. Department of Energy (DOE) in the Office of Science (SC). An important aspect of NBL's mission is to conduct the measurement evaluation program, a program designed to provide independent verification of internal analytical quality control as practiced by DOE facilities in nuclear material accountability measurements.

The measurement evaluation program consists of two parts: the Safeguards Measurement Evaluation (SME) program for evaluating destructive analyses measurement results (e.g., titration, mass spectrometry) of uranium and plutonium bearing materials, and the Calorimetry Exchange (CALEX) program for evaluating non-destructive analyses measurement results (e.g., calorimetry and gamma ray spectrometry) of plutonium. This annual report describes the SME program activities in CY 2007. A separate report has been prepared and issued for the CALEX program.

B. SAFEGUARDS MEASUREMENT EVALUATION PROGRAM

Material control and accountability measurements are essential elements of nuclear material safeguards. The accountability measurements are carried out either by destructive or non-destructive analyses. The methods must be capable of providing quantitative results within the acceptable limits of accuracy and precision. Large bias and/or poor precision compromise the ability to detect material loss in processing/by theft/by diversion.

In the SME program, the elemental and isotopic-abundance measurement results of uranium and plutonium test samples are evaluated for accuracy, precision, and conformity to method/material specific International Target Values (ITVs)¹. These measurements are usually made as part of the routine analysis with full cognizance to internal quality control practices. If the results reveal inadequate performance (such as insufficient accuracy, poor precision, bias, day-to-day variation) then the laboratories are advised to make improvements. The usual remedial measures are measurement procedures review, implementing proper instrument calibration procedures, checking the validity of calibration and quality control standards, and re-training analysts.

¹ Aigner H., Binner R., Kuhn E. et al. International Target Values 2000 for Measurement Uncertainties in Safeguarding Nuclear Materials. International Atomic Energy Agency Report STR-327

C. CY 2007 SAFEGUARDS MEASUREMENT EVALUATION PROGRAM

The activities and accomplishments of the CY 2007 SME program are described below. The information provided in this section is essentially the same as in the Summary section.

- a) **Test samples:** Only uranium test samples were sent for analyses. The shipments suffered delays and pointed out the need to streamline NBL shipping procedures (packaging, preparation of shipping documents and finally shipping).

Plutonium samples could not be sent because NBL plutonium laboratories were in “stand down” mode. The plutonium laboratories are expected to resume operation in CY 2009.

- b) **Test samples analyses:** Sixteen laboratories participated in the CY 2007 SME program and analyzed uranium test samples for concentration and/or isotopes abundance (see Table S1 in Summary). Approved methods of analyses were used with full cognizance to internal quality control practices.
- c) **Performance evaluation reports:** Measurement results were evaluated in comparison to the respective characterized values and performance evaluation reports were sent to the participants. In a number of cases, unacceptably long delays occurred between the submission of results and generation of evaluation reports, mainly due to insufficient help; part time help provided by two NBL statisticians was lost . Additional help was provided in late 2007 to catch up with the backlog and to operate the program in conformity to performance metrics; two chemists were assigned to the program to work part time in the program. With this help, the CY 2008 program is expected to register significant improvements – smoother shipping operations and shorter “turn-around” time for evaluation reports.
- d) **Safeguards Measurement Evaluation System:** A new safeguards Measurement Evaluation System (SMES) - a combination of database and a statistical evaluation system - was developed by Chickasaw Nation Industries computer professionals (under DOE-CH contract). It replaces a similar but less efficient system based on FoxPro. The new SMES is web based and will allow participants to enter their own measurement

results and retrieve evaluation reports. The new system, when it becomes fully functional, is expected to cut down the “turn around” time for evaluation reports to two weeks or less.

Quality assurance tests for the new SMES were completed; privacy statements, risk assessment and security plan documents were submitted to the management for review and approval. Data stored in the FoxPro system were migrated to the new SMES. The new SMES is now ready for routine use.

CY 2007 measurement results were evaluated using the new SMES; data were entered manually by SME program staff and checked for data entry errors manually. In CY 2008 SME program participants in DOE laboratories will have direct access for data entry and retrieval of performance evaluation reports. International laboratories will be provided access at a later date.

- e) **Measurement Evaluation Program Annual Meeting:** The Measurement Evaluation Program Annual Meeting was held in Tucson, Arizona on July 7, 2007, a day prior to the start of the 48th INMM Annual Meeting. About thirty scientists/technicians attended the meeting and presented seventeen technical papers. The minutes of the meeting was prepared and sent to the attendees. Holding the NBL meeting at the same location as the INMM meeting is proving to be cost-effective – savings in travel costs and marginal increase only in lodging cost (one more day of lodging).
- f) **New test samples:** A new test sample of uranium oxide (U_3O_8) was made. This sample will be available for analysis in CY 2008 SME program.
- g) **New participants:** Efforts were made to increase the number of participants in the SME program. At least four more laboratories are expected to join; two of these will be NRC laboratories in the U.S., and two others are outside the U.S.
- h) **ABACC workshops:** Two back-to-back workshops were held in August 2007, one in Argentina, and the other in Brazil. The workshops were of one week duration each and attended by about 20 scientists/technicians from each country. Subject matter experts from NBL were instructors. Hands-on training was provided in NBL modified method of

Davies-Gray titration for uranium determination. Instructions needed for measurement uncertainty evaluations in Davies-Gray titration and thermal ionization mass spectrometry results were also provided. These instructions included hands-on training in using "Kragten" spread sheets for uncertainty estimations. The spread sheet - a simple and elegant software tool – based on an original work by Kragten was developed further by NIST statisticians to calculate uncertainties according to ISO Guide to The Expression of Uncertainty in Measurements. The spread sheets were given to the attendees free of charge. Note that a large portion of the course material on statistics and uncertainty estimations were obtained from William Guthrie and Thomas Vetter of NIST. The NBL instructors thank them for permission to use the NIST material in the two workshops.

At these workshops, the attendees demonstrated the procedures of uranium analyses currently in use at their facilities using auto titrators and others semi-automated procedures. The close interactions and technical exchanges with the attendees helped NBL to learn about the strengths and weaknesses in nuclear material measurement techniques as practiced in the ABACC network laboratories. The attendees were appreciative of the opportunity given to them to participate in the workshop. These types of assistance and interactions certainly promote goodwill among nations in the international safeguard programs.

- i) **Brazilian Safeguards Measurement Evaluation Program:** NBL will provide help to LASAL (Laboratorio de Salvaguardas in Rio de Janeiro) to set up a Measurement Evaluation program for Brazilian laboratories. This collaborative work involves characterizing uranium oxide pellets made in Brazil for use as test samples in that program. The pellets were analyzed at NBL in early 2008. NIST statisticians are now evaluating the measurement results to define the uranium concentration and uranium isotopes abundances.
- j) **Atomic Weapons Establishment (AWE) Laboratory Evaluation Program:** NBL has agreed to participate in CY 2008 AWE uranium metal exchange program. Depleted and enriched uranium metal samples will be analyzed for uranium concentration and uranium isotopes abundance.

- k) **Los Alamos National Laboratory Plutonium Metal Exchange Program:** This year, NBL was unable to analyze plutonium metal samples of the LANL program. Until 2005, NBL contributed measurement results for plutonium assay by coulometry and plutonium isotopes by TIMS. It is expected that NBL will resume analyses of the exchange samples in CY 2009.

Peter Mason of the Standards and Evaluation Division represented NBL at the Plutonium Metal Exchange Workshop in Los Alamos, Mexico in June 2007. He presented two papers at the workshop: "Introducing the New Safeguards Measurement (SME) Program Database" and "New ICPMS Instrumentation at NBL".

- l) **48th INMM Annual Meeting, Tucson, Arizona, July 2007:** Seven members of staff from NBL attended the 48th INMM annual meeting. One paper "New Brunswick Laboratory Measurement Evaluation Program: New Database Development by B. Srinivasan, Michael Soriano, Joseph Waggoner and Ron Kuziel" was presented at the meeting and was also published in the proceedings of the meeting (included as appendix C to this report). Joseph Waggoner (Standards Evaluation Division) made the oral presentation.
- m) **American Chemical Society Meeting, Boston, Massachusetts:** An invited paper entitled "Measurement Evaluation in Chronometry and Preparation of Chronometry Test Material Standards by B. Srinivasan, Paul Croatto, Steve Goldberg, Usha Narayanan and Jon Neuhoff" was presented at the American Chemical Society Meeting in August 2007. Stephan Vogt (Standards Evaluation Division) represented NBL and presented the paper.
- n) **Conference on Nuclear safety and Nuclear Education (Obninsk, Russia):** A paper on "Nuclear Material Sample Measurement Comparison and Performance Evaluation in Support of Nonproliferation Programs" was presented at the Obninsk meeting in October 2007. Kattathu Mathew (Standards Evaluation Division) represented NBL and presented the paper. The same paper with a modified title "New Brunswick Laboratory Safeguards Measurement Evaluation Program: Operational Features by B. Srinivasan, Kattathu Mathew, Paul Croatto, Usha Narayanan and Jon Neuhoff" was published in the International Journal of Nuclear Knowledge Management (included as appendix D to this report).

C.1. SME Program Participants

DOE laboratories participation is mandated by the requirement in Chapter II.4.e. (7) of DOE Manual 474.1-1 of November 2000: "*Each facility's measurement control program must include participation in appropriate inter-laboratory control programs to provide independent verification of internal analytical quality control.*" In addition to DOE laboratories, facilities outside the U.S. also participate on a voluntary basis with DOE approval. Table 1 lists the CY 2007 program participants.

Table 1. CY 2007 SME Program Participants in Uranium Sample Analysis

ABACC LABORATORIES (a group of laboratories in Argentina and Brazil)
IDAHO NATIONAL LABORATORY (DOE Contractor Laboratory)
LOS ALAMOS NATIONAL LABORATORY (DOE Contractor Laboratory)
NEW BRUNSWICK LABORATORY (DOE Laboratory)
SAVANNAH RIVER SITE (DOE Contractor Laboratory)
TOKAI SAFEGUARDS ANALYTICAL LABORATORY (Japan)
URENCO (Capenhurst) LTD (U.K.)
Y-12 Plant – BWXT LLC. (DOE Contractor Laboratory)

C.2. Materials and Measurement Methods

Table 2 shows the test samples used in the program for uranium concentration determinations and the methods of analyses. Table 3 shows the same type of information for uranium isotopes abundance measurements. The participants are identified by code letters to provide confidentiality.

Table 2. Materials and methods used to evaluate uranium assay. The participating laboratories are identified by code letters only. Numbers next to codes refer to number of times the laboratory participated in the program. For example, B6 means laboratory B participated in the program six times during CY 2007.

Method	UNH Solutions	UO ₂ Pellets	UF ₆	UO ₃ Powder
Dichromate Titration (Davies-Gray)	AB6 AD1 BA2 BC1 B6 F4 G2	AC2 AD1 AE1 BA3 BC2 BF1 F2 T2	AB2 AE3 EA1	AE2 BA2 BF2 F3
IDMS	AW1 A2 B3			
XRF	A2			

Notes: UNH, uranyl nitrate solutions. UO₂, uranium dioxide pellets. UF₆, uranium hexafluoride. UO₃, uranium oxide powder. IDMS, isotope dilution mass spectrometry. XRF, X-ray fluorescence.

Table 3. Materials and methods used to evaluate uranium isotopic abundance measurements. The participant laboratories are identified by code letters only. Numbers next to codes refer to number of times the laboratory participated in the program. For example, BE5 means laboratory BE participated in the program five times during CY 2007.

Method	HEU	LEU
TIMS	A1 B3 F1 AW1	AA6 BC1 A1 AW1 B3 F4 T3
ICPMS		BE5 EA3
GSMS		BC1

Notes. HEU is highly-enriched uranium containing ≥ 20 wt % ²³⁵U. LEU is low-enriched uranium containing < 20 wt % ²³⁵U. TIMS, thermal ionization mass spectrometry. ICPMS, inductively coupled plasma mass spectrometry. GSMS, gas source mass spectrometry.

C.3. Test Materials, Shipping, and Analysis

Test materials: The SME Program test samples are made from Certified Reference Materials (CRMs) or Working Reference Materials (WRMs), or custom-made. The test samples are characterized for elemental concentrations and/or isotopic abundances, and the characterized values are used as reference values in evaluating measurement results submitted by SME program participants.

Shipping test samples and analyses: Test samples for a particular calendar year are shipped to participants in the last quarter of the previous year. CY 2007 sample shipments were delayed with concomitant reduction in participation frequency. The shipping procedures and operations are expected to be streamlined in CY 2008. NBL has hired a shipping specialist recently who is expected to provide the much needed help for on-time shipment of SME program samples.

C.4. Database and Statistical Evaluation Programs

The measurement results submitted by the participating laboratories are entered manually into a SMES database and verified manually for data entry errors. The results are evaluated using the statistics application programs. Performance evaluation reports are sent to participant laboratories and their supervisory organizations.

In CY 2007, in a limited number of cases, measurement results were entered and evaluation reports generated using the new SMES and the older FoxPro® database/evaluation system. The double entry/double evaluations were done as a part of quality assurance tests for the new system.

C.5. Statistical Evaluation of Measurement Results

The measurement results are evaluated using statistical techniques. First, the percent relative difference (% RD) of each experimental result is calculated with respect to the corresponding reference value, the latter obtained from characterization measurements. The % RD is defined as follows:

$$\% \text{ RD} = 100 \times \{(\text{observed value} - \text{reference value}) / \text{reference value}\}$$

Next, each set of % RDs is examined for outliers using a number of statistical tests. A particular result is identified as a potential outlying value if at least two of the statistical tests show it to be an “outlier” at $\geq 99\%$ significance level. The outliers are removed only after reviews by the statistician and/or the Measurement Evaluation Program Coordinator. The data set, without outliers, is then tested to identify significant sources of variation (attributable to day-to-day and/or analyst-to-analyst differences) using standard one-factor analysis of variance (ANOVA). If the ANOVA results indicate no significant variation, then the standard uncertainty is the simple standard deviation (σ) of the results divided by the square root of n , where n is the number of measurements. The coverage factor is the Student’s 95 % “t” factor with $n-1$ degrees of freedom. For example, in a set of 8 results showing no day-to-day or analyst-to-analyst variation, the number of degrees of freedom is 7, and the coverage factor is 2.36.

If the ANOVA results indicate significant day-to-day and/or analyst-to-analyst variation ($\geq 95\%$), then the standard uncertainty in the mean % RD is estimated from a combination of the mean square for the “error” and the mean square for the “model” quantities from the ANOVA, with degrees of freedom determined from Satterthwaite’s approximation. For measurements done on two days (or by two analysts), the formula for estimating the standard uncertainty in the mean % RD is reduced to the square root of the mean square for the “model” quantity obtained from ANOVA results. In this case, the coverage factor is 12.71 (i.e., the Student’s 95 % “t” factor with one degree of freedom).

The uncertainties shown in the statistical reports are the 95 % confidence limit (C.L.) of means. In the figures accompanying the reports, the 95 % confidence interval (C.I.) of the mean is constructed from the C.L. Note that the C.I. represents the interval containing all values between the mean % RD minus the C.L. and the mean % RD plus the C.L. Thus, the 95 % C.L. of the mean are just the two end points of the C.I.

A measurement is considered to be bias-free if the 95 % C.I. included zero. Otherwise, measurement bias is indicated. The simple standard deviation (σ) of the % RDs represents the precision of the measurement results.

C.6. Examples of Statistical Evaluation Reports

Two examples of the statistical analysis reports are shown in Figs.1 and 2, the former showing uranium assay results from Davies-Gray titration, and the latter from Isotope Dilution Mass Spectrometry (IDMS) measurements. There are 8 results in each set from analyses of two samples in duplicate and on two different days.

There are no outliers in Fig.1. There is no evidence for significant day-to-day variation; the statistical significance is 44.3 %. Note that variations are considered significant if they exceed 95 %, and marginally significant if the value is between 90 and 95 %. The mean % RD value is -0.154 and the 95 % C.L. uncertainty is 0.070, the latter calculated using a coverage factor of 2.36 corresponding to 7 degrees of freedom. The mean value extended by the confidence limit (-0.154 ± 0.070) does not include zero, thereby indicating negative bias in the measurements. The standard deviation (a measure of precision) of the results is 0.083.

There are no outliers in Fig. 2. However, there is evidence for significant day-to-day variation (statistical significance of 96.6 %). The mean % RD value is 0.015 and the uncertainty at 95 % C.L. is 1.319, the latter calculated using a coverage factor of 12.7 corresponding to 1 degree of freedom. The mean value extended by the confidence limit (0.015 ± 1.319) overlaps with zero, indicating no statistically-significant bias. But, this conclusion is not meaningful since the uncertainty is very large. The standard deviation (a measure of precision) of the results is 0.149.

The bias and precision International Target Values (ITVs) are shown at the bottom of the reports. These are referred to as $u(s)$ and $u(r)$ in the April 2000 draft of the “International Target Values 2000 for Measurement Uncertainties in Safeguarding Nuclear Materials”.

In Fig.1, the mean % RD of -0.154 is beyond the bias ITV of 0.1 %; the precision of 0.083 is within the precision ITV of 0.1 %. The measurement suffers from negative bias as seen in the evaluation result of (-0.154 ± 0.070). In Fig.2, the mean % RD of 0.015 is within the ITV of 0.1 %, and the precision of 0.149 is also within the ITV of 0.15 %. However, no conclusion is possible regarding bias because of the large estimated uncertainty in the measurement results (0.015 ± 1.319) - a consequence of day-to-day variation.

Statistical reports, similar to those shown in Figures 1 and 2, are generated for each set of results (material/method/analyses date specific reports). The reports are sent to the laboratories and to the supervisory organizations along with a performance evaluation letter stating the conclusions of the statistical reports. For DOE laboratories, the site offices are responsible for initiating action to bring about improvements if bias and/or precision in measurements are not within the respective target values. If needed, NBL will provide assistance through site visits, procedure reviews, and training sessions.

Figure 1
SAMPLE DATA EVALUATION REPORT
No statistically significant difference due to analysis day

U.S. Department of Energy
 New Brunswick Laboratory
 Safeguards Measurement Evaluation Program
 Data Evaluation Report

Day to Day ANOVA analysis

Report for Laboratory: XX

U02 Pellet – U Concentration

Davies-Gray Titration

Date of Report: November 30, 2003

Sample Number	Aliquant Number	Analysis Date	Reported %U	% Relative Difference	Analyst Code
95EU0079-1	1	11/03/03	88.126	-0.0034	XXX
95EU0079-1	2	11/03/03	87.990	-0.1577	XXX
95EU0079-2	1	11/03/03	88.031	-0.1112	XXX
95EU0079-2	2	11/03/03	87.892	-0.2689	XXX
95EU0079-1	3	11/04/03	88.030	-0.1123	XXX
95EU0079-1	4	11/04/03	87.950	-0.2031	XXX
95EU0079-2	3	11/04/03	87.922	-0.2349	XXX
95EU0079-2	4	11/04/03	88.002	-0.1441	XXX

Number of Results Analyzed	8
Mean % Difference	-0.154
Mean Absolute % Difference	0.154
95% C.L. of Mean (df = 7)	0.070
Standard Deviation	0.083
Between-Day Standard Deviation (df = 1)	0.054
Within-Day Standard Deviation (df = 6)	0.087
Statistical Significance of Between-Day Standard Deviation	44.3%

International target value for bias in Davies-Gray Titration is 0.1%.

International target value for precision in Davies-Gray Titration is 0.1%.


Figure 1 (cont.)

Figure 2
SAMPLE DATA EVALUATION REPORT
Statistically significant difference due to analysis day

U.S. Department of Energy
 New Brunswick Laboratory
 Safeguards Measurement Evaluation Program
 Data Evaluation Report

Day to Day ANOVA analysis

Report for Laboratory: XX

UNH Solution – U Concentration

IDMS


Date of Report: November 30, 2003

Sample Number	Aliquant Number	Analysis Date	Reported %U	% Relative Difference	Analyst Code
94NU0021-023	1	11/03/03	1.0000	-0.0590	XXX
94NU0021-023	2	11/03/03	1.0003	-0.0290	XXX
94NU0023-079	1	11/03/03	0.9991	-0.0080	XXX
94NU0023-079	2	11/03/03	0.9996	-0.2582	XXX
94NU0021-023	3	11/04/03	1.0022	0.1609	XXX
94NU0021-023	4	11/04/03	1.0004	-0.0190	XXX
94NU0023-079	3	11/04/03	1.0004	0.1221	XXX
94NU0023-079	4	11/04/03	1.0013	0.2122	XXX

Number of Results Analyzed	8
Mean % Difference	0.015
Mean Absolute % Difference	0.109
95% C.L. of Mean (df = 1)	1.319
Standard Deviation	0.149
Between-Day Standard Deviation (df = 1)	0.294
Within-Day Standard Deviation (df = 6)	0.107
Statistical Significance of Between-Day Standard Deviation	96.6%

International target value for bias in IDMS is 0.1%.

International target value for precision in IDMS is 0.15%.

Figure 2 (cont.)

C.7. SMES: Design Features and Availability to Users

SMES was developed by computer professionals from Chickasaw Nation Industries (contractor to DOE Chicago Office) and it will be maintained by them. The design features are shown in Section C.7.1.

C.7.1. SMES Design

- Modern technology: The SMES employs modern web-based technologies and uses a well supported modern database system (SQL).
- Modular programming: The SMES employs modular programming techniques and reusable code, and will be easy to maintain. The design allows for future expansions of the program, such as increase in the number of participants and evaluation of results from new methods of analyses.
- Java enterprise server architecture: SMES is designed around the Java 2 Enterprise Edition (J2EE) architecture. The system uses a dedicated J2EE Application Server.
- SQL database storage: SMES data is securely stored on an SQL database server for quick retrieval and updating of data; data backups are automated.
- Secure platform independent thin client: Laboratories will be able to enter their own measurement results to SMES via the Internet, eliminating the need to mail the data to NBL. SMES will support the most popular current web browsers with Secure Socket Layers (SSL) and will require no special browser add-ons.
- Role-based security: The system provides a number of access roles including those required for data entry, data validation and published report retrieval. SMES will provide access to participating laboratories and oversight agencies (e.g., DOE area office). Note that participant laboratories will have access only to their own data and reports.

- Security through the Web: SMES will provide secure access through the Internet by using User ID/passwords, role-based access and encryption with considerations of confidentiality of data submitted and reports generated.
- Historical Data: All historical data contained in the FoxPro® database will be migrated to SMES.
- Calculation Techniques: The time proven statistical analysis tools (e.g., outlier tests, calculation of mean and standard deviation of %RDs, tests to determine day-to-day and analyst-to-analyst variations, determination of 95% C.I. etc.) originally written for the FoxPro® application will be retained.
- Quality assurance/documentation: The DOE standards for software development, change control and quality assurance are incorporated in design and maintenance.

C.7.2. SMES Availability

The SMES became available in CY 2007 for evaluating measurement results and generating statistical reports; only SME program staff had access to the system. In late 2007, documents related to privacy, risk assessment and security were prepared and submitted to management for review and approval. The approvals have been obtained. The system will be available in CY 2008 for DOE laboratory personnel to enter data and retrieve performance evaluation reports. Training will be provided by SME program staff.

D. ANALYSES RESULTS AND REPORTING FORMAT

The uranium measurement results submitted by the participating laboratories in 2005-2007 are shown in Appendices A and B; results evaluated in the first quarter of CY 2008 but belonging to analyses in the last quarter of CY 2007 are also included. The evaluations of uranium concentration measurement results (also referred to as uranium assay results) are discussed in Section E.1 to E.4., and uranium isotopes abundance results in Section E.5. In these discussions, the laboratories are identified by code letters only.

The measurement results were evaluated in terms of mean % RD and the standard deviation of the mean % RD for each material/method/laboratory combination. Tables 4 to 12 show the code letters for the participants, the methods of analyses, the number of results (outliers removed), mean % RDs, standard deviations, bias target values and precision target values.

The % RDs and standard deviations shown in Tables 4 to 12 are also shown graphically in Figures 3 to 20. There are two types of figures: the material-measurement skeletal figures (odd numbers figures between Figure 3 and Figure 20) to evaluate bias, and the material-measurement line figures (even number figures between Figure 4 and Figure 20) to evaluate precision.

In the odd numbered figures, the mean % RDs are shown as diamonds. The vertical line represents the standard deviation. The bias target values are shown as dotted horizontal lines. If the diamonds (extended by the respective standard deviation of the results) fall within the horizontal lines, then the measurements are said to satisfy the bias target values; those falling outside fail. The magnitude of bias (if any) can be estimated only with reference to the mean % RD and its uncertainty as 95 % C.L. No bias is indicated if the mean % RD extended by the 95 % C.L. includes zero. If it fails to include zero, bias is indicated; above zero indicates positive bias and below zero indicates negative bias.

The even number figures between Figure 4 and Figure 20 show precisions achieved in the measurements. The vertical line represents the standard deviation associated with each set of mean % RDs. If the top of the vertical line is below the corresponding precision target value - shown as a dotted horizontal line - then the laboratory has satisfied the precision target value. If the vertical line extends beyond the dotted horizontal line, then the laboratory has failed the precision criterion. In these figures, the diamonds represent the absolute values of the mean % RDs. The measurements are assumed to be bias-free if the diamonds fall on the abscissa or very close to it. The magnitude of bias can be estimated only with reference to % RD taken in conjunction with uncertainty (as 95 % C.L.).

Figures 21 to 76 in Section F show evaluations of uranium assay and uranium isotopes abundance measurement results submitted during 2005-2007. Unlike Figures 3 to 20 where results from all laboratories are shown, the three year evaluation figures are laboratory specific.

E. PERFORMANCE EVALUATION: MATERIAL BY MATERIAL

The uranium assay evaluations are discussed in Sections E.1 to E.4 and uranium isotopes abundance evaluations in Section E.5.

E.1. Uranyl Nitrate Solutions

Test samples of uranyl nitrate solutions were made from both enriched uranium (> 0.7 % in ^{235}U) and natural uranium. Three different types of uranyl nitrate solutions were made: one solution from 50 % enriched material, three solutions from 90 % enriched material, and three solutions from natural uranium. The uranium concentrations of these solutions were in the range of 7 to 10 mg uranium per gram of solution. The uranium contents of the three natural uranium solutions differed from each other by no more than 0.2 %. The uranium contents of the three solutions from 90 % material also differed from each other by not more than 0.2 %. These solutions, with closely-spaced but distinguishably different values for concentrations, are ideal samples to test accuracy and precision in uranium concentration measurements.

E.1.1. Preparation and Packaging for Shipment

The uranyl nitrate solutions were sent to participating laboratories in flame-sealed glass ampoules with break-off tips. Each ampoule was packed in a plastic bag. The bag was wrapped in absorbent cushioning material and sealed in another large plastic bag. The large bag was then kept inside a screw-cap fiberboard can for shipping.

E.1.2. Reference Value and Uncertainty

NBL used a modified Davies and Gray titration procedure to characterize the uranium concentrations of the test samples in the ampoules. The uncertainties (95 % C.L.) in uranium concentrations were as follows: $\pm 0.1\%$ for the 50 % enriched uranium solution, $\pm 0.02\%$ for the 90 % enriched uranium solutions, and in the range of ± 0.02 to $\pm 0.05\%$ for the natural uranium solutions.

A separate experiment demonstrated that the solutions did not suffer concentration change as a result of flame sealing. Samples withdrawn from sealed ampoules of natural uranium solutions and from the original stock, showed negligible differences between them; the uranium concentrations agreed within a few hundredths of one percent.

E.1.3. Performance Evaluation

Seven laboratories determined the uranium concentrations of the test samples (uranyl nitrate solutions) using Davies & Gray titration; three laboratories used isotope dilution mass spectrometry (IDMS) and one laboratory used X-ray fluorescence (XRF).

The evaluation results are shown in Table 4, along with the target values for each method. Fig. 3 and Fig. 4 are graphical representation of results in Table 4.

D& G results in conformity with bias target value:	Laboratories AB, AD, F, and G
D&G results in conformity with precision target value:	Laboratories AB, AD, BC, F, and G
D& G results non-conformity with bias target value:	Laboratories BA, BC, and B
D& G results non-conformity with precision target value:	Laboratories BA and B
IDMS results in conformity with bias target value:	Laboratories AW and A
IDMS results in conformity with precision target value:	Laboratory AW
IDMS results non-conformity with bias target value:	Laboratory B
IDMS results non-conformity with precision target value:	Laboratories A and B
XRF results in conformity with bias target value:	Laboratory A
XRF results in conformity with precision target value:	Laboratory A

Table 4. Inter-laboratory performance summary for uranium assay in UNH solutions

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
Davies & Gray Titration	AB	-0.071	0.079	48	0.1	0.1
	AD	-0.056	0.062	12	0.1	0.1
	BA	-0.433	0.443	36	0.1	0.1
	BC	-3.949	0.062	24	0.1	0.1
	B	0.177	0.329	40	0.1	0.1
	F	0.037	0.067	32	0.1	0.1
	G	-0.070	0.051	16	0.1	0.1
	AW	-0.031	0.090	12	0.1	0.15
IDMS	A	0.059	0.357	16	0.1	0.15
	B	-0.666	0.557	24	0.1	0.15
X-Ray Fluorescence [#]	A	0.235	0.268	16	0.5	0.5

ITV's are not available for X-ray Fluorescence and are assumed to be same as DOE target values.



Figure 3

Figure 4

E.2. Enriched Uranium Dioxide (UO_2) Pellet

The uranium dioxide (UO_2) pellets were originally made in a single batch at the Westinghouse Commercial Nuclear Fuel Division (a NRC licensee), using a high temperature sintering process at 1700 °C for 20 hours in a reducing atmosphere. The UO_2 pellets are known to be stable. They suffer no compositional change on exposure to air and are resistant to moisture uptake. The pellets served as a test material for both uranium assay and uranium isotopic abundance measurements. The ^{235}U content is about 4.5 %.

E.2.1. Preparation and Packaging for Shipment

The UO_2 pellets were wrapped in low-lint tissue to prevent chipping, placed in snap-cap glass bottles, and the bottles sealed in plastic bags. The bottles were shipped in fiberboard cans.

E.2.2. Reference Value and Uncertainty

The elemental uranium concentration of the pellets was determined by the NBL high-precision titration method. A uranium metal assay standard was used for quality control and traceability. The uranium concentration was measured with an uncertainty (as 95 % C.L) of about $\pm 0.02\%$.

E.2.3. Performance Evaluation

Eight laboratories analyzed the uranium dioxide pellets for uranium concentration using Davies-Gray Titration. The mean of % RDs along with uncertainties are shown in Table 5 along with the target values and also in Figs. 5 & 6.

D& G results in conformity with bias target value:	Laboratories AC, AD, AE, BA, BC, BF, F and T
D&G results in conformity with precision target value:	Laboratories AC, AD, AE, BA, BF, F and A
D& G results non-conformity with bias target value:	None
D& G results non-conformity with precision target value:	Laboratory BC

Table 5. Inter-laboratory performance summary for uranium assay in UO₂ Pellets

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
Davies-Gray Titration	AC	-0.011	0.024	16	0.1	0.1
	AD	-0.066	0.010	8	0.1	0.1
	AE	-0.064	0.068	8	0.1	0.1
	BA	-0.096	0.077	51	0.1	0.1
	BC	0.093	0.124	52	0.1	0.1
	BF	-0.003	0.074	18	0.1	0.1
	F	-0.007	0.041	27	0.1	0.1
	T	0.032	0.090	16	0.1	0.1



Figure 5

Figure 6

E.3. Uranium Hexafluoride (UF₆)

In FY 1993, Portsmouth Gaseous Diffusion Plant donated two sampling manifolds to NBL for transferring UF₆ from 2S cylinders to P-10 tubes. One of the two manifolds was used to transfer natural UF₆, and the other for enriched material. These manifolds have been taken out of service. Now, NBL is relying on Portsmouth Gaseous Diffusion facility for the preparation of SME test samples. The UF₆ samples analyzed this year were made by Portsmouth.

E.3.1. Preparation and Packaging for Shipment

The Portsmouth Gaseous Diffusion facility prepared and packaged UF₆ test samples in P-10 tubes. Each test sample contained 7 to 12 g of UF₆.

E.3.2. Reference Value and Uncertainty

The UF₆ test samples used in the SME Program since October 2004 were not characterized for assay because of the “stand down” of laboratory activities at NBL. Calculated values (based upon the assumption of 100 % purity) were used instead. Characterized isotopic values were based upon Portsmouth data which were verified at NBL using gas source mass spectrometry (GSMS).

E.3.3. Performance Evaluation

Three laboratories reported results for uranium assay in UF₆ using Davies-Gray Titration. The mean of % RDs along with uncertainties are shown in Table 6 and also in Figs.7 & 8. Laboratories AB and EA were in conformity with both bias and precision target values. Laboratory AE was in conformity with precision target value but missed the bias target value.

Table 6. Performance summary for uranium assay in UF₆

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
Davies-Gray Titration	AB	-0.050	0.048	12	0.1	0.1
	AE	-0.152	0.094	16	0.1	0.1
	EA	-0.079	0.036	8	0.1	0.1



Figure 7

Figure 8

E.4. Uranium Oxide (UO_3) Powder

UO_3 powder is an ideal test material to monitor the capability of a laboratory in analyzing hygroscopic materials. It was used as a test material several years ago, but was discontinued for sometime in between because of a perceived lack of interest in this material. A few years ago, it was re-introduced as a test material at the request of a participant laboratory. Four different laboratories analyzed UO_3 powder in CY 2007.

E.4.1. Preparation and Packaging for Shipment

The test samples come from preparations done several years ago; they were packaged under dry nitrogen atmosphere into pharmaceutical vials with Teflon-lined stoppers. The vials were crimp sealed, then sealed in plastic, and packaged in cardboard tubes for shipping.

E.4.2. Reference Value and Uncertainty

The elemental concentration of uranium in UO_3 material was characterized through analysis of 8 different samples using the NBL-modified Davies and Gray titration method. Quality control and traceability were provided through analysis of CRM 112-A (a uranium metal assay standard). The uranium content of the test samples differed from the original value by about 0.064 %, the new value being lower. The uncertainty (95 % C.L.) in the new measurements was 0.012 %. Apparently, the concentration of uranium in the UO_3 material was not altered to a significant extent.

E.4.3. Performance Evaluation

Four laboratories analyzed the UO_3 test samples for uranium concentration using Davies & Gray method. The mean % RDs along with uncertainties are shown in Table 7 and also in Figs.9 & 10. Laboratories AE, BF and F were in conformity with both bias and precision target values; laboratory BA met the precision target value but missed the bias target value.

A fifth laboratory reported difficulties in analyzing the UO_3 samples. They suspected that moisture uptake might have compromised the integrity of the test samples. They requested NBL to redo the characterization experiments and provide a procedure for analyzing the samples for uranium concentration on a moisture-free basis. These experiments were done at NBL in CY 2007 with inconclusive results (see Section F).

Table 7. Inter-laboratory performance summary of uranium assay in UO₃

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
Davies-Gray Titration	AE	-0.047	0.049	16	0.1	0.1
	BA	-0.165	0.098	44	0.1	0.1
	BF	-0.009	0.097	16	0.1	0.1
	F	-0.082	0.070	80	0.1	0.1



Figure 9

Figure 10

E.5. ^{235}U Enrichment

Both high and low enriched uranium test samples are available for evaluating isotopic abundance results. Highly-enriched Uranium (HEU) test samples include three uranyl nitrate solutions with 90 % enrichment, and one uranyl nitrate solution with 50 % enrichment. Low-Enriched Uranium (LEU) samples comprise one uranyl nitrate solution with 4 % enrichment, solid UO_2 pellets of about 4 % enrichment, hydrolyzed UF_6 of about 4.5 % enrichment, UO_3 powder of about 0.9% enrichment, and several UF_6 samples with varying levels of enrichment.

E.5.1. Preparation and Packaging for Shipment

The uranyl nitrate solutions were packaged in flame-sealed glass ampoules with a break-off tip. The ampoules were sealed in plastic, wrapped in absorbent cushioning, sealed in plastic again, and packaged in cardboard tubes for shipping. Each solution contained 5 to 10 mg uranium/g solution.

The UO_2 pellets were packaged in a snap-cap glass bottle with a low-lint tissue for cushioning to prevent chipping. The glass bottles are sealed in plastic, and packaged in cardboard tubes for shipping.

The UO_3 powder samples were packaged under dry nitrogen atmosphere in pharmaceutical vials with Teflon-lined stoppers. The vials were crimp sealed, then sealed in plastic, and packaged in cardboard tubes for shipping.

The UF_6 test samples in P-10 tubes were packed in sealed plastic bags and shipped in cardboard containers with screw caps.

E.5.2. Reference Value and Uncertainty

The uranium isotopic abundances in the UNH, UO_2 , and UO_3 test materials were characterized by thermal ionization mass spectrometry (TIMS). The experimental results were corrected for mass fractionation effects. The correction factors were determined through analyses of appropriate Certified Reference Materials performed under the same conditions as the test materials.

UF_6 material was characterized by TIMS and/or gas source mass spectrometry (GSMS). The TIMS measurements required hydrolyzed UF_6 samples, and GSMS measurements were made on UF_6 itself.

The uncertainties (95 % C.L.) in ^{235}U abundance by TIMS were as follows: 0.02 % for the 4 % enriched uranyl nitrate solution; < 0.01 % for the 50 % and 90 % enriched solutions; 0.07 % for UO_2 pellets; and 0.053 % for UF_6 . The uncertainties for the uranyl nitrate solutions did not include the uncertainties in determining the mass fractionation correction factors, whereas the uncertainties in UO_2 and UF_6 included mass fractionation correction factor uncertainties.

E.5.3. Performance Evaluation

Low enriched UNH solutions: Six laboratories analyzed the UNH solutions (LEU) via thermal ionization mass spectrometry (TIMS) and two laboratories analyzed via ICPMS. The % RDs along with uncertainties are shown in Table 8 and also in Figs. 11 & 12. Note that international target values are available for TIMS measurements only; the ICPMS target values are assumed to be the same as TIMS.

TIMS/ICPMS results in conformity with bias target value:	Laboratories A, F, T and EA
TIMS/ICPMS results in conformity with precision target value:	Laboratories A, F, T and EA
TIMS/ICPMS results non-conformity with bias target value:	Laboratories AA, AW, B, and BE
TIMS/ICPMS results non-conformity with precision target value:	Laboratories AA, AW, B, and BE

Table 8. Inter-laboratory performance summary for ^{235}U enrichment in LEU materials

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
TIMS	AA	-0.119	0.131	20	0.1	0.1
	A	0.026	0.068	8	0.1	0.1
	AW	0.113	0.137	3	0.1	0.1
	B	0.102	0.152	20	0.1	0.1
	F	-0.005	0.026	12	0.1	0.1
	T	0.029	0.019	8	0.1	0.1
ICPMS [#]	BE	-1.009	1.230	32	0.1	0.1
	EA	-0.009	0.023	8	0.1	0.1

ITV's are not available for ICPMS and are assumed to be the same as those for TIMS.



Figure 11

Figure 12

Low enriched UF₆: Three laboratories analyzed the UF₆ samples, using three different types of instruments: TIMS, ICPMS and GSMS. The % RDs along with uncertainties are shown in Table 9 and also in Figs. 13 & 14. The target values for GSMS are lower than TIMS/ICPMS target values. Note that target values are available only for TIMS; ICPMS target values are assumed to be the same as TIMS.

TIMS/ICPMS results in conformity with bias target value:	Laboratories AA and EA
TIMS/ICPMS results in conformity with precision target value:	Laboratory EA
TIMS/ICPMS results non-conformity with bias target value:	None
TIMS/ICPMS results non-conformity with precision target value:	Laboratory AA
GSMS results in conformity with bias target value:	Laboratory BC
GSMS results in conformity with precision target value:	None
GSMS results non-conformity with bias target value:	None
GSMS results non-conformity with precision target value:	Laboratory BC

Table 9. Inter-laboratory performance summary for ²³⁵U enrichment in UF₆

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
TIMS	AA	0.080	0.112	28	0.1	0.1
GSMS	BC	0.017	0.250	12	0.05	0.05
ICPMS [#]	EA	0.008	0.065	24	0.1	0.1

ITV's are not available for ICPMS and are assumed to be the same as those for TIMS.



Figure 13

Figure 14

Low enriched UO₂ pellet: Six laboratories analyzed the UO₂ pellets. Four of the six laboratories used thermal ionization mass spectrometry (TIMS) and one laboratory used ICPMS. The % RDs along with uncertainties are shown in Table 10 and also in Figs. 15 & 16. Note that international target values are available for TIMS measurements only; the ICPMS target values are assumed to be the same as TIMS.

TIMS/ICPMS results in conformity with bias target value:	Laboratories BC, F and T
TIMS/ICPMS results in conformity with precision target value:	Laboratories AA, BC, F and T
TIMS/ICPMS results non-conformity with bias target value:	Laboratories AA and BE
TIMS/ICPMS results non-conformity with precision target value:	Laboratory BE

Table 10. Inter-laboratory performance summary for ²³⁵U enrichment in UO₂

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
TIMS	AA	-0.128	0.044	8	0.1	0.1
	BC	0.029	0.092	8	0.1	0.1
	F	-0.018	0.021	16	0.1	0.1
	T	0.027	0.027	16	0.1	0.1
ICPMS [#]	BE	-0.975	1.142	48	0.1	0.1

ITV's are not available for ICPMS and are assumed to be the same as those for TIMS.



Figure 15

Figure 16

UO₃ powder: Only one laboratory analyzed the UO₃ samples via thermal ionization mass spectrometry (TIMS). The % RD and the uncertainty are shown in Table 11 and in Figs. 17 & 18. The results are in conformity to both bias and precision target values.

Table 11. Inter-laboratory performance summary for ²³⁵U enrichment in UO₃

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
TIMS	F	-0.071	0.033	8	0.1	0.1

Figure 17





Figure 18

High enriched UNH solution: Four laboratories analyzed the UNH solutions using thermal ionization mass spectrometry (TIMS). The % RDs along with uncertainties are shown in Table 12 and in Figs. 19 & 20. All laboratories were able to measure ^{235}U abundance within the bias and precision target values.

Table 12. Inter-laboratory performance summary for ^{235}U enrichment in UNH (HEU)

Method	Lab Code	Mean % RD	Standard deviation	N	ITV	
					Bias	Precision
TIMS	A	0.010	0.023	8	0.05	0.05
	B	0.028	0.028	20	0.05	0.05
	F	0.008	0.010	19	0.05	0.05
	AW	0.013	0.022	12	0.05	0.05



Figure 19

Figure 20

E.6. Plutonium Assay and Isotopic Abundance

Plutonium test samples could not be sent to participants because of the “stand down” of plutonium laboratory operations.

F. SPECIAL STUDY: MOISTURE IN UO₃ TEST SAMPLES

One of the laboratories in the SME program (laboratory A) suspected that the integrity of the UO₃ powder test samples might have been compromised by moisture uptake. They wanted NBL to verify whether there are any differences between samples analyzed as received and samples analyzed after drying to remove the moisture. NBL analyzed eight UO₃ sample but with no definite conclusions. NBL will repeat the analysis in CY 2008 this time using a well defined statistical sampling and analysis plan to determine whether moisture uptake is significant. If found to be significant, analysis procedure for these samples will be modified – sample drying before weighing samples for analysis. If simple drying procedures do not yield reproducible uranium assay results, then this particular set of test samples will be withdrawn.

G. LONG TERM EVALUTION OF URANIUM MEASUREMENTS

The uranium assay and isotopic results submitted by the participating laboratories during the years 2005 to 2007 are evaluated to provide long-term trends. The evaluation results are shown in Figs. 21 to 76. The % RDs calculated from the submitted results are shown in Figs. 21 to 76. Each figure represents results from one laboratory only for a specific material/method combination. For example, Fig. 21 shows results from laboratory AB for uranyl nitrate solution analysis by D&G titration.


Figure 21


Figure 22


Figure 23


Figure 24


Figure 25


Figure 26


Figure 27


Figure 28


Figure 29


Figure 30


Figure 31


Figure 32


Figure 33


Figure 34


Figure 35


Figure 36


Figure 37


Figure 38


Figure 39


Figure 40


Figure 41


Figure 42


Figure 43


Figure 44


Figure 45


Figure 46


Figure 47


Figure 48


Figure 49


Figure 50


Figure 51


Figure 52


Figure 53


Figure 54


Figure 55


Figure 56


Figure 57


Figure 58


Figure 59


Figure 60

Figure 61

**New Brunswick Laboratory Safeguards Measurement Evaluation Program
U235 Enrichment - HEU**





Figure 62

Figure 63





Figure 64


Figure 65


Figure 66


Figure 67

Figure 68





Figure 69


Figure 70


Figure 71


Figure 72


Figure 73


Figure 74



Figure 75

Figure 76

[This page intentionally left blank.]

APPENDICES

Appendix A: Uranium Assay Results

Appendix B: Uranium Isotopic Results

Appendix C: NBL Measurement Evaluation Program: New Database Development

Appendix D: New Brunswick Laboratory Safeguards Measurement Evaluation Program:
Operational Features

Key to symbols in the tables in the appendices

Material Symbols

UNH	Uranyl Nitrate Solution
UO ₂	Uranium Dioxide Pellet
UF ₆	Uranium Hexafluoride
UO ₃	Uranium Trioxide Powder
HEU	Highly Enriched Uranium
LEU	Low Enriched Uranium

Method Type Symbols

DG	Davies-Gray Titration
IDMS	Isotope Dilution Mass Spectrometry
XRFL	X-Ray Fluorescence - Liquid
XRFS	X-Ray Fluorescence - Solid
TIMS	Thermal Ionization Mass Spectrometry
GSMS	Gas Source Mass Spectrometry
ICPMS	Inductively Coupled Plasma Mass Spectrometry

[This page intentionally left blank.]

Appendix A: Uranium Assay Results

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
UNH	DG	AB	7/11/07	1.0032	-0.082	FP-AV
UNH	DG	AB	7/11/07	1.0037	-0.032	FP-AV
UNH	DG	AB	7/11/07	1.0035	-0.052	FP-AV
UNH	DG	AB	7/11/07	1.0105	-0.302	FP-AV
UNH	DG	AB	7/11/07	1.0123	-0.124	FP-AV
UNH	DG	AB	7/11/07	1.012	-0.154	FP-AV
UNH	DG	AB	7/12/07	1.0044	0.038	FP-AV
UNH	DG	AB	7/12/07	1.004	-0.002	FP-AV
UNH	DG	AB	7/12/07	1.0042	0.018	FP-AV
UNH	DG	AB	7/12/07	1.0121	-0.144	FP-AV
UNH	DG	AB	7/12/07	1.012	-0.154	FP-AV
UNH	DG	AB	7/12/07	1.013	-0.055	FP-AV
UNH	DG	AB	3/4/08	1.00376	-0.026	AV-FP
UNH	DG	AB	3/4/08	1.00389	-0.013	AV-FP
UNH	DG	AB	3/4/08	1.00409	0.007	AV-FP
UNH	DG	AB	3/4/08	1.01232	-0.122	AV-FP
UNH	DG	AB	3/4/08	1.01217	-0.137	AV-FP
UNH	DG	AB	3/4/08	1.0124	-0.114	AV-FP
UNH	DG	AB	3/5/08	1.00416	0.014	AV-FP
UNH	DG	AB	3/5/08	1.00418	0.016	AV-FP
UNH	DG	AB	3/5/08	1.00394	-0.008	AV-FP
UNH	DG	AB	3/5/08	1.01314	-0.041	AV-FP
UNH	DG	AB	3/5/08	1.0124	-0.114	AV-FP
UNH	DG	AB	3/5/08	1.01228	-0.126	AV-FP
UNH	DG	AB	7/11/07	1.0032	-0.082	FP-AV
UNH	DG	AB	7/11/07	1.0037	-0.032	FP-AV
UNH	DG	AB	7/11/07	1.0035	-0.052	FP-AV
UNH	DG	AB	7/12/07	1.0044	0.038	FP-AV
UNH	DG	AB	7/12/07	1.004	-0.002	FP-AV
UNH	DG	AB	7/12/07	1.0042	0.018	FP-AV
UNH	DG	AB	7/11/07	1.0105	-0.302	FP-AV
UNH	DG	AB	7/11/07	1.0123	-0.124	FP-AV
UNH	DG	AB	7/11/07	1.012	-0.154	FP-AV
UNH	DG	AB	7/12/07	1.0121	-0.144	FP-AV
UNH	DG	AB	7/12/07	1.012	-0.154	FP-AV
UNH	DG	AB	7/12/07	1.013	-0.055	FP-AV
UNH	DG	AB	3/4/08	1.00376	-0.026	AV-FP
UNH	DG	AB	3/4/08	1.00389	-0.013	AV-FP
UNH	DG	AB	3/4/08	1.00409	0.007	AV-FP
UNH	DG	AB	3/5/08	1.00416	0.014	AV-FP
UNH	DG	AB	3/5/08	1.00418	0.016	AV-FP
UNH	DG	AB	3/5/08	1.00394	-0.008	AV-FP
UNH	DG	AB	3/4/08	1.01232	-0.122	AV-FP
UNH	DG	AB	3/4/08	1.01217	-0.137	AV-FP
UNH	DG	AB	3/4/08	1.0124	-0.114	AV-FP
UNH	DG	AB	3/5/08	1.01314	-0.041	AV-FP
UNH	DG	AB	3/5/08	1.0124	-0.114	AV-FP

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
Date	Result					
UNH	DG	AB	3/5/08	1.01228	-0.126	AV-FP
UNH	DG	AD	3/4/08	1.00376	-0.026	AV-FP
UNH	DG	AD	3/4/08	1.00389	-0.013	AV-FP
UNH	DG	AD	3/4/08	1.00409	0.007	AV-FP
UNH	DG	AD	3/4/08	1.01232	-0.122	AV-FP
UNH	DG	AD	3/4/08	1.01217	-0.137	AV-FP
UNH	DG	AD	3/4/08	1.0124	-0.114	AV-FP
UNH	DG	AD	3/5/08	1.00416	0.014	AV-EA
UNH	DG	AD	3/5/08	1.00418	0.016	AV-EA
UNH	DG	AD	3/5/08	1.00394	-0.008	AV-EA
UNH	DG	AD	3/5/08	1.01314	-0.041	AV-EA
UNH	DG	AD	3/5/08	1.0124	-0.114	AV-EA
UNH	DG	AD	3/5/08	1.01228	-0.126	AV-EA
UNH	DG	BA	11/26/07	1.0139	0.034	
UNH	DG	BA	11/26/07	1.015	0.142	
UNH	DG	BA	11/26/07	1.0142	0.063	
UNH	DG	BA	11/26/07	1.0146	0.103	
UNH	DG	BA	11/26/07	1.0135	-0.006	
UNH	DG	BA	11/26/07	1.0102	-0.332	
UNH	DG	BA	11/26/07	1.014	0.043	
UNH	DG	BA	11/26/07	1.0139	0.034	
UNH	DG	BA	11/27/07	1.0133	-0.026	
UNH	DG	BA	11/27/07	1.0147	0.112	
UNH	DG	BA	11/27/07	1.0136	0.004	
UNH	DG	BA	11/27/07	1.0138	0.024	
UNH	DG	BA	11/27/07	1.0112	-0.233	
UNH	DG	BA	11/27/07	1.0126	-0.095	
UNH	DG	BA	11/27/07	1.0113	-0.223	
UNH	DG	BA	11/27/07	1.0142	0.063	
UNH	DG	BA	3/17/08	0.9925	-1.147	
UNH	DG	BA	3/17/08	0.9877	-1.625	
UNH	DG	BA	3/17/08	0.9957	-0.829	
UNH	DG	BA	3/17/08	0.9966	-0.739	
UNH	DG	BA	3/17/08	0.9959	-0.809	
UNH	DG	BA	3/17/08	0.9966	-0.739	
UNH	DG	BA	3/17/08	0.9985	-0.550	
UNH	DG	BA	3/17/08	0.9951	-0.888	
UNH	DG	BA	3/17/08	0.9984	-0.560	
UNH	DG	BA	3/17/08	0.9962	-0.779	
UNH	DG	BA	3/17/08	0.9987	-0.530	
UNH	DG	BA	3/18/08	0.9973	-0.669	
UNH	DG	BA	3/18/08	0.9929	-1.108	
UNH	DG	BA	3/18/08	0.9956	-0.839	
UNH	DG	BA	3/18/08	1	-0.400	
UNH	DG	BA	3/18/08	1.0018	-0.221	
UNH	DG	BA	3/18/08	0.9969	-0.709	
UNH	DG	BA	3/18/08	0.9975	-0.649	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
			<u>Date</u>	<u>Result</u>		
UNH	DG	BA	3/18/08	0.9967	-0.729	
UNH	DG	BA	3/18/08	0.9963	-0.769	
UNH	DG	BC	9/27/07	0.9644	-3.946	ICO
UNH	DG	BC	9/27/07	0.9639	-3.996	ICO
UNH	DG	BC	9/27/07	0.9643	-3.956	ICO
UNH	DG	BC	9/27/07	0.964	-3.986	ICO
UNH	DG	BC	9/27/07	0.9639	-3.996	ICO
UNH	DG	BC	9/27/07	0.9647	-3.916	ICO
UNH	DG	BC	9/27/07	0.9743	-3.873	ICO
UNH	DG	BC	9/27/07	0.974	-3.903	ICO
UNH	DG	BC	9/27/07	0.9743	-3.873	ICO
UNH	DG	BC	9/27/07	0.974	-3.903	ICO
UNH	DG	BC	9/27/07	0.9742	-3.883	ICO
UNH	DG	BC	9/27/07	0.9741	-3.893	ICO
UNH	DG	BC	10/2/07	0.9639	-3.996	ICO
UNH	DG	BC	10/2/07	0.9631	-4.076	ICO
UNH	DG	BC	10/2/07	0.9637	-4.016	ICO
UNH	DG	BC	10/2/07	0.9638	-4.006	ICO
UNH	DG	BC	10/2/07	0.9632	-4.066	ICO
UNH	DG	BC	10/2/07	0.9638	-4.006	ICO
UNH	DG	BC	10/2/07	0.9734	-3.962	ICO
UNH	DG	BC	10/2/07	0.9736	-3.943	ICO
UNH	DG	BC	10/2/07	0.9745	-3.854	ICO
UNH	DG	BC	10/2/07	0.9738	-3.923	ICO
UNH	DG	BC	10/2/07	0.9738	-3.923	ICO
UNH	DG	BC	10/2/07	0.9742	-3.883	ICO
UNH	DG	B	1/26/06	1.0063	0.227	3747
UNH	DG	B	1/30/06	1.0064	0.237	5011
UNH	DG	B	1/25/06	1.003	0.245	3747
UNH	DG	B	1/26/06	1.0029	0.235	3747
UNH	DG	B	1/30/06	1.0029	0.235	5011
UNH	DG	B	1/18/06	1.0042	0.018	5011
UNH	DG	B	1/18/06	1.0037	-0.032	5011
UNH	DG	B	1/18/06	1.0015	0.095	5011
UNH	DG	B	1/18/06	1.001	0.045	5011
UNH	DG	B	1/19/06	1.0013	0.075	5011
UNH	DG	B	1/19/06	1.0012	0.065	5011
UNH	DG	B	3/29/06	1.00182	-0.048	4905
UNH	DG	B	3/29/06	1.00181	-0.049	4905
UNH	DG	B	4/27/06	1.0011	-0.120	3747
UNH	DG	B	4/27/06	0.9991	-0.319	3747
UNH	DG	B	3/29/06	1.00706	0.475	4905
UNH	DG	B	3/29/06	1.00706	0.475	4905
UNH	DG	B	4/27/06	1.008	0.569	3747
UNH	DG	B	4/27/06	1.0043	0.200	3747
UNH	DG	B	7/24/06	1.00015	-0.215	796
UNH	DG	B	7/24/06	1.00225	-0.005	796

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
UNH	DG	B	7/24/06	1.00024	-0.031	796
UNH	DG	B	7/24/06	1.0014	0.085	796
UNH	DG	B	10/9/06	1.0043	0.200	374
UNH	DG	B	10/9/06	1.0046	0.229	374
UNH	DG	B	10/11/06	1.0018	0.125	374
UNH	DG	B	10/11/06	1.0021	0.155	374
UNH	DG	B	7/24/07	1.0143	0.073	5011
UNH	DG	B	7/24/07	1.0176	0.399	5011
UNH	DG	B	7/24/07	1.0134	-0.016	5011
UNH	DG	B	7/24/07	1.0139	0.034	5011
UNH	DG	B	7/25/07	1.0137	0.014	3375
UNH	DG	B	7/25/07	1.0164	0.280	3375
UNH	DG	B	7/25/07	1.0172	0.359	3375
UNH	DG	B	7/25/07	1.0152	0.162	3375
UNH	DG	B	10/30/07	1.0177	0.408	753
UNH	DG	B	10/30/07	1.0187	0.507	753
UNH	DG	B	10/30/07	1.0176	0.399	753
UNH	DG	B	10/30/07	1.0135	-0.006	753
UNH	DG	B	11/4/07	1.0173	0.369	747
UNH	DG	B	11/4/07	1.0176	0.399	747
UNH	DG	B	11/4/07	1.0177	0.408	747
UNH	DG	B	11/4/07	1.0185	0.487	747
UNH	DG	B	1/17/08	1.0188	0.517	966
UNH	DG	B	1/17/08	1.0172	0.359	966
UNH	DG	B	1/18/08	1.0177	0.408	747
UNH	DG	B	1/18/08	1.018	0.438	747
UNH	DG	B	1/18/08	1.0177	0.408	747
UNH	DG	B	1/18/08	1.0173	0.369	747
UNH	DG	B	1/19/08	1.0132	-0.036	686
UNH	DG	B	1/19/08	1.0221	0.843	686
UNH	DG	B	3/27/08	1.0155	0.191	4905
UNH	DG	B	3/27/08	1.0132	-0.036	4905
UNH	DG	B	4/1/08	1.0106	-0.292	4686
UNH	DG	B	4/1/08	1.0141	0.053	4686
UNH	DG	B	4/13/08	1.0011	-1.229	3747
UNH	DG	B	4/13/08	1.0167	0.310	3747
UNH	DG	B	4/13/08	1.011	-0.253	3747
UNH	DG	B	4/13/08	1.0157	0.211	3747
UNH	DG	F	11/17/05	1.0032	-0.082	237
UNH	DG	F	11/17/05	1.0034	-0.062	237
UNH	DG	F	11/18/05	1.0041	0.008	237
UNH	DG	F	11/18/05	1.004	-0.002	237
UNH	DG	F	11/21/05	1.0043	0.028	231
UNH	DG	F	11/21/05	1.0044	0.038	231
UNH	DG	F	11/22/05	1.0043	0.028	231
UNH	DG	F	11/22/05	1.0049	0.088	231
UNH	DG	F	11/17/05	0.9999	-0.065	237

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
			<u>Date</u>	<u>Result</u>		
UNH	DG	F	11/17/05	0.9995	-0.105	237
UNH	DG	F	11/18/05	1	-0.055	237
UNH	DG	F	11/18/05	1	-0.055	237
UNH	DG	F	11/21/05	0.9999	-0.065	231
UNH	DG	F	11/21/05	0.9998	-0.075	231
UNH	DG	F	11/22/05	1.0012	0.065	231
UNH	DG	F	11/22/05	1.0026	0.205	231
UNH	DG	F	9/25/07	1.005	0.098	GS
UNH	DG	F	9/25/07	1.005	0.098	GS
UNH	DG	F	9/25/07	1.001	0.045	GS
UNH	DG	F	9/25/07	1	-0.055	GS
UNH	DG	F	9/26/07	1.005	0.098	GS
UNH	DG	F	9/26/07	1.005	0.098	GS
UNH	DG	F	9/26/07	1.001	0.045	GS
UNH	DG	F	9/26/07	1.002	0.145	GS
UNH	DG	F	9/13/07	1.005	0.098	MM
UNH	DG	F	9/13/07	1.005	0.098	MM
UNH	DG	F	9/13/07	1.001	0.045	MM
UNH	DG	F	9/13/07	1.001	0.045	MM
UNH	DG	F	9/14/07	1.006	0.197	MM
UNH	DG	F	9/14/07	1.006	0.197	MM
UNH	DG	F	9/14/07	1.001	0.045	MM
UNH	DG	F	9/14/07	1.001	0.045	MM
UNH	DG	F	7/25/07	1.004	-0.002	GS
UNH	DG	F	7/25/07	1.004	-0.002	GS
UNH	DG	F	7/25/07	1.003	0.070	GS
UNH	DG	F	7/25/07	1.002	-0.030	GS
UNH	DG	F	7/26/07	1.004	-0.002	GS
UNH	DG	F	7/26/07	1.004	-0.002	GS
UNH	DG	F	7/26/07	1.002	-0.030	GS
UNH	DG	F	7/26/07	1.002	-0.030	GS
UNH	DG	F	7/26/07	1.004	-0.002	MM
UNH	DG	F	7/26/07	1.004	-0.002	MM
UNH	DG	F	7/26/07	1.002	-0.030	MM
UNH	DG	F	7/26/07	1.002	-0.030	MM
UNH	DG	F	7/26/07	1.002	-0.030	MM
UNH	DG	F	7/27/07	1.004	-0.002	MM
UNH	DG	F	7/27/07	1.004	-0.002	MM
UNH	DG	F	7/27/07	1.002	-0.030	MM
UNH	DG	G	12/9/04	1.00423	0.021	
UNH	DG	G	12/10/04	1.00405	0.003	
UNH	DG	G	12/10/04	1.00392	-0.010	
UNH	DG	G	12/9/04	1.00057	0.002	
UNH	DG	G	12/9/04	1.00059	0.004	
UNH	DG	G	12/10/04	1.0008	0.025	
UNH	DG	G	12/10/04	1.00069	0.014	
UNH	DG	G	3/23/05	1.00412	0.010	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
			<u>Date</u>	<u>Result</u>		
UNH	DG	G	3/23/05	1.00425	0.023	
UNH	DG	G	3/24/05	1.00381	-0.021	
UNH	DG	G	3/24/05	1.00389	-0.013	
UNH	DG	G	3/23/05	1.00253	0.023	
UNH	DG	G	3/23/05	1.00219	-0.011	
UNH	DG	G	3/24/05	1.00201	-0.029	
UNH	DG	G	3/24/05	1.00219	-0.011	
UNH	DG	G	10/4/05	1.00419	0.017	
UNH	DG	G	10/4/05	1.00397	-0.005	
UNH	DG	G	10/5/05	1.00413	0.011	
UNH	DG	G	10/5/05	1.00378	-0.024	
UNH	DG	G	10/4/05	1.00077	0.022	
UNH	DG	G	10/4/05	1.00043	-0.012	
UNH	DG	G	10/5/05	1.00073	0.018	
UNH	DG	G	10/5/05	1.00079	0.024	
UNH	DG	G	2/22/06	1.00414	0.012	
UNH	DG	G	2/22/06	1.00388	-0.014	
UNH	DG	G	2/23/06	1.00379	-0.023	
UNH	DG	G	2/23/06	1.00411	0.009	
UNH	DG	G	2/22/06	1.00041	-0.014	
UNH	DG	G	2/22/06	1.00052	-0.003	
UNH	DG	G	2/23/06	1.00055	0.000	
UNH	DG	G	2/23/06	1.00052	-0.003	
UNH	DG	G	4/19/06	1.00428	0.026	
UNH	DG	G	4/19/06	1.00425	0.023	
UNH	DG	G	4/20/06	1.00399	-0.003	
UNH	DG	G	4/20/06	1.00404	0.002	
UNH	DG	G	4/19/06	1.00252	0.022	
UNH	DG	G	4/19/06	1.00228	-0.002	
UNH	DG	G	4/20/06	1.00261	0.031	
UNH	DG	G	4/20/06	1.0021	-0.020	
UNH	DG	G	8/24/06	1.00045	-0.010	
UNH	DG	G	8/24/06	1.00057	0.002	
UNH	DG	G	8/24/06	1.00211	-0.019	
UNH	DG	G	8/24/06	1.00205	-0.025	
UNH	DG	G	8/25/06	1.00047	-0.008	
UNH	DG	G	8/25/06	1.0006	0.005	
UNH	DG	G	8/25/06	1.00217	-0.013	
UNH	DG	G	8/25/06	1.00224	-0.006	
UNH	DG	G	11/16/07	10.1329	-0.027	EL,KG
UNH	DG	G	11/16/07	10.1375	0.019	EL,KG
UNH	DG	G	11/16/07	10.1237	-0.117	EL,KG
UNH	DG	G	11/16/07	10.1254	-0.101	EL,KG
UNH	DG	G	12/6/07	10.1353	-0.003	EL,KG
UNH	DG	G	12/6/07	10.1395	0.038	EL,KG
UNH	DG	G	12/6/07	10.1311	-0.044	EL,KG
UNH	DG	G	12/6/07	10.1254	-0.101	EL,KG

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
			<u>Date</u>	<u>Result</u>		
UNH	DG	G	1/30/08	10.1257	-0.098	
UNH	DG	G	1/30/08	10.1253	-0.102	
UNH	DG	G	1/30/08	10.1284	-0.071	
UNH	DG	G	1/30/08	10.1238	-0.116	
UNH	DG	G	1/31/08	10.1269	-0.086	
UNH	DG	G	1/31/08	10.1261	-0.094	
UNH	DG	G	1/31/08	10.1222	-0.132	
UNH	DG	G	1/31/08	10.1274	-0.081	
UNH	DG	V	3/21/06	1.00227	-0.174	1
UNH	DG	V	3/21/06	1.00217	-0.184	1
UNH	DG	V	3/21/06	1.00177	-0.224	1
UNH	DG	V	4/5/06	1.00249	-0.152	1
UNH	DG	V	4/5/06	1.00229	-0.172	1
UNH	DG	V	4/5/06	1.00339	-0.063	1
UNH	DG	V	4/25/06	1.00356	-0.046	1
UNH	DG	V	4/25/06	1.00356	-0.046	1
UNH	DG	V	4/25/06	1.00356	-0.046	1
UNH	DG	V	3/21/06	0.99927	-0.128	1
UNH	DG	V	3/21/06	1.00017	-0.038	1
UNH	DG	V	3/21/06	0.99967	-0.088	1
UNH	DG	V	4/5/06	0.99911	-0.144	1
UNH	DG	V	4/5/06	1.00011	-0.044	1
UNH	DG	V	4/5/06	0.99982	-0.073	1
UNH	DG	V	4/25/06	1.00044	-0.011	1
UNH	DG	V	4/25/06	1.0006	0.005	1
UNH	DG	V	4/25/06	1.00052	-0.003	1

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
<u>Date</u>	<u>Result</u>					
UNH	IDMS	AW	5/21/07	6.824	0.101	
UNH	IDMS	AW	5/21/07	6.8131	-0.059	
UNH	IDMS	AW	5/21/07	6.8187	0.023	
UNH	IDMS	AW	5/21/07	6.8113	-0.085	
UNH	IDMS	AW	5/21/07	4.6939	-0.155	
UNH	IDMS	AW	5/21/07	4.6948	-0.136	
UNH	IDMS	AW	6/1/07	4.6961	-0.108	
UNH	IDMS	AW	6/4/07	6.8111	-0.088	
UNH	IDMS	AW	6/4/07	6.821	0.057	
UNH	IDMS	AW	6/4/07	4.7053	0.087	
UNH	IDMS	AW	6/4/07	4.6985	-0.057	
UNH	IDMS	AW	6/4/07	4.7034	0.047	
UNH	IDMS	A	11/8/04	1.0024	-0.161	GPW/JM
UNH	IDMS	A	11/8/04	1.0036	-0.042	GPW/JM
UNH	IDMS	A	11/10/04	1.0037	-0.032	DLB/JM
UNH	IDMS	A	11/10/04	1.0048	0.078	DLB/JM
UNH	IDMS	A	11/8/04	1.0011	-0.120	GPW/JM
UNH	IDMS	A	11/8/04	1.003	0.070	GPW/JM
UNH	IDMS	A	11/10/04	1.0013	-0.100	DLB/JM
UNH	IDMS	A	11/10/04	1.0038	0.150	DLB/JM
UNH	IDMS	A	9/14/05	1.0038	-0.022	BLM/GPW
UNH	IDMS	A	9/15/05	1.0044	0.038	DLB/GPW
UNH	IDMS	A	9/19/05	1.0042	0.018	WS/GPW
UNH	IDMS	A	9/19/05	1.0041	0.008	WS/GPW
UNH	IDMS	A	9/14/05	1.0004	-0.015	BLM/GPW
UNH	IDMS	A	9/14/05	1.0004	-0.015	BLM/GPW
UNH	IDMS	A	9/15/05	1.0008	0.025	DLB/GPW
UNH	IDMS	A	9/15/05	1.0008	0.025	DLB/GPW
UNH	IDMS	A	9/19/05	1.0008	0.025	WS/GPW
UNH	IDMS	A	9/19/05	1.0007	0.015	WS/GPW
UNH	IDMS	A	6/21/07	1.0155	0.191	BLM/JM
UNH	IDMS	A	6/21/07	1.0118	-0.174	BLM/JM
UNH	IDMS	A	6/21/07	1.0126	-0.095	BLM/JM
UNH	IDMS	A	6/21/07	1.0191	0.547	BLM/JM
UNH	IDMS	A	6/25/07	1.0151	0.152	PRR/JM
UNH	IDMS	A	6/25/07	1.0142	0.063	PRR/JM
UNH	IDMS	A	6/25/07	1.0154	0.182	PRR/JM
UNH	IDMS	A	6/25/07	1.0139	0.034	PRR/JM
UNH	IDMS	A	9/25/07	1.0123	-0.124	BLM/JM
UNH	IDMS	A	9/25/07	1.0131	-0.045	BLM/JM
UNH	IDMS	A	9/25/07	1.0126	-0.095	BLM/JM
UNH	IDMS	A	9/25/07	1.0116	-0.193	BLM/JM
UNH	IDMS	A	9/26/07	1.0112	-0.233	LSB/JM
UNH	IDMS	A	9/26/07	1.0252	1.148	LSB/JM
UNH	IDMS	A	9/26/07	1.0106	-0.292	LSB/JM
UNH	IDMS	A	9/26/07	1.0124	-0.114	LSB/JM

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
			<u>Date</u>	<u>Result</u>		
UNH	IDMS	B	1/4/05	0.4698	-0.068	CDN
UNH	IDMS	B	1/4/05	0.4696	-0.111	CDN
UNH	IDMS	B	1/8/05	0.4654	-1.004	JLB
UNH	IDMS	B	1/8/05	0.4655	-0.983	JLB
UNH	IDMS	B	1/4/05	0.4593	-1.296	CDN
UNH	IDMS	B	1/4/05	0.4591	-1.339	CDN
UNH	IDMS	B	1/8/05	0.4562	-1.962	JLB
UNH	IDMS	B	1/8/05	0.4563	-1.941	JLB
UNH	IDMS	B	1/9/06	0.9978	-0.620	PAM
UNH	IDMS	B	1/9/06	0.9984	-0.560	PAM
UNH	IDMS	B	1/12/06	1.0015	-0.251	DDN
UNH	IDMS	B	1/12/06	1.0006	-0.341	DDN
UNH	IDMS	B	1/9/06	0.9957	-0.485	PAM
UNH	IDMS	B	1/9/06	0.9977	-0.285	PAM
UNH	IDMS	B	1/12/06	1.0057	0.515	DDN
UNH	IDMS	B	1/12/06	1.0018	0.125	DDN
UNH	IDMS	B	3/4/06	0.9957	-0.829	WPB
UNH	IDMS	B	3/4/06	0.9962	-0.779	WPB
UNH	IDMS	B	3/5/06	0.9914	-1.257	DDN
UNH	IDMS	B	3/5/06	1	-0.400	DDN
UNH	IDMS	B	3/4/06	0.9962	-0.435	WPB
UNH	IDMS	B	3/4/06	1.0004	-0.015	WPB
UNH	IDMS	B	3/5/06	0.9899	-1.064	DDN
UNH	IDMS	B	3/5/06	0.9893	-1.124	DDN
UNH	IDMS/U-233 spike	B	4/12/06	1.0007	-0.331	JMG
UNH	IDMS/U-233 spike	B	4/12/06	0.9982	-0.580	JMG
UNH	IDMS/U-233 spike	B	4/13/06	0.9916	-1.237	CPT
UNH	IDMS/U-233 spike	B	4/13/06	0.9932	-1.078	CPT
UNH	IDMS/U-233 spike	B	4/12/06	0.9958	-0.819	JMG
UNH	IDMS/U-233 spike	B	4/12/06	0.9956	-0.839	JMG
UNH	IDMS/U-233 spike	B	4/13/06	0.9959	-0.809	CPT
UNH	IDMS/U-233 spike	B	4/13/06	0.9971	-0.689	CPT
UNH	IDMS	B	7/22/06	0.9955	-0.678	PWB
UNH	IDMS	B	7/22/06	0.9981	-0.419	PWB
UNH	IDMS	B	7/22/06	0.9975	-0.305	PWB
UNH	IDMS	B	7/22/06	0.994	-0.655	PWB
UNH	IDMS	B	7/31/06	1.0167	1.437	DDB
UNH	IDMS	B	7/31/06	1.0039	0.160	DDB
UNH	IDMS	B	8/1/06	0.9909	-0.964	DDB
UNH	IDMS	B	8/1/06	0.9877	-1.284	DDB
UNH	IDMS	B	1/18/08	10.092	-0.430	DDN
UNH	IDMS	B	1/18/08	10.108	-0.272	DDN
UNH	IDMS	B	1/18/08	10.053	-0.815	DDN
UNH	IDMS	B	1/18/08	10.048	-0.864	DDN
UNH	IDMS	B	1/26/08	10.075	-0.598	JCP
UNH	IDMS	B	1/26/08	10.053	-0.815	JCP
UNH	IDMS	B	1/26/08	10.076	-0.588	JCP

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
<u>Date</u>	<u>Result</u>					
UNH	IDMS	B	1/26/08	10.062	-0.726	JCP
UNH	IDMS	B	3/25/08	10.035	-0.993	PAM
UNH	IDMS	B	3/25/08	10.029	-1.052	PAM
UNH	IDMS	B	3/25/08	10.021	-1.131	PAM
UNH	IDMS	B	3/25/08	10.042	-0.923	PAM
UNH	IDMS	B	3/27/08	10.031	-1.032	LLB
UNH	IDMS	B	3/27/08	10.024	-1.101	LLB
UNH	IDMS	B	3/27/08	10.046	-0.884	LLB
UNH	IDMS	B	3/27/08	10.028	-1.062	LLB
UNH	IDMS	G	5/5/06	0.4702	0.017	3061
UNH	IDMS	G	5/10/06	0.4701	-0.004	3060
UNH	IDMS	G	5/10/06	0.4703	0.038	3061
UNH	IDMS	G	5/10/06	0.4699	-0.047	3073
UNH	IDMS	G	5/10/06	0.4701	-0.004	3074
UNH	IDMS	G	5/5/06	0.4629	-0.013	3062
UNH	IDMS	G	5/5/06	0.463	0.009	3063
UNH	IDMS	G	5/10/06	0.4633	0.073	3075
UNH	IDMS	G	5/11/06	0.4632	0.052	3076
UNH	IDMS	G	5/11/06	0.4651	-0.049	3064
UNH	IDMS	G	5/11/06	0.4648	-0.114	3065
UNH	IDMS	G	5/18/06	0.4649	-0.092	3081
UNH	IDMS	G	5/5/06	0.4654	0.015	3066
UNH	IDMS	G	5/5/06	0.4656	0.058	3067
UNH	IDMS	G	5/11/06	0.4652	-0.028	3080
UNH	XRFL	A	11/4/04	1.002	-0.030	MER/SJB
UNH	XRFL	A	11/4/04	1.001	-0.130	MER/SJB
UNH	XRFL	A	11/30/04	0.999	-0.329	MER/RBD
UNH	XRFL	A	11/30/04	1.003	0.070	MER/RBD
UNH	XRFL	A	11/4/04	0.998	-0.255	MER/SJB
UNH	XRFL	A	11/4/04	1	-0.055	MER/SJB
UNH	XRFL	A	11/30/04	0.998	-0.255	MER/RBD
UNH	XRFL	A	11/30/04	0.998	-0.255	MER/RBD
UNH	XRFL	A	9/29/05	1.002	-0.201	ACB
UNH	XRFL	A	9/29/05	1.002	-0.201	ACB
UNH	XRFL	A	10/3/05	1.002	-0.201	ACB
UNH	XRFL	A	10/3/05	1.004	-0.002	ACB
UNH	XRFL	A	9/29/05	1	-0.055	ACB
UNH	XRFL	A	9/29/05	1.001	0.045	ACB
UNH	XRFL	A	10/3/05	0.997	-0.355	ACB
UNH	XRFL	A	10/3/05	0.999	-0.155	ACB
UNH	XRFL	A	6/20/07	1.014	0.043	MER
UNH	XRFL	A	6/20/07	1.015	0.142	MER
UNH	XRFL	A	6/20/07	1.014	0.043	MER
UNH	XRFL	A	6/20/07	1.019	0.537	MER
UNH	XRFL	A	7/12/07	1.02	0.635	MER
UNH	XRFL	A	7/12/07	1.016	0.241	MER

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
			<u>Date</u>	<u>Result</u>		
UNH	XRFL	A	7/12/07	1.019	0.537	MER
UNH	XRFL	A	7/12/07	1.02	0.635	MER
UNH	XRFL	A	9/27/07	1.017	0.339	MER
UNH	XRFL	A	9/27/07	1.015	0.142	MER
UNH	XRFL	A	9/27/07	1.014	0.043	MER
UNH	XRFL	A	9/27/07	1.011	-0.253	MER
UNH	XRFL	A	10/1/07	1.012	-0.154	MER
UNH	XRFL	A	10/1/07	1.016	0.241	MER
UNH	XRFL	A	10/1/07	1.018	0.438	MER
UNH	XRFL	A	10/1/07	1.015	0.142	MER

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	AB	3/30/06	88.128	-0.001	AV
UO ₂	DG	AB	3/30/06	88.149	0.023	AV
UO ₂	DG	AB	3/30/06	88.123	-0.007	AV
UO ₂	DG	AB	3/31/06	88.091	-0.043	AV
UO ₂	DG	AB	3/31/06	88.143	0.016	AV
UO ₂	DG	AB	3/31/06	88.103	-0.030	AV
UO ₂	DG	AB	3/30/06	88.124	-0.006	AV
UO ₂	DG	AB	3/30/06	88.120	-0.010	AV
UO ₂	DG	AB	3/30/06	88.044	-0.096	AV
UO ₂	DG	AB	3/31/06	88.076	-0.060	AV
UO ₂	DG	AB	3/31/06	88.071	-0.066	AV
UO ₂	DG	AB	3/31/06	88.085	-0.050	AV
UO ₂	DG	AC	6/5/06	88.028	-0.115	AL
UO ₂	DG	AC	6/5/06	88.037	-0.104	AL
UO ₂	DG	AC	6/6/06	87.970	-0.180	NDS
UO ₂	DG	AC	6/6/06	87.983	-0.166	NDS
UO ₂	DG	AC	6/5/06	87.986	-0.162	NDS
UO ₂	DG	AC	6/5/06	87.977	-0.172	NDS
UO ₂	DG	AC	6/6/06	87.972	-0.178	AL
UO ₂	DG	AC	6/6/06	87.964	-0.187	AL
UO ₂	DG	AC	11/16/06	88.128	-0.001	AL
UO ₂	DG	AC	11/16/06	88.137	0.009	AL
UO ₂	DG	AC	11/16/06	88.096	-0.037	NDS
UO ₂	DG	AC	11/16/06	88.097	-0.036	NDS
UO ₂	DG	AC	11/17/06	88.129	0.000	NDS
UO ₂	DG	AC	11/17/06	88.147	0.020	NDS
UO ₂	DG	AC	11/17/06	88.098	-0.035	AL
UO ₂	DG	AC	11/17/06	88.11	-0.022	AL
UO ₂	DG	AC	11/20/07	88.071	-0.066	AL
UO ₂	DG	AC	11/20/07	88.102	-0.031	AL
UO ₂	DG	AC	11/21/07	88.116	-0.015	NDS
UO ₂	DG	AC	11/21/07	88.128	-0.001	NDS
UO ₂	DG	AC	11/28/07	88.131	0.002	NDS
UO ₂	DG	AC	11/28/07	88.147	0.020	NDS
UO ₂	DG	AC	11/29/07	88.138	0.010	AL
UO ₂	DG	AC	11/29/07	88.127	-0.002	AL
UO ₂	DG	AD	6/2/06	88.097	-0.036	EB/CD
UO ₂	DG	AD	6/2/06	88.126	-0.003	EB/CD
UO ₂	DG	AD	6/2/06	88.115	-0.016	EB/CD
UO ₂	DG	AD	6/2/06	88.087	-0.048	EB/CD
UO ₂	DG	AD	6/5/06	88.113	-0.018	EB/CD
UO ₂	DG	AD	6/5/06	88.121	-0.009	EB/CD
UO ₂	DG	AD	6/5/06	88.120	-0.010	EB/CD
UO ₂	DG	AD	6/5/06	88.109	-0.023	EB/CD
UO ₂	DG	AD	11/23/06	88.059	-0.079	EB
UO ₂	DG	AD	11/23/06	88.069	-0.068	CD

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	AD	11/23/06	88.075	-0.061	EB
UO ₂	DG	AD	11/23/06	88.064	-0.074	CD
UO ₂	DG	AD	11/25/06	88.088	-0.047	EB
UO ₂	DG	AD	11/25/06	88.065	-0.073	CD
UO ₂	DG	AD	11/25/06	88.075	-0.061	EB
UO ₂	DG	AD	11/25/06	88.07	-0.067	CD
UO ₂	DG	AE	4/26/06	88.0960	-0.037	JR-JM
UO ₂	DG	AE	4/26/06	88.1055	-0.027	JR-JM
UO ₂	DG	AE	4/26/06	88.1239	-0.006	JR-JM
UO ₂	DG	AE	4/26/06	88.1190	-0.011	JR-JM
UO ₂	DG	AE	4/27/06	88.1469	0.020	JR-JM
UO ₂	DG	AE	4/27/06	88.1042	-0.028	JR-JM
UO ₂	DG	AE	4/27/06	88.1658	0.042	JR-JM
UO ₂	DG	AE	4/27/06	88.1668	0.043	JR-JM
UO ₂	DG	AE	11/23/06	88.122	-0.008	JR-EL
UO ₂	DG	AE	11/23/06	88.164	0.040	JR-EL
UO ₂	DG	AE	11/23/06	88.091	-0.043	JR-EL
UO ₂	DG	AE	11/23/06	88.095	-0.039	JR-EL
UO ₂	DG	AE	11/24/06	88.065	-0.073	JR-EL
UO ₂	DG	AE	11/24/06	88.006	-0.140	JR-EL
UO ₂	DG	AE	11/24/06	88.06	-0.078	JR-EL
UO ₂	DG	AE	11/24/06	87.977	-0.172	JR-EL
UO ₂	DG	BA	4/6/06	88.152	0.092	
UO ₂	DG	BA	4/6/06	88.083	0.014	
UO ₂	DG	BA	4/6/06	88.125	0.061	
UO ₂	DG	BA	4/6/06	88.052	-0.022	
UO ₂	DG	BA	4/6/06	88.099	0.032	
UO ₂	DG	BA	4/7/06	88.032	-0.044	
UO ₂	DG	BA	4/7/06	88.005	-0.075	
UO ₂	DG	BA	4/7/06	88.027	-0.050	
UO ₂	DG	BA	4/7/06	87.936	-0.153	
UO ₂	DG	BA	4/7/06	88.025	-0.052	
UO ₂	DG	BA	4/6/06	88.031	-0.045	
UO ₂	DG	BA	4/6/06	88.027	-0.050	
UO ₂	DG	BA	4/6/06	88.058	-0.015	
UO ₂	DG	BA	4/6/06	88.140	0.078	
UO ₂	DG	BA	4/6/06	88.099	0.032	
UO ₂	DG	BA	4/7/06	87.951	-0.136	
UO ₂	DG	BA	4/7/06	88.048	-0.026	
UO ₂	DG	BA	4/7/06	87.930	-0.160	
UO ₂	DG	BA	4/7/06	87.997	-0.084	
UO ₂	DG	BA	4/7/06	87.966	-0.119	
UO ₂	DG	BA	11/23/06	88.055	-0.084	
UO ₂	DG	BA	11/23/06	88.15	0.024	
UO ₂	DG	BA	11/23/06	88.114	-0.017	
UO ₂	DG	BA	11/23/06	88.094	-0.040	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	BA	11/23/06	88.131	0.002	
UO ₂	DG	BA	11/23/06	88.099	-0.034	
UO ₂	DG	BA	11/23/06	88.137	0.009	
UO ₂	DG	BA	11/23/06	88.08	-0.056	
UO ₂	DG	BA	11/24/06	88.102	-0.031	
UO ₂	DG	BA	11/24/06	88.08	-0.056	
UO ₂	DG	BA	11/24/06	88.124	-0.006	
UO ₂	DG	BA	11/24/06	88.136	0.008	
UO ₂	DG	BA	11/24/06	88.132	0.003	
UO ₂	DG	BA	11/24/06	88.152	0.026	
UO ₂	DG	BA	11/24/06	88.062	-0.076	
UO ₂	DG	BA	11/24/06	88.113	-0.018	
UO ₂	DG	BA	11/26/07	88.009	-0.136	
UO ₂	DG	BA	11/26/07	88.001	-0.145	
UO ₂	DG	BA	11/26/07	87.987	-0.161	
UO ₂	DG	BA	11/26/07	88.02	-0.124	
UO ₂	DG	BA	11/26/07	87.97	-0.180	
UO ₂	DG	BA	11/26/07	87.977	-0.172	
UO ₂	DG	BA	11/26/07	87.991	-0.157	
UO ₂	DG	BA	11/27/07	87.972	-0.178	
UO ₂	DG	BA	11/27/07	88.01	-0.135	
UO ₂	DG	BA	11/27/07	87.982	-0.167	
UO ₂	DG	BA	11/27/07	87.928	-0.228	
UO ₂	DG	BA	11/27/07	87.962	-0.189	
UO ₂	DG	BA	11/27/07	87.956	-0.196	
UO ₂	DG	BA	11/27/07	87.914	-0.244	
UO ₂	DG	BA	3/17/08	88.12	-0.010	
UO ₂	DG	BA	3/17/08	88.053	-0.086	
UO ₂	DG	BA	3/17/08	88.105	-0.027	
UO ₂	DG	BA	3/17/08	88.052	-0.087	
UO ₂	DG	BA	3/17/08	88.052	-0.087	
UO ₂	DG	BA	3/17/08	87.961	-0.191	
UO ₂	DG	BA	3/17/08	87.998	-0.149	
UO ₂	DG	BA	3/17/08	88.087	-0.048	
UO ₂	DG	BA	3/17/08	88.087	-0.048	
UO ₂	DG	BA	3/17/08	88.023	-0.120	
UO ₂	DG	BA	3/17/08	88.103	-0.030	
UO ₂	DG	BA	3/17/08	87.961	-0.191	
UO ₂	DG	BA	3/18/08	88.105	-0.027	
UO ₂	DG	BA	3/18/08	87.953	-0.200	
UO ₂	DG	BA	3/18/08	87.977	-0.172	
UO ₂	DG	BA	3/18/08	88.122	-0.008	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	BA	3/18/08	88.084	-0.051	
UO ₂	DG	BA	3/18/08	87.966	-0.185	
UO ₂	DG	BA	3/18/08	87.958	-0.194	
UO ₂	DG	BA	3/18/08	88.041	-0.100	
UO ₂	DG	BA	3/18/08	88.025	-0.118	
UO ₂	DG	BC	4/13/06	88.06433	-0.073	ICO
UO ₂	DG	BC	4/13/06	88.15492	0.029	ICO
UO ₂	DG	BC	4/13/06	88.15719	0.032	ICO
UO ₂	DG	BC	4/13/06	88.08713	-0.048	ICO
UO ₂	DG	BC	4/13/06	88.05359	-0.086	ICO
UO ₂	DG	BC	4/13/06	88.01272	-0.132	ICO
UO ₂	DG	BC	4/13/06	87.97265	-0.177	ICO
UO ₂	DG	BC	4/13/06	88.02056	-0.123	ICO
UO ₂	DG	BC	4/13/06	88.00236	-0.144	ICO
UO ₂	DG	BC	4/13/06	88.02031	-0.123	ICO
UO ₂	DG	BC	4/18/06	88.12967	0.001	ICO
UO ₂	DG	BC	4/18/06	88.10743	-0.024	ICO
UO ₂	DG	BC	4/18/06	88.10258	-0.030	ICO
UO ₂	DG	BC	4/18/06	88.06009	-0.078	ICO
UO ₂	DG	BC	4/18/06	88.16297	0.039	ICO
UO ₂	DG	BC	4/18/06	88.01585	-0.128	ICO
UO ₂	DG	BC	4/18/06	88.03035	-0.112	ICO
UO ₂	DG	BC	4/18/06	88.01072	-0.134	ICO
UO ₂	DG	BC	4/18/06	88.03399	-0.108	ICO
UO ₂	DG	BC	4/18/06	88.01799	-0.126	ICO
UO ₂	DG	BC	10/17/06	88.07628	-0.060	ICO
UO ₂	DG	BC	10/17/06	88.10536	-0.027	ICO
UO ₂	DG	BC	10/17/06	88.11326	-0.018	ICO
UO ₂	DG	BC	10/17/06	88.09782	-0.035	ICO
UO ₂	DG	BC	10/17/06	88.07518	-0.061	ICO
UO ₂	DG	BC	10/17/06	88.20877	0.091	ICO
UO ₂	DG	BC	10/17/06	88.13921	0.012	ICO
UO ₂	DG	BC	10/17/06	88.16356	0.039	ICO
UO ₂	DG	BC	10/17/06	88.19854	0.079	ICO
UO ₂	DG	BC	10/17/06	88.15852	0.033	ICO
UO ₂	DG	BC	10/19/06	88.05198	-0.087	ICO
UO ₂	DG	BC	10/19/06	88.08551	-0.049	ICO
UO ₂	DG	BC	10/19/06	88.05898	-0.079	ICO
UO ₂	DG	BC	10/19/06	88.05865	-0.080	ICO
UO ₂	DG	BC	10/19/06	88.05625	-0.083	ICO
UO ₂	DG	BC	10/19/06	88.1391	0.011	ICO
UO ₂	DG	BC	10/19/06	88.15435	0.029	ICO
UO ₂	DG	BC	10/19/06	88.15698	0.032	ICO
UO ₂	DG	BC	10/19/06	88.16761	0.044	ICO
UO ₂	DG	BC	10/19/06	88.13891	0.011	ICO
UO ₂	DG	BC	10/17/07	88.3862	0.292	ICO

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	BC	10/17/07	88.2957	0.189	ICO
UO ₂	DG	BC	10/17/07	88.3437	0.244	ICO
UO ₂	DG	BC	10/17/07	88.3748	0.279	ICO
UO ₂	DG	BC	10/17/07	88.389	0.295	ICO
UO ₂	DG	BC	10/17/07	88.3956	0.303	ICO
UO ₂	DG	BC	10/17/07	88.2712	0.161	ICO
UO ₂	DG	BC	10/17/07	88.2899	0.183	ICO
UO ₂	DG	BC	10/17/07	88.1584	0.033	ICO
UO ₂	DG	BC	10/17/07	88.1217	-0.008	ICO
UO ₂	DG	BC	10/17/07	88.0746	-0.062	ICO
UO ₂	DG	BC	10/17/07	88.0775	-0.058	ICO
UO ₂	DG	BC	10/17/07	88.3706	0.274	ICO
UO ₂	DG	BC	10/17/07	88.364	0.267	ICO
UO ₂	DG	BC	10/17/07	88.2394	0.125	ICO
UO ₂	DG	BC	10/17/07	88.2341	0.119	ICO
UO ₂	DG	BC	10/24/07	88.145	0.018	ICO
UO ₂	DG	BC	10/24/07	88.1168	-0.014	ICO
UO ₂	DG	BC	10/24/07	88.1436	0.017	ICO
UO ₂	DG	BC	10/24/07	88.1153	-0.016	ICO
UO ₂	DG	BC	10/24/07	88.3143	0.210	ICO
UO ₂	DG	BC	10/24/07	88.3189	0.215	ICO
UO ₂	DG	BC	10/24/07	88.2798	0.171	ICO
UO ₂	DG	BC	10/24/07	88.2844	0.176	ICO
UO ₂	DG	BC	10/24/07	88.3438	0.244	ICO
UO ₂	DG	BC	10/24/07	88.3682	0.271	ICO
UO ₂	DG	BC	10/24/07	88.3091	0.204	ICO
UO ₂	DG	BC	10/24/07	88.3335	0.232	ICO
UO ₂	DG	BC	10/24/07	88.2801	0.171	ICO
UO ₂	DG	BC	10/24/07	88.3115	0.207	ICO
UO ₂	DG	BC	10/24/07	88.2435	0.130	ICO
UO ₂	DG	BC	10/24/07	88.2749	0.166	ICO
UO ₂	DG	BF	4/4/06	88.061	-0.077	ABC
UO ₂	DG	BF	4/4/06	88.097	-0.036	ABC
UO ₂	DG	BF	4/4/06	88.090	-0.044	ABC
UO ₂	DG	BF	4/12/06	88.122	-0.008	ABC
UO ₂	DG	BF	4/12/06	88.056	-0.083	ABC
UO ₂	DG	BF	4/12/06	88.110	-0.022	ABC
UO ₂	DG	BF	4/12/06	88.138	0.010	ABC
UO ₂	DG	BF	4/12/06	88.133	0.005	ABC
UO ₂	DG	BF	4/12/06	88.148	0.022	ABC
UO ₂	DG	BF	4/12/06	88.147	0.020	ABC
UO ₂	DG	BF	4/13/06	88.068	-0.069	ABC
UO ₂	DG	BF	4/13/06	88.142	0.015	ABC
UO ₂	DG	BF	4/13/06	88.160	0.035	ABC
UO ₂	DG	BF	4/13/06	88.141	0.014	ABC
UO ₂	DG	BF	1/15/07	88.122	-0.008	ABC
UO ₂	DG	BF	1/15/07	88.121	-0.009	ABC

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	BF	1/15/07	87.955	-0.197	ABC
UO ₂	DG	BF	1/15/07	87.998	-0.149	ABC
UO ₂	DG	BF	1/15/07	88.148	0.022	ABC
UO ₂	DG	BF	1/15/07	88.125	-0.005	ABC
UO ₂	DG	BF	1/16/07	88.127	-0.002	ABC
UO ₂	DG	BF	1/16/07	88.147	0.020	ABC
UO ₂	DG	BF	1/16/07	88.194	0.074	ABC
UO ₂	DG	BF	1/16/07	88.204	0.085	ABC
UO ₂	DG	BF	1/16/07	88.145	0.018	ABC
UO ₂	DG	BF	1/16/07	88.224	0.108	ABC
UO ₂	DG	BF	1/16/07	88.105	-0.027	ABC
UO ₂	DG	BF	1/16/07	88.178	0.056	ABC
UO ₂	DG	BF	1/16/07	88.125	-0.005	ABC
UO ₂	DG	BF	1/16/07	88.126	-0.003	ABC
UO ₂	DG	BF	1/16/07	88.138	0.010	ABC
UO ₂	DG	BF	1/16/07	88.094	-0.040	ABC
UO ₂	DG	F	9/14/07	88.154	0.028	MM
UO ₂	DG	F	9/14/07	88.098	-0.035	MM
UO ₂	DG	F	9/14/07	88.082	-0.053	MM
UO ₂	DG	F	9/14/07	88.129	0.000	MM
UO ₂	DG	F	9/14/07	88.108	-0.024	MM
UO ₂	DG	F	9/14/07	88.139	0.011	MM
UO ₂	DG	F	9/14/07	88.108	-0.024	MM
UO ₂	DG	F	9/14/07	88.123	-0.007	MM
UO ₂	DG	F	9/25/07	88.131	0.002	MM
UO ₂	DG	F	9/25/07	88.049	-0.091	MM
UO ₂	DG	F	9/25/07	87.991	-0.157	MM
UO ₂	DG	F	9/25/07	88.148	0.022	MM
UO ₂	DG	F	9/25/07	88.131	0.002	MM
UO ₂	DG	F	9/25/07	88.11	-0.022	MM
UO ₂	DG	F	9/25/07	88.158	0.033	MM
UO ₂	DG	F	9/25/07	88.169	0.045	MM
UO ₂	DG	F	10/2/07	88.119	-0.011	AV
UO ₂	DG	F	10/2/07	88.14	0.012	AV
UO ₂	DG	F	10/2/07	88.148	0.022	AV
UO ₂	DG	F	10/2/07	88.146	0.019	AV
UO ₂	DG	F	10/2/07	88.131	0.002	AV
UO ₂	DG	F	10/2/07	88.122	-0.008	AV
UO ₂	DG	F	10/4/07	88.137	0.009	AV
UO ₂	DG	F	10/4/07	88.135	0.007	AV
UO ₂	DG	F	10/4/07	88.143	0.016	AV
UO ₂	DG	F	10/4/07	88.155	0.030	AV
UO ₂	DG	F	10/4/07	88.12	-0.010	AV

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	T	12/19/05	88.20	0.081	
UO ₂	DG	T	12/19/05	88.26	0.149	
UO ₂	DG	T	12/19/05	88.24	0.126	
UO ₂	DG	T	12/19/05	88.15	0.024	
UO ₂	DG	T	12/21/05	88.13	0.001	
UO ₂	DG	T	12/21/05	88.18	0.058	
UO ₂	DG	T	12/21/05	88.10	-0.033	
UO ₂	DG	T	12/21/05	88.11	-0.022	
UO ₂	DG	T	4/12/06	88.23	0.115	
UO ₂	DG	T	4/12/06	88.20	0.081	
UO ₂	DG	T	4/12/06	88.17	0.047	
UO ₂	DG	T	4/12/06	88.22	0.103	
UO ₂	DG	T	4/19/06	88.04	-0.101	
UO ₂	DG	T	4/19/06	88.12	-0.010	
UO ₂	DG	T	4/19/06	88.15	0.024	
UO ₂	DG	T	4/19/06	88.10	-0.033	
UO ₂	DG	T	10/16/06	88.15	0.024	
UO ₂	DG	T	10/16/06	88.15	0.024	
UO ₂	DG	T	10/16/06	88.1	-0.033	
UO ₂	DG	T	10/16/06	88.12	-0.010	
UO ₂	DG	T	10/24/06	88.03	-0.112	
UO ₂	DG	T	10/24/06	88.18	0.058	
UO ₂	DG	T	10/24/06	88.03	-0.112	
UO ₂	DG	T	8/23/06	88.18	0.058	
UO ₂	DG	T	8/23/06	88.16	0.035	
UO ₂	DG	T	8/23/06	88.16	0.035	
UO ₂	DG	T	8/23/06	88.1	-0.033	
UO ₂	DG	T	8/25/06	88.1	-0.033	
UO ₂	DG	T	8/25/06	88.12	-0.010	
UO ₂	DG	T	8/25/06	88.19	0.069	
UO ₂	DG	T	8/25/06	88.16	0.035	
UO ₂	DG	T	11/7/07	88.13	0.001	
UO ₂	DG	T	11/7/07	88.09	-0.044	
UO ₂	DG	T	11/7/07	88.1	-0.033	
UO ₂	DG	T	11/7/07	88.21	0.092	
UO ₂	DG	T	11/9/07	88.25	0.137	
UO ₂	DG	T	11/9/07	88.27	0.160	
UO ₂	DG	T	11/9/07	88.25	0.137	
UO ₂	DG	T	11/9/07	88.28	0.171	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₂	DG	V	3/21/06	88.031	-0.111	1
UO ₂	DG	V	3/21/06	88.038	-0.103	1
UO ₂	DG	V	3/21/06	88.032	-0.110	1
UO ₂	DG	V	4/5/06	88.085	-0.050	1
UO ₂	DG	V	4/5/06	88.089	-0.045	1
UO ₂	DG	V	4/5/06	88.136	0.008	1
UO ₂	DG	V	4/25/06	88.113	-0.018	1
UO ₂	DG	V	4/25/06	88.074	-0.062	1
UO ₂	DG	V	4/25/06	88.145	0.018	1
UO ₂	DG	V	3/21/06	88.070	-0.067	1
UO ₂	DG	V	3/21/06	88.092	-0.042	1
UO ₂	DG	V	3/21/06	88.034	-0.108	1
UO ₂	DG	V	4/5/06	87.990	-0.158	1
UO ₂	DG	V	4/5/06	88.108	-0.024	1
UO ₂	DG	V	4/5/06	88.038	-0.103	1
UO ₂	DG	V	4/25/06	88.143	0.016	1
UO ₂	DG	V	4/25/06	88.093	-0.041	1
UO ₂	DG	V	4/25/06	88.139	0.011	1

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₃	DG	AE	10/3/07	82.697	0.031	JR-JM-EL
UO ₃	DG	AE	10/3/07	82.665	-0.007	JR-JM-EL
UO ₃	DG	AE	10/3/07	82.675	0.005	JR-JM-EL
UO ₃	DG	AE	10/3/07	82.669	-0.002	JR-JM-EL
UO ₃	DG	AE	10/4/07	82.667	-0.005	JR-JM-EL
UO ₃	DG	AE	10/4/07	82.59	-0.098	JR-JM-EL
UO ₃	DG	AE	10/4/07	82.663	-0.010	JR-JM-EL
UO ₃	DG	AE	10/4/07	82.649	-0.027	JR-JM-EL
UO ₃	DG	AE	3/17/08	82.586	-0.103	JR-JM-EL
UO ₃	DG	AE	3/17/08	82.571	-0.121	JR-JM-EL
UO ₃	DG	AE	3/17/08	82.589	-0.099	JR-JM-EL
UO ₃	DG	AE	3/17/08	82.624	-0.057	JR-JM-EL
UO ₃	DG	AE	3/18/08	82.66	-0.013	JR-JM-EL
UO ₃	DG	AE	3/18/08	82.602	-0.083	JR-JM-EL
UO ₃	DG	AE	3/18/08	82.619	-0.063	JR-JM-EL
UO ₃	DG	AE	3/18/08	82.584	-0.105	JR-JM-EL
UO ₃	DG	BA	11/26/07	82.599	-0.087	
UO ₃	DG	BA	11/26/07	82.606	-0.079	
UO ₃	DG	BA	11/26/07	82.616	-0.067	
UO ₃	DG	BA	11/26/07	82.608	-0.076	
UO ₃	DG	BA	11/26/07	82.635	-0.044	
UO ₃	DG	BA	11/26/07	82.618	-0.064	
UO ₃	DG	BA	11/26/07	82.581	-0.109	
UO ₃	DG	BA	11/26/07	82.616	-0.067	
UO ₃	DG	BA	11/27/07	82.409	-0.317	
UO ₃	DG	BA	11/27/07	82.479	-0.232	
UO ₃	DG	BA	11/27/07	82.578	-0.112	
UO ₃	DG	BA	11/27/07	82.485	-0.225	
UO ₃	DG	BA	11/27/07	82.59	-0.098	
UO ₃	DG	BA	11/27/07	82.664	-0.008	
UO ₃	DG	BA	11/27/07	82.614	-0.069	
UO ₃	DG	BA	11/27/07	82.62	-0.062	
UO ₃	DG	BA	3/19/08	82.485	-0.225	
UO ₃	DG	BA	3/19/08	82.671	0.000	
UO ₃	DG	BA	3/19/08	82.592	-0.096	
UO ₃	DG	BA	3/19/08	82.417	-0.307	
UO ₃	DG	BA	3/19/08	82.468	-0.246	
UO ₃	DG	BA	3/19/08	82.557	-0.138	
UO ₃	DG	BA	3/19/08	82.543	-0.155	
UO ₃	DG	BA	3/19/08	82.576	-0.115	
UO ₃	DG	BA	3/19/08	82.421	-0.302	
UO ₃	DG	BA	3/19/08	82.469	-0.244	
UO ₃	DG	BA	3/19/08	82.405	-0.322	
UO ₃	DG	BA	3/19/08	82.431	-0.290	
UO ₃	DG	BA	3/19/08	82.42	-0.304	
UO ₃	DG	BA	3/19/08	82.426	-0.296	
UO ₃	DG	BA	3/20/08	82.528	-0.173	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₃	DG	BA	3/20/08	82.582	-0.108	
UO ₃	DG	BA	3/20/08	82.624	-0.057	
UO ₃	DG	BA	3/20/08	82.635	-0.044	
UO ₃	DG	BA	3/20/08	82.529	-0.172	
UO ₃	DG	BA	3/20/08	82.6	-0.086	
UO ₃	DG	BA	3/20/08	82.562	-0.132	
UO ₃	DG	BA	3/20/08	82.488	-0.221	
UO ₃	DG	BA	3/20/08	82.471	-0.242	
UO ₃	DG	BA	3/20/08	82.451	-0.266	
UO ₃	DG	BA	3/20/08	82.43	-0.292	
UO ₃	DG	BA	3/20/08	82.418	-0.306	
UO ₃	DG	BA	3/20/08	82.51	-0.195	
UO ₃	DG	BA	3/20/08	82.507	-0.198	
UO ₃	DG	BF	7/27/07	82.772	0.122	ABC
UO ₃	DG	BF	7/27/07	82.768	0.117	ABC
UO ₃	DG	BF	8/1/07	82.732	0.074	ABC
UO ₃	DG	BF	8/1/07	82.737	0.080	ABC
UO ₃	DG	BF	10/3/07	82.635	-0.044	ABC
UO ₃	DG	BF	10/3/07	82.646	-0.030	ABC
UO ₃	DG	BF	10/8/07	82.597	-0.090	ABC
UO ₃	DG	BF	10/8/07	82.584	-0.105	ABC
UO ₃	DG	BF	3/27/08	82.767	0.116	ABC
UO ₃	DG	BF	3/27/08	82.575	-0.116	ABC
UO ₃	DG	BF	3/27/08	82.58	-0.110	ABC
UO ₃	DG	BF	3/27/08	82.569	-0.123	ABC
UO ₃	DG	BF	3/28/08	82.764	0.112	ABC
UO ₃	DG	BF	3/28/08	82.625	-0.056	ABC
UO ₃	DG	BF	3/28/08	82.589	-0.099	ABC
UO ₃	DG	BF	3/28/08	82.68	0.011	ABC
UO ₃	DG	F	9/18/07	82.613	-0.070	NH
UO ₃	DG	F	9/18/07	82.627	-0.053	NH
UO ₃	DG	F	9/18/07	82.663	-0.010	NH
UO ₃	DG	F	9/18/07	82.634	-0.045	NH
UO ₃	DG	F	9/18/07	82.632	-0.047	NH
UO ₃	DG	F	9/18/07	82.65	-0.025	NH
UO ₃	DG	F	9/18/07	82.637	-0.041	NH
UO ₃	DG	F	9/18/07	82.611	-0.073	NH
UO ₃	DG	F	9/19/07	82.616	-0.067	NH
UO ₃	DG	F	9/19/07	82.595	-0.092	NH
UO ₃	DG	F	9/19/07	82.586	-0.103	NH
UO ₃	DG	F	9/19/07	82.635	-0.044	NH
UO ₃	DG	F	9/19/07	82.602	-0.083	NH
UO ₃	DG	F	9/19/07	82.603	-0.082	NH
UO ₃	DG	F	9/19/07	82.592	-0.096	NH
UO ₃	DG	F	9/19/07	82.6	-0.086	NH
UO ₃	DG	F	9/21/07	82.624	-0.057	NH

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₃	DG	F	9/21/07	82.647	-0.029	NH
UO ₃	DG	F	9/21/07	82.621	-0.060	NH
UO ₃	DG	F	9/21/07	82.651	-0.024	NH
UO ₃	DG	F	9/21/07	82.603	-0.082	NH
UO ₃	DG	F	9/21/07	82.622	-0.059	NH
UO ₃	DG	F	9/21/07	82.69	0.023	NH
UO ₃	DG	F	9/21/07	82.641	-0.036	NH
UO ₃	DG	F	9/24/07	82.635	-0.044	NH
UO ₃	DG	F	9/24/07	82.663	-0.010	NH
UO ₃	DG	F	9/24/07	82.65	-0.025	NH
UO ₃	DG	F	9/24/07	82.662	-0.011	NH
UO ₃	DG	F	9/24/07	82.641	-0.036	NH
UO ₃	DG	F	9/24/07	82.607	-0.077	NH
UO ₃	DG	F	9/24/07	82.711	0.048	NH
UO ₃	DG	F	9/24/07	82.7	0.035	NH
UO ₃	DG	F	9/18/07	82.617	-0.065	MM
UO ₃	DG	F	9/18/07	82.55	-0.146	MM
UO ₃	DG	F	9/18/07	82.53	-0.171	MM
UO ₃	DG	F	9/18/07	82.605	-0.080	MM
UO ₃	DG	F	9/18/07	82.586	-0.103	MM
UO ₃	DG	F	9/18/07	82.615	-0.068	MM
UO ₃	DG	F	9/18/07	82.474	-0.238	MM
UO ₃	DG	F	9/18/07	82.463	-0.252	MM
UO ₃	DG	F	9/19/07	82.597	-0.090	MM
UO ₃	DG	F	9/19/07	82.572	-0.120	MM
UO ₃	DG	F	9/19/07	82.607	-0.077	MM
UO ₃	DG	F	9/19/07	82.573	-0.119	MM
UO ₃	DG	F	9/19/07	82.615	-0.068	MM
UO ₃	DG	F	9/19/07	82.608	-0.076	MM
UO ₃	DG	F	9/19/07	82.52	-0.183	MM
UO ₃	DG	F	9/19/07	82.577	-0.114	MM
UO ₃	DG	F	9/21/07	82.569	-0.123	MM
UO ₃	DG	F	9/21/07	82.567	-0.126	MM
UO ₃	DG	F	9/21/07	82.541	-0.157	MM
UO ₃	DG	F	9/21/07	82.51	-0.195	MM
UO ₃	DG	F	9/21/07	82.541	-0.157	MM
UO ₃	DG	F	9/21/07	82.561	-0.133	MM
UO ₃	DG	F	9/21/07	82.443	-0.276	MM
UO ₃	DG	F	9/21/07	82.5	-0.207	MM
UO ₃	DG	F	9/24/07	82.62	-0.062	MM
UO ₃	DG	F	9/24/07	82.624	-0.057	MM
UO ₃	DG	F	9/24/07	82.558	-0.137	MM
UO ₃	DG	F	9/24/07	82.606	-0.079	MM
UO ₃	DG	F	9/24/07	82.591	-0.097	MM
UO ₃	DG	F	9/24/07	82.557	-0.138	MM

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₃	DG	F	9/24/07	82.46	-0.255	MM
UO ₃	DG	F	9/24/07	82.421	-0.302	MM
UO ₃	DG	F	3/11/08	82.68	0.011	MM
UO ₃	DG	F	3/11/08	82.715	0.053	MM
UO ₃	DG	F	3/11/08	82.65	-0.025	MM
UO ₃	DG	F	3/11/08	82.598	-0.088	MM
UO ₃	DG	F	3/13/08	82.594	-0.093	MM
UO ₃	DG	F	3/13/08	82.612	-0.071	MM
UO ₃	DG	F	3/13/08	82.633	-0.046	MM
UO ₃	DG	F	3/13/08	82.636	-0.042	MM
UO ₃	DG	F	3/13/08	82.598	-0.088	NH
UO ₃	DG	F	3/13/08	82.669	-0.002	NH
UO ₃	DG	F	3/13/08	82.635	-0.044	NH
UO ₃	DG	F	3/13/08	82.637	-0.041	NH
UO ₃	DG	F	3/14/08	82.682	0.013	NH
UO ₃	DG	F	3/14/08	82.635	-0.044	NH
UO ₃	DG	F	3/14/08	82.596	-0.091	NH
UO ₃	DG	F	3/14/08	82.636	-0.042	NH
UO ₃	DG	V	3/21/06	82.588	-0.100	1
UO ₃	DG	V	3/21/06	82.620	-0.062	1
UO ₃	DG	V	3/21/06	82.593	-0.094	1
UO ₃	DG	V	4/5/06	82.577	-0.114	1
UO ₃	DG	V	4/5/06	82.581	-0.109	1
UO ₃	DG	V	4/5/06	82.539	-0.160	1
UO ₃	DG	V	4/25/06	82.610	-0.074	1
UO ₃	DG	V	4/25/06	82.608	-0.076	1
UO ₃	DG	V	3/21/06	82.609	-0.075	1
UO ₃	DG	V	3/21/06	82.648	-0.028	1
UO ₃	DG	V	3/21/06	82.651	-0.024	1
UO ₃	DG	V	4/5/06	82.586	-0.103	1
UO ₃	DG	V	4/5/06	82.597	-0.090	1
UO ₃	DG	V	4/5/06	82.577	-0.114	1
UO ₃	DG	V	4/25/06	82.598	-0.088	1
UO ₃	DG	V	4/25/06	82.647	-0.029	1
UO ₃	DG	V	4/25/06	82.623	-0.058	1
UO ₃	IDMS	A	11/9/04	82.78	0.132	BLM/JM
UO ₃	IDMS	A	11/9/04	82.59	-0.098	BLM/JM
UO ₃	IDMS	A	12/7/04	82.67	-0.001	WJS/JM
UO ₃	IDMS	A	12/7/04	82.77	0.120	WJS/JM
UO ₃	IDMS	A	12/7/04	82.71	0.047	WJS/JM
UO ₃	IDMS	A	12/7/04	82.71	0.047	WJS/JM
UO ₃	IDMS	A	11/9/04	82.48	-0.231	BLM/JM
UO ₃	IDMS	A	11/9/04	82.51	-0.195	BLM/JM
UO ₃	IDMS	A	3/23/05	82.68	0.011	BLM/JM
UO ₃	IDMS	A	3/23/05	82.69	0.023	BLM/JM
UO ₃	IDMS	A	3/24/05	82.48	-0.231	WJS/JM

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₃	IDMS	A	3/24/05	82.77	0.120	WJS/JM
UO ₃	IDMS	A	3/23/05	82.76	0.108	WJS/JM
UO ₃	IDMS	A	3/23/05	82.64	-0.037	WJS/JM
UO ₃	IDMS	A	3/24/05	82.67	-0.001	BLM/JM
UO ₃	IDMS	A	3/24/05	82.53	-0.171	BLM/JM
UO ₃	IDMS	A	9/14/05	82.58	-0.110	BLM/JM
UO ₃	IDMS	A	9/14/05	82.68	0.011	BLM/JM
UO ₃	IDMS	A	9/15/05	82.60	-0.086	DB/JM
UO ₃	IDMS	A	9/15/05	82.58	-0.110	DB/JM
UO ₃	IDMS	A	9/19/05	82.53	-0.171	WJS/JM
UO ₃	IDMS	A	9/19/05	82.68	0.011	WJS/JM
UO ₃	IDMS	A	9/14/05	82.55	-0.146	BLM/JM
UO ₃	IDMS	A	9/14/05	82.61	-0.074	BLM/JM
UO ₃	IDMS	A	9/15/05	82.66	-0.013	DB/JM
UO ₃	IDMS	A	9/15/05	82.76	0.108	DB/JM
UO ₃	IDMS	A	9/19/05	82.63	-0.050	WJS/JM
UO ₃	IDMS	A	9/19/05	82.58	-0.110	WJS/JM
UO ₃	IDMS	B	3/2/06	82.0	-0.812	PAM
UO ₃	IDMS	B	3/2/06	81.8	-1.054	PAM
UO ₃	IDMS	B	3/9/06	82.0	-0.812	DDN
UO ₃	IDMS	B	3/9/06	82.4	-0.328	DDN
UO ₃	IDMS	B	3/2/06	82.1	-0.691	PAM
UO ₃	IDMS	B	3/2/06	81.8	-1.054	PAM
UO ₃	IDMS	B	3/9/06	82.5	-0.207	DDN
UO ₃	IDMS	B	3/9/06	82.8	0.156	DDN
UO ₃	XRFL	A	11/4/04	82.35	-0.388	MER/SJB
UO ₃	XRFL	A	11/4/04	82.20	-0.570	MER/SJB
UO ₃	XRFL	A	11/30/04	82.23	-0.533	MER/RBD
UO ₃	XRFL	A	11/30/04	82.15	-0.630	MER/RBD
UO ₃	XRFL	A	11/4/04	82.41	-0.316	MER/SJB
UO ₃	XRFL	A	11/4/04	82.49	-0.219	MER/SJB
UO ₃	XRFL	A	11/30/04	82.15	-0.630	MER/RBD
UO ₃	XRFL	A	11/30/04	82.24	-0.521	MER/RBD
UO ₃	XRFL	A	4/15/05	82.09	-0.703	MER/ACB
UO ₃	XRFL	A	4/15/05	82.39	-0.340	MER/ACB
UO ₃	XRFL	A	4/18/05	82.29	-0.461	MER/ACB
UO ₃	XRFL	A	4/18/05	82.22	-0.546	MER/ACB
UO ₃	XRFL	A	4/15/05	82.33	-0.412	MER/ACB
UO ₃	XRFL	A	4/15/05	82.39	-0.340	MER/ACB
UO ₃	XRFL	A	4/18/05	82.55	-0.146	MER/ACB
UO ₃	XRFL	A	4/18/05	82.41	-0.316	MER/ACB
UO ₃	XRFL	A	9/15/05	82.87	0.241	ACB
UO ₃	XRFL	A	9/15/05	82.77	0.120	ACB
UO ₃	XRFL	A	9/27/05	82.51	-0.195	ACB
UO ₃	XRFL	A	9/27/05	82.64	-0.037	ACB
UO ₃	XRFL	A	9/15/05	82.81	0.168	ACB

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UO ₃	XRFL	A	9/15/05	82.71	0.047	ACB
UO ₃	XRFL	A	9/27/05	82.48	-0.231	ACB
UO ₃	XRFL	A	9/27/05	82.34	-0.400	ACB
UO ₃	XRFS	A	11/15/04	82.22	-0.546	MER/RBD
UO ₃	XRFS	A	11/15/04	82.16	-0.618	MER/RBD
UO ₃	XRFS	A	12/1/04	81.95	-0.872	MER/SJB
UO ₃	XRFS	A	12/1/04	81.95	-0.872	MER/SJB
UO ₃	XRFS	A	11/15/04	82.06	-0.739	MER/RBD
UO ₃	XRFS	A	11/15/04	81.94	-0.884	MER/RBD
UO ₃	XRFS	A	12/1/04	81.94	-0.884	MER/SJB
UO ₃	XRFS	A	12/1/04	82.17	-0.606	MER/SJB
UO ₃	XRFS	A	3/9/05	82.71	0.047	MER/ACB
UO ₃	XRFS	A	3/9/05	82.65	-0.025	MER/ACB
UO ₃	XRFS	A	3/10/05	82.56	-0.134	MER/ACB
UO ₃	XRFS	A	3/10/05	82.46	-0.255	MER/ACB
UO ₃	XRFS	A	3/9/05	82.52	-0.183	MER/ACB
UO ₃	XRFS	A	3/9/05	82.42	-0.304	MER/ACB
UO ₃	XRFS	A	3/10/05	82.67	-0.001	MER/ACB
UO ₃	XRFS	A	3/10/05	82.39	-0.340	MER/ACB
UO ₃	XRFS	A	9/13/05	82.28	-0.473	ACB
UO ₃	XRFS	A	9/13/05	82.21	-0.558	ACB
UO ₃	XRFS	A	9/15/05	82.45	-0.267	ACB
UO ₃	XRFS	A	9/15/05	82.38	-0.352	ACB
UO ₃	XRFS	A	9/13/05	82.24	-0.521	ACB
UO ₃	XRFS	A	9/13/05	82.19	-0.582	ACB
UO ₃	XRFS	A	9/15/05	82.35	-0.388	ACB
UO ₃	XRFS	A	9/15/05	82.55	-0.146	ACB

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
<u>Date</u>	<u>Result</u>					
UF ₆	DG	AB	4/5/06	67.451	-0.247	FP
UF ₆	DG	AB	4/5/06	67.444	-0.257	FP
UF ₆	DG	AB	4/5/06	67.413	-0.303	FP
UF ₆	DG	AB	4/6/06	67.413	-0.303	FP
UF ₆	DG	AB	4/6/06	67.388	-0.340	FP
UF ₆	DG	AB	4/6/06	67.396	-0.328	FP
UF ₆	DG	AB	4/5/06	67.468	-0.222	FP
UF ₆	DG	AB	4/5/06	67.476	-0.210	FP
UF ₆	DG	AB	4/5/06	67.496	-0.181	FP
UF ₆	DG	AB	4/6/06	67.429	-0.280	FP
UF ₆	DG	AB	4/6/06	67.422	-0.290	FP
UF ₆	DG	AB	4/6/06	67.439	-0.265	FP
UF ₆	DG	AB	7/11/07	67.575	-0.064	FP-AV
UF ₆	DG	AB	7/11/07	67.602	-0.024	FP-AV
UF ₆	DG	AB	7/11/07	67.558	-0.089	FP-AV
UF ₆	DG	AB	7/12/07	67.573	-0.067	FP-AV
UF ₆	DG	AB	7/12/07	67.565	-0.079	FP-AV
UF ₆	DG	AB	7/12/07	67.582	-0.053	FP-AV
UF ₆	DG	AB	3/13/08	67.577	-0.061	AV-FP
UF ₆	DG	AB	3/13/08	67.605	-0.019	AV-FP
UF ₆	DG	AB	3/13/08	67.556	-0.092	AV-FP
UF ₆	DG	AB	3/14/08	67.669	0.075	AV-FP
UF ₆	DG	AB	3/14/08	67.547	-0.105	AV-FP
UF ₆	DG	AB	3/14/08	67.602	-0.024	AV-FP
UF ₆	DG	AE	4/5/06	67.431	-0.277	JR-JM
UF ₆	DG	AE	4/5/06	67.414	-0.302	JR-JM
UF ₆	DG	AE	4/5/06	67.663	0.066	JR-JM
UF ₆	DG	AE	4/5/06	67.700	0.121	JR-JM
UF ₆	DG	AE	4/5/06	67.427	-0.283	JR-JM
UF ₆	DG	AE	4/5/06	67.438	-0.266	JR-JM
UF ₆	DG	AE	4/5/06	67.694	0.112	JR-JM
UF ₆	DG	AE	4/5/06	67.698	0.118	JR-JM
UF ₆	DG	AE	11/28/06	67.592	-0.039	JR-EL
UF ₆	DG	AE	11/28/06	67.554	-0.095	JR-EL
UF ₆	DG	AE	11/28/06	67.408	-0.311	JR-EL
UF ₆	DG	AE	11/28/06	67.434	-0.272	JR-EL
UF ₆	DG	AE	11/29/06	67.485	-0.197	JR-EL
UF ₆	DG	AE	11/29/06	67.537	-0.120	JR-EL
UF ₆	DG	AE	11/29/06	67.384	-0.346	JR-EL
UF ₆	DG	AE	11/29/06	67.444	-0.257	JR-EL
UF ₆	DG	AE	10/3/07	67.525	-0.138	JR-JM-EL
UF ₆	DG	AE	10/3/07	67.571	-0.070	JR-JM-EL
UF ₆	DG	AE	10/4/07	67.561	-0.084	JR-JM-EL
UF ₆	DG	AE	10/4/07	67.545	-0.108	JR-JM-EL

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UF ₆	DG	AE	3/17/08	67.56	-0.086	JR-JM-EL
UF ₆	DG	AE	3/17/08	67.551	-0.099	JR-JM-EL
UF ₆	DG	AE	3/18/08	67.537	-0.120	JR-JM-EL
UF ₆	DG	AE	3/18/08	67.553	-0.096	JR-JM-EL
UF ₆	DG	V	3/21/06	67.553	-0.037	1
UF ₆	DG	V	3/21/06	67.625	0.070	1
UF ₆	DG	V	3/21/06	67.592	0.021	1
UF ₆	DG	V	4/5/06	67.516	-0.092	1
UF ₆	DG	V	4/5/06	67.505	-0.108	1
UF ₆	DG	V	4/5/06	67.538	-0.059	1
UF ₆	DG	V	4/25/06	67.583	0.007	1
UF ₆	DG	V	4/25/06	67.605	0.040	1
UF ₆	DG	V	4/25/06	67.607	0.043	1
UF ₆	DG	V	3/21/06	67.569	-0.013	1
UF ₆	DG	V	3/21/06	67.537	-0.061	1
UF ₆	DG	V	3/21/06	67.564	-0.021	1
UF ₆	DG	V	4/5/06	67.530	-0.071	1
UF ₆	DG	V	4/5/06	67.497	-0.120	1
UF ₆	DG	V	4/5/06	67.480	-0.145	1
UF ₆	DG	V	4/25/06	67.543	-0.052	1
UF ₆	DG	V	4/25/06	67.579	0.001	1
UF ₆	DG	V	4/25/06	67.565	-0.019	1
UF ₆	DG	EA	6/13/07	67.58	-0.040	NP2
UF ₆	DG	EA	6/13/07	67.58	-0.040	NP2
UF ₆	DG	EA	6/13/07	67.55	-0.084	NP2
UF ₆	DG	EA	6/13/07	67.52	-0.129	NP2
UF ₆	DG	EA	6/14/07	67.57	-0.055	NP1
UF ₆	DG	EA	6/14/07	67.57	-0.055	NP1
UF ₆	DG	EA	6/14/07	67.53	-0.114	NP1
UF ₆	DG	EA	6/14/07	67.53	-0.114	NP1
UF ₆	DG	AB	4/5/06	67.451	-0.247	FP
UF ₆	DG	AB	4/5/06	67.444	-0.257	FP
UF ₆	DG	AB	4/5/06	67.413	-0.303	FP

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
UF ₆	DG	AB	4/6/06	67.413	-0.303	FP
UF ₆	DG	AB	4/6/06	67.388	-0.340	FP
UF ₆	DG	AB	4/6/06	67.396	-0.328	FP
UF ₆	DG	AB	4/5/06	67.468	-0.222	FP
UF ₆	DG	AB	4/5/06	67.476	-0.210	FP
UF ₆	DG	AB	4/5/06	67.496	-0.181	FP
UF ₆	DG	AB	4/6/06	67.429	-0.280	FP
UF ₆	DG	AB	4/6/06	67.422	-0.290	FP
UF ₆	DG	AB	4/6/06	67.439	-0.265	FP
UF ₆	DG	AE	4/5/06	67.431	-0.277	JR-JM
UF ₆	DG	AE	4/5/06	67.414	-0.302	JR-JM
UF ₆	DG	AE	4/5/06	67.663	0.066	JR-JM
UF ₆	DG	AE	4/5/06	67.700	0.121	JR-JM
UF ₆	DG	AE	4/5/06	67.427	-0.283	JR-JM
UF ₆	DG	AE	4/5/06	67.438	-0.266	JR-JM
UF ₆	DG	AE	4/5/06	67.694	0.112	JR-JM
UF ₆	DG	AE	4/5/06	67.698	0.118	JR-JM
UF ₆	DG	V	3/21/06	67.553	-0.037	1
UF ₆	DG	V	3/21/06	67.625	0.070	1
UF ₆	DG	V	3/21/06	67.592	0.021	1
UF ₆	DG	V	4/5/06	67.516	-0.092	1
UF ₆	DG	V	4/5/06	67.505	-0.108	1
UF ₆	DG	V	4/5/06	67.538	-0.059	1
UF ₆	DG	V	4/25/06	67.583	0.007	1
UF ₆	DG	V	4/25/06	67.605	0.040	1
UF ₆	DG	V	4/25/06	67.607	0.043	1
UF ₆	DG	V	3/21/06	67.569	-0.013	1
UF ₆	DG	V	3/21/06	67.537	-0.061	1
UF ₆	DG	V	3/21/06	67.564	-0.021	1
UF ₆	DG	V	4/5/06	67.530	-0.071	1
UF ₆	DG	V	4/5/06	67.497	-0.120	1
UF ₆	DG	V	4/5/06	67.480	-0.145	1
UF ₆	DG	V	4/25/06	67.543	-0.052	1
UF ₆	DG	V	4/25/06	67.579	0.001	1
UF ₆	DG	V	4/25/06	67.565	-0.019	1

[This page intentionally left blank.]

Appendix B: Uranium Isotopic Results

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
HEU	TIMS	A	11/10/04	89.882	-0.010	BLM/JM
HEU	TIMS	A	11/10/04	89.885	-0.007	BLM/JM
HEU	TIMS	A	11/9/04	89.888	-0.003	WJS/GPW
HEU	TIMS	A	11/9/04	89.888	-0.003	WJS/GPW
HEU	TIMS	A	11/10/04	90.328	-0.010	BLM/JM
HEU	TIMS	A	11/10/04	90.331	-0.007	BLM/JM
HEU	TIMS	A	11/9/04	90.328	-0.010	WJS/GPW
HEU	TIMS	A	11/9/04	90.334	-0.004	WJS/GPW
HEU	TIMS	A	9/14/05	51.334	0.019	WS/JM
HEU	TIMS	A	9/14/05	51.330	0.011	WS/JM
HEU	TIMS	A	9/19/05	51.303	-0.042	WS/JM
HEU	TIMS	A	9/19/05	51.313	-0.022	WS/JM
HEU	TIMS	A	9/14/05	89.702	0.026	WS/JM
HEU	TIMS	A	9/14/05	89.697	0.020	WS/JM
HEU	TIMS	A	9/19/05	89.670	-0.010	WS/JM
HEU	TIMS	A	9/19/05	89.669	-0.011	WS/JM
HEU	TIMS	A	7/17/07	51.333	0.017	BLM/PRR
HEU	TIMS	A	7/18/07	51.305	-0.038	BLM/PRR
HEU	TIMS	A	7/18/07	51.325	0.001	BLM/PRR
HEU	TIMS	A	7/17/07	51.333	0.017	BLM/PRR
HEU	TIMS	A	7/17/07	51.336	0.022	BLM/PRR
HEU	TIMS	A	7/17/07	51.337	0.024	BLM/PRR
HEU	TIMS	A	7/17/07	51.325	0.001	BLM/PRR
HEU	TIMS	A	7/17/07	51.343	0.036	BLM/PRR
HEU	TIMS	B	1/8/05	89.6706	-0.009	JLB
HEU	TIMS	B	1/8/05	89.6709	-0.009	JLB
HEU	TIMS	B	1/10/05	89.6868	0.009	MDM
HEU	TIMS	B	1/10/05	89.6781	-0.001	MDM
HEU	TIMS	B	1/8/05	90.3376	0.000	JLB
HEU	TIMS	B	1/8/05	90.3453	0.009	JLB
HEU	TIMS	B	1/10/05	90.3425	0.006	MDM
HEU	TIMS	B	1/10/05	90.3469	0.011	MDM
HEU	TIMS	B	12/14/05	51.3445	0.039	PAM
HEU	TIMS	B	12/14/05	51.3451	0.040	PAM
HEU	TIMS	B	12/15/05	51.3140	-0.020	TEB
HEU	TIMS	B	12/15/05	51.3449	0.040	TEB
HEU	TIMS	B	12/14/05	89.8980	0.008	PAM
HEU	TIMS	B	12/14/05	89.8822	-0.010	PAM
HEU	TIMS	B	12/15/05	89.8965	0.006	TEB
HEU	TIMS	B	12/15/05	89.8546	-0.041	TEB
HEU	TIMS	B	1/29/06	51.3255	0.002	PAM
HEU	TIMS	B	1/29/06	51.3245	0.000	PAM
HEU	TIMS	B	3/3/06	51.3254	0.002	DDN
HEU	TIMS	B	3/3/06	51.3287	0.008	DDN
HEU	TIMS	B	1/29/06	89.9017	0.012	PAM
HEU	TIMS	B	1/29/06	89.8946	0.004	PAM
HEU	TIMS	B	3/3/06	89.8888	-0.003	DDN

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		<u>Analyst</u>
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
HEU	TIMS	B	3/3/06	89.8987	0.008	DDN
HEU	TIMS	B	3/28/06	89.6567	-0.025	DDB
HEU	TIMS	B	3/28/06	89.6820	0.004	DDB
HEU	TIMS	B	4/5/06	89.6910	0.014	DDN
HEU	TIMS	B	4/5/06	89.6988	0.022	DDN
HEU	TIMS	B	10/11/06	90.3263	-0.012	LLB
HEU	TIMS	B	10/11/06	90.334	-0.004	LLB
HEU	TIMS	B	7/25/06	90.346	0.010	PAM
HEU	TIMS	B	7/25/06	90.3431	0.007	PAM
HEU	TIMS	B	1/17/08	90.3459	0.010	DDN
HEU	TIMS	B	1/17/08	90.3436	0.007	DDN
HEU	TIMS	B	1/26/08	90.3437	0.007	JCP
HEU	TIMS	B	1/26/08	90.3456	0.009	JCP
HEU	TIMS	B	1/17/08	90.3518	0.016	DDN
HEU	TIMS	B	1/17/08	90.3512	0.016	DDN
HEU	TIMS	B	1/26/08	90.3521	0.017	JCP
HEU	TIMS	B	1/26/08	90.3514	0.016	JCP
HEU	TIMS	B	3/25/08	51.3613	0.072	PAM
HEU	TIMS	B	3/25/08	51.3608	0.071	PAM
HEU	TIMS	B	3/27/08	51.3613	0.072	LLB
HEU	TIMS	B	3/27/08	51.3617	0.072	LLB
HEU	TIMS	B	3/25/08	51.3475	0.045	PAM
HEU	TIMS	B	3/25/08	51.3471	0.044	PAM
HEU	TIMS	B	3/27/08	51.3492	0.048	LLB
HEU	TIMS	B	3/27/08	51.3491	0.048	LLB
HEU	TIMS	F	1/12/07	89.68024	0.002	RE
HEU	TIMS	F	1/12/07	89.68079	0.002	RE
HEU	TIMS	F	1/13/07	89.68045	0.002	RE
HEU	TIMS	F	1/13/07	89.68072	0.002	RE
HEU	TIMS	F	1/12/07	90.33964	0.003	RE
HEU	TIMS	F	1/12/07	90.33819	0.001	RE
HEU	TIMS	F	1/13/07	90.34	0.003	RE
HEU	TIMS	F	1/13/07	90.33858	0.002	RE
HEU	TIMS	F	3/13/07	51.3386	0.027	RBT
HEU	TIMS	F	3/13/07	51.3376	0.026	RBT
HEU	TIMS	F	3/14/07	51.3381	0.026	RBT
HEU	TIMS	F	3/14/07	51.3379	0.026	RBT
HEU	TIMS	F	7/11/07	89.901	0.011	KJM
HEU	TIMS	F	7/11/07	89.894	0.003	KJM
HEU	TIMS	F	7/12/07	89.893	0.002	KJM
HEU	TIMS	F	7/12/07	89.898	0.008	KJM
HEU	TIMS	F	7/11/07	90.3397	0.003	KJM
HEU	TIMS	F	7/12/07	90.3406	0.004	KJM
HEU	TIMS	F	7/12/07	90.3416	0.005	KJM

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
HEU	TIMS	G	5/3/06	89.6817	0.003	
HEU	TIMS	G	5/3/06	89.6811	0.003	
HEU	TIMS	G	5/3/06	89.8908	0.000	
HEU	TIMS	G	5/3/06	89.8908	0.000	
HEU	TIMS	G	5/3/06	90.3402	0.003	
HEU	TIMS	G	5/3/06	90.3407	0.004	
HEU	TIMS	G	5/3/06	90.3356	-0.002	
HEU	TIMS	G	5/3/06	90.3351	-0.002	
HEU	TIMS	G	5/5/06	51.3442	0.038	
HEU	TIMS	G	5/5/06	51.3458	0.042	
HEU	TIMS	G	5/8/06	51.3435	0.037	
HEU	TIMS	G	5/8/06	51.3457	0.041	
HEU	TIMS	AW	5/21/07	51.3215	-0.006	
HEU	TIMS	AW	5/21/07	51.3436	0.037	
HEU	TIMS	AW	6/4/07	51.3488	0.047	
HEU	TIMS	AW	5/21/07	51.3366	0.024	
HEU	TIMS	AW	5/21/07	51.3494	0.049	
HEU	TIMS	AW	6/4/07	51.3295	0.010	
HEU	TIMS	AW	5/21/07	89.6919	0.015	
HEU	TIMS	AW	6/1/07	89.6825	0.004	
HEU	TIMS	AW	6/4/07	89.6629	-0.018	
HEU	TIMS	AW	5/21/07	89.6893	0.012	
HEU	TIMS	AW	6/4/07	89.6753	-0.004	
HEU	TIMS	AW	6/4/07	89.6681	-0.012	
HEU	TIMS	W	9/5/06	90.3421	0.005	Kuehn
HEU	TIMS	W	9/6/06	90.3425	0.006	Kuehn
HEU	TIMS	W	9/7/06	90.3431	0.007	Kuehn
HEU	TIMS	W	9/8/06	90.343	0.006	Kuehn
HEU	TIMS	W	9/5/06	90.3435	0.007	Kuehn
HEU	TIMS	W	9/6/06	90.3415	0.005	Kuehn
HEU	TIMS	W	9/7/06	90.3425	0.006	Kuehn
HEU	TIMS	W	9/8/06	90.3431	0.007	Kuehn

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		<u>Analyst</u>
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	AA	4/5/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/5/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/6/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/6/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/6/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/20/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/20/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/21/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/21/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/21/06	4.0051	-0.078	ALEG
LEU	TIMS	AA	4/20/06	0.7100	-0.125	ALEG
LEU	TIMS	AA	4/20/06	0.7119	0.142	ALEG
LEU	TIMS	AA	4/21/06	0.7117	0.114	ALEG
LEU	TIMS	AA	4/21/06	0.7127	0.254	ALEG
LEU	TIMS	AA	4/21/06	0.7107	-0.027	ALEG
LEU	TIMS	AA	4/20/06	0.7109	0.001	ALEG
LEU	TIMS	AA	4/20/06	0.7109	0.001	ALEG
LEU	TIMS	AA	4/21/06	0.7127	0.254	ALEG
LEU	TIMS	AA	4/21/06	0.7088	-0.294	ALEG
LEU	TIMS	AA	4/21/06	0.7107	-0.027	ALEG
LEU	TIMS	AA	11/27/06	0.7119	0.142	AL/EG
LEU	TIMS	AA	11/27/06	0.711	0.015	AL/EG
LEU	TIMS	AA	11/28/06	0.7095	-0.196	AL/EG
LEU	TIMS	AA	11/28/06	0.7111	0.029	AL/EG
LEU	TIMS	AA	11/27/06	0.7125	0.226	AL/EG
LEU	TIMS	AA	11/27/06	0.7131	0.311	AL/EG
LEU	TIMS	AA	11/28/06	0.7123	0.198	AL/EG
LEU	TIMS	AA	11/28/06	0.7127	0.254	AL/EG
LEU	TIMS	AA	10/4/07	1.3046	0.184	MLG
LEU	TIMS	AA	10/4/07	1.3049	0.207	MLG
LEU	TIMS	AA	10/4/07	1.3013	-0.069	MLG
LEU	TIMS	AA	10/17/07	1.3018	-0.031	MLG
LEU	TIMS	AA	10/17/07	1.3026	0.031	MLG
LEU	TIMS	AA	10/5/07	1.3053	0.238	MLG
LEU	TIMS	AA	10/5/07	1.302	-0.015	MLG
LEU	TIMS	AA	10/5/07	1.3036	0.108	MLG
LEU	TIMS	AA	10/17/07	1.3029	0.054	MLG
LEU	TIMS	AA	10/17/07	1.3031	0.069	MLG
LEU	TIMS	AA	10/5/07	4.3968	0.120	MLG
LEU	TIMS	AA	10/5/07	4.3902	-0.031	MLG
LEU	TIMS	AA	10/17/07	4.3875	-0.092	MLG
LEU	TIMS	AA	10/17/07	4.3868	-0.108	MLG
LEU	TIMS	AA	10/17/07	4.3867	-0.110	MLG
LEU	TIMS	AA	10/8/07	4.3904	-0.026	MLG
LEU	TIMS	AA	10/8/07	4.39	-0.035	MLG
LEU	TIMS	AA	10/8/07	4.3939	0.054	MLG

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	AA	10/17/07	4.3869	-0.106	MLG
LEU	TIMS	AA	10/17/07	4.3876	-0.090	MLG
LEU	TIMS	AA	12/3/07	4.4514	-0.190	MLG
LEU	TIMS	AA	12/3/07	4.4577	-0.048	MLG
LEU	TIMS	AA	12/3/07	4.4523	-0.170	MLG
LEU	TIMS	AA	12/7/07	4.4436	-0.365	MLG
LEU	TIMS	AA	12/7/07	4.4483	-0.259	MLG
LEU	TIMS	AA	12/3/07	4.4575	-0.053	MLG
LEU	TIMS	AA	12/3/07	4.4528	-0.158	MLG
LEU	TIMS	AA	12/3/07	4.4587	-0.026	MLG
LEU	TIMS	AA	12/7/07	4.4434	-0.369	MLG
LEU	TIMS	AA	12/7/07	4.4453	-0.327	MLG
LEU	TIMS	AA	11/30/06	4.0051	-0.078	AL/EG
LEU	TIMS	AA	11/30/06	4.0027	-0.138	AL/EG
LEU	TIMS	AA	12/1/06	4.0058	-0.061	AL/EG
LEU	TIMS	AA	12/1/06	4.0009	-0.183	AL/EG
LEU	TIMS	AA	11/30/06	4.0026	-0.140	AL/EG
LEU	TIMS	AA	11/30/06	4.0042	-0.101	AL/EG
LEU	TIMS	AA	12/1/06	4.0012	-0.175	AL/EG
LEU	TIMS	AA	12/1/06	4.0023	-0.148	AL/EG
LEU	TIMS	AA	3/13/08	2.9862	0.064	MLG
LEU	TIMS	AA	3/13/08	2.9848	0.017	MLG
LEU	TIMS	AA	3/13/08	2.9867	0.080	MLG
LEU	TIMS	AA	3/18/08	2.9853	0.034	MLG
LEU	TIMS	AA	3/18/08	2.9832	-0.037	MLG
LEU	TIMS	AA	3/13/08	2.9878	0.117	MLG
LEU	TIMS	AA	3/13/08	2.9844	0.003	MLG
LEU	TIMS	AA	3/13/08	2.9847	0.013	MLG
LEU	TIMS	AA	3/18/08	2.9882	0.131	MLG
LEU	TIMS	AA	3/18/08	2.9864	0.070	MLG
LEU	TIMS	BC	4/17/06	4.0046	-0.091	MRPP
LEU	TIMS	BC	4/17/06	4.0090	0.019	MRPP
LEU	TIMS	BC	4/26/06	4.0067	-0.038	MRPP
LEU	TIMS	BC	4/26/06	4.0074	-0.021	MRPP
LEU	TIMS	BC	4/17/06	4.0066	-0.041	MRPP
LEU	TIMS	BC	4/17/06	4.0126	0.109	MRPP
LEU	TIMS	BC	4/26/06	4.0077	-0.013	MRPP
LEU	TIMS	BC	4/26/06	4.0096	0.034	MRPP
LEU	TIMS	BC	10/9/06	4.0101	0.047	MRPP
LEU	TIMS	BC	10/9/06	4.0095	0.032	MRPP
LEU	TIMS	BC	10/16/06	4.0122	0.099	MRPP
LEU	TIMS	BC	10/16/06	4.0083	0.002	MRPP
LEU	TIMS	BC	10/9/06	4.0054	-0.071	MRPP
LEU	TIMS	BC	10/9/06	4.0138	0.139	MRPP
LEU	TIMS	BC	10/16/06	4.0128	0.114	MRPP
LEU	TIMS	BC	10/16/06	4.0031	-0.128	MRPP

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	A	9/26/07	4.3921	0.013	MB/JM
LEU	TIMS	A	9/27/07	4.3913	-0.005	MB/JM
LEU	TIMS	A	9/27/07	4.3946	0.070	MB/JM
LEU	TIMS	A	9/26/07	4.3895	-0.046	MB/JM
LEU	TIMS	A	9/12/07	4.3916	0.001	MB/JM
LEU	TIMS	A	9/11/07	4.3932	0.038	MB/JM
LEU	TIMS	A	9/11/07	4.3901	-0.033	MB/JM
LEU	TIMS	A	9/12/07	4.3989	0.168	MB/JM
LEU	TIMS	AW	5/16/07	4.3916	0.001	
LEU	TIMS	AW	5/21/07	4.4032	0.266	
LEU	TIMS	AW	6/4/07	4.3947	0.072	
LEU	TIMS	B	1/9/06	0.7099	-0.214	PAM
LEU	TIMS	B	1/9/06	0.7110	-0.060	PAM
LEU	TIMS	B	1/12/06	0.7117	0.039	DDN
LEU	TIMS	B	1/12/06	0.7111	-0.046	DDN
LEU	TIMS	B	1/9/06	0.7111	-0.046	PAM
LEU	TIMS	B	1/9/06	0.7121	0.095	PAM
LEU	TIMS	B	1/12/06	0.7103	-0.158	DDN
LEU	TIMS	B	1/12/06	0.7088	-0.369	DDN
LEU	TIMS	B	3/2/06	0.8835	1.439	PAM
LEU	TIMS	B	3/2/06	0.8820	1.267	PAM
LEU	TIMS	B	3/9/06	0.8811	1.164	DDN
LEU	TIMS	B	3/9/06	0.8815	1.210	DDN
LEU	TIMS	B	3/2/06	0.8832	1.405	PAM
LEU	TIMS	B	3/2/06	0.8823	1.302	PAM
LEU	TIMS	B	3/9/06	0.8821	1.279	DDN
LEU	TIMS	B	3/9/06	0.8816	1.221	DDN
LEU	TIMS	B	3/4/06	0.7094	-0.285	WPB
LEU	TIMS	B	3/4/06	0.7119	0.067	WPB
LEU	TIMS	B	3/5/06	0.7113	-0.017	DDN
LEU	TIMS	B	3/5/06	0.7114	-0.003	DDN
LEU	TIMS	B	3/4/06	0.7110	-0.060	WPB
LEU	TIMS	B	3/4/06	0.7102	-0.172	WPB
LEU	TIMS	B	3/5/06	0.7117	0.039	DDN
LEU	TIMS	B	3/5/06	0.7104	-0.144	DDN
LEU	TIMS	B	3/28/06	4.3860	-0.126	DDB
LEU	TIMS	B	3/28/06	4.3904	-0.026	DDB
LEU	TIMS	B	4/5/06	4.3918	0.006	DDN
LEU	TIMS	B	4/5/06	4.3965	0.113	DDN
LEU	TIMS	B	4/12/06	0.7137	0.320	JMG
LEU	TIMS	B	4/12/06	0.7127	0.179	JMG
LEU	TIMS	B	4/13/06	0.7114	-0.003	CPT
LEU	TIMS	B	4/13/06	0.7116	0.025	CPT
LEU	TIMS	B	4/12/06	0.7131	0.236	JMG
LEU	TIMS	B	4/12/06	0.7109	-0.074	JMG

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	B	4/13/06	0.7114	-0.003	CPT
LEU	TIMS	B	4/13/06	0.7111	-0.046	CPT
LEU	TIMS	B	7/22/06	0.7133	0.264	PWB
LEU	TIMS	B	7/22/06	0.7122	0.109	PWB
LEU	TIMS	B	7/31/06	0.7121	0.095	DDB
LEU	TIMS	B	7/31/06	0.7127	0.179	DDB
LEU	TIMS	B	7/22/06	0.7122	0.109	PWB
LEU	TIMS	B	7/22/06	0.7135	0.292	PWB
LEU	TIMS	B	8/1/06	0.7126	0.165	DDB
LEU	TIMS	B	8/1/06	0.7138	0.334	DDB
LEU	TIMS	B	7/25/06	4.3886	-0.067	PAM
LEU	TIMS	B	7/25/06	4.3806	-0.249	PAM
LEU	TIMS	B	10/11/06	4.3845	-0.160	LLB
LEU	TIMS	B	10/11/06	4.3855	-0.138	LLB
LEU	TIMS	B	1/18/08	0.7124	0.234	DDN
LEU	TIMS	B	1/18/08	0.7121	0.192	DDN
LEU	TIMS	B	1/26/08	0.7122	0.206	JCP
LEU	TIMS	B	1/26/08	0.7126	0.263	JCP
LEU	TIMS	B	1/18/08	0.7124	0.234	DDN
LEU	TIMS	B	1/18/08	0.712	0.178	DDN
LEU	TIMS	B	1/26/08	0.7118	0.150	JCP
LEU	TIMS	B	1/26/08	0.7121	0.192	JCP
LEU	TIMS	B	3/25/08	0.7115	0.108	PAM
LEU	TIMS	B	3/25/08	0.7122	0.206	PAM
LEU	TIMS	B	3/27/08	0.7101	-0.089	LLB
LEU	TIMS	B	3/27/08	0.7116	0.122	LLB
LEU	TIMS	B	3/25/08	0.7118	0.150	PAM
LEU	TIMS	B	3/25/08	0.7116	0.122	PAM
LEU	TIMS	B	3/27/08	0.7122	0.206	LLB
LEU	TIMS	B	3/27/08	0.712	0.178	LLB
LEU	TIMS	F	3/13/07	4.39049	-0.024	RBT
LEU	TIMS	F	3/13/07	4.39055	-0.023	RBT
LEU	TIMS	F	3/14/07	4.39122	-0.007	RBT
LEU	TIMS	F	3/14/07	4.39085	-0.016	RBT
LEU	TIMS	F	9/19/07	4.45874	-0.025	RE
LEU	TIMS	F	9/19/07	4.46042	0.012	RE
LEU	TIMS	F	9/20/07	4.46146	0.036	RBT
LEU	TIMS	F	9/20/07	4.461	0.025	RBT
LEU	TIMS	F	9/19/07	4.45999	0.003	RE
LEU	TIMS	F	9/19/07	4.45923	-0.014	RE
LEU	TIMS	F	9/20/07	4.45739	-0.055	RBT
LEU	TIMS	F	9/20/07	4.46096	0.025	RBT
LEU	TIMS	F	9/19/07	4.00767	-0.014	RE
LEU	TIMS	F	9/19/07	4.00685	-0.034	RE
LEU	TIMS	F	9/20/07	4.00736	-0.022	RBT

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	F	9/20/07	4.00782	-0.010	RBT
LEU	TIMS	F	9/19/07	4.00708	-0.029	RE
LEU	TIMS	F	9/19/07	4.00727	-0.024	RE
LEU	TIMS	F	9/20/07	4.00681	-0.035	RBT
LEU	TIMS	F	9/20/07	4.00692	-0.033	RBT
LEU	TIMS	F	9/19/07	4.00673	-0.037	RE
LEU	TIMS	F	9/19/07	4.0067	-0.038	RE
LEU	TIMS	F	9/20/07	4.0086	0.009	RBT
LEU	TIMS	F	9/20/07	4.0081	-0.003	RBT
LEU	TIMS	F	9/19/07	4.0066	-0.041	RE
LEU	TIMS	F	9/19/07	4.00743	-0.020	RE
LEU	TIMS	F	9/20/07	4.00948	0.031	RBT
LEU	TIMS	F	9/20/07	4.00857	0.008	RBT
LEU	TIMS	F	11/21/07	0.88138	-0.070	KJM
LEU	TIMS	F	11/21/07	0.88115	-0.096	KJM
LEU	TIMS	F	11/23/07	0.88135	-0.074	KJM
LEU	TIMS	F	11/23/07	0.88088	-0.127	KJM
LEU	TIMS	F	11/21/07	0.88192	-0.009	KJM
LEU	TIMS	F	11/21/07	0.88144	-0.063	KJM
LEU	TIMS	F	11/23/07	0.88141	-0.067	KJM
LEU	TIMS	F	11/23/07	0.88148	-0.059	KJM
LEU	TIMS	G	5/8/06	4.3939	0.054	
LEU	TIMS	G	5/8/06	4.3948	0.074	
LEU	TIMS	G	5/8/06	4.3928	0.029	
LEU	TIMS	G	5/8/06	4.3948	0.074	
LEU	TIMS	T	1/12/06	4.010	0.044	
LEU	TIMS	T	1/12/06	4.009	0.019	
LEU	TIMS	T	1/12/06	4.011	0.069	
LEU	TIMS	T	1/12/06	4.009	0.019	
LEU	TIMS	T	1/19/06	4.009	0.019	
LEU	TIMS	T	1/19/06	4.009	0.019	
LEU	TIMS	T	1/19/06	4.010	0.044	
LEU	TIMS	T	1/19/06	4.009	0.019	
LEU	TIMS	T	4/21/06	4.010	0.044	
LEU	TIMS	T	4/21/06	4.011	0.069	
LEU	TIMS	T	4/21/06	4.010	0.044	
LEU	TIMS	T	4/21/06	4.012	0.094	
LEU	TIMS	T	4/24/06	4.009	0.019	
LEU	TIMS	T	4/24/06	4.009	0.019	
LEU	TIMS	T	4/24/06	4.009	0.019	
LEU	TIMS	T	4/24/06	4.008	-0.006	
LEU	TIMS	T	8/28/06	4.012	0.094	
LEU	TIMS	T	8/28/06	4.01	0.044	
LEU	TIMS	T	8/28/06	4.01	0.044	
LEU	TIMS	T	8/28/06	4.009	0.019	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	T	8/29/06	4.01	0.044	
LEU	TIMS	T	8/29/06	4.01	0.044	
LEU	TIMS	T	8/29/06	4.012	0.094	
LEU	TIMS	T	8/29/06	4.009	0.019	
LEU	TIMS	T	10/17/06	4.01	0.044	
LEU	TIMS	T	10/17/06	4.009	0.019	
LEU	TIMS	T	10/17/06	4.009	0.019	
LEU	TIMS	T	10/17/06	4.01	0.044	
LEU	TIMS	T	10/19/06	4.01	0.044	
LEU	TIMS	T	10/19/06	4.008	-0.006	
LEU	TIMS	T	10/19/06	4.009	0.019	
LEU	TIMS	T	10/19/06	4.01	0.044	
LEU	TIMS	T	4/26/07	4.01	0.044	
LEU	TIMS	T	4/26/07	4.009	0.019	
LEU	TIMS	T	4/26/07	4.008	-0.006	
LEU	TIMS	T	4/26/07	4.008	-0.006	
LEU	TIMS	T	5/16/07	4.011	0.069	
LEU	TIMS	T	5/16/07	4.009	0.019	
LEU	TIMS	T	5/16/07	4.009	0.019	
LEU	TIMS	T	5/16/07	4.009	0.019	
LEU	TIMS	T	11/29/07	4.009	0.019	
LEU	TIMS	T	11/29/07	4.009	0.019	
LEU	TIMS	T	11/29/07	4.012	0.094	
LEU	TIMS	T	11/29/07	4.008	-0.006	
LEU	TIMS	T	12/11/07	4.009	0.019	
LEU	TIMS	T	12/11/07	4.009	0.019	
LEU	TIMS	T	12/11/07	4.01	0.044	
LEU	TIMS	T	12/11/07	4.01	0.044	
LEU	TIMS	V	3/24/06	4.3881	-0.078	5
LEU	TIMS	V	4/3/06	4.3943	0.063	5
LEU	TIMS	V	4/3/06	4.3922	0.015	5
LEU	TIMS	V	4/3/06	4.3913	-0.005	5
LEU	TIMS	V	4/25/06	4.3920	0.010	5
LEU	TIMS	V	4/25/06	4.3917	0.004	5
LEU	TIMS	V	4/25/06	4.3922	0.015	5
LEU	TIMS	V	3/24/06	4.3888	-0.062	5
LEU	TIMS	V	3/24/06	4.3898	-0.040	5
LEU	TIMS	V	3/24/06	4.3913	-0.005	5
LEU	TIMS	V	4/3/06	4.3937	0.049	5
LEU	TIMS	V	4/3/06	4.3934	0.042	5
LEU	TIMS	V	4/3/06	4.3938	0.051	5
LEU	TIMS	V	4/25/06	4.3932	0.038	5
LEU	TIMS	V	4/25/06	4.3922	0.015	5
LEU	TIMS	V	4/25/06	4.3922	0.015	5
LEU	TIMS	V	3/24/06	4.4581	-0.040	5
LEU	TIMS	V	3/24/06	4.4567	-0.071	5
LEU	TIMS	V	3/24/06	4.4568	-0.069	5

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis Date</u>	<u>Reported Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	TIMS	V	4/3/06	4.4654	0.124	5
LEU	TIMS	V	4/4/06	4.4630	0.070	5
LEU	TIMS	V	4/4/06	4.4625	0.059	5
LEU	TIMS	V	4/25/06	4.4606	0.017	5
LEU	TIMS	V	4/25/06	4.4612	0.030	5
LEU	TIMS	V	4/25/06	4.4599	0.001	5
LEU	TIMS	V	3/24/06	4.4580	-0.042	5
LEU	TIMS	V	3/24/06	4.4596	-0.006	5
LEU	TIMS	V	3/24/06	4.4611	0.028	5
LEU	TIMS	V	4/4/06	4.4566	-0.073	5
LEU	TIMS	V	4/4/06	4.4598	-0.001	5
LEU	TIMS	V	4/4/06	4.4604	0.012	5
LEU	TIMS	V	4/26/06	4.4608	0.021	5
LEU	TIMS	V	4/26/06	4.4616	0.039	5
LEU	TIMS	V	4/26/06	4.4612	0.030	5
LEU	TIMS	V	3/21/06	4.0057	-0.063	5
LEU	TIMS	V	3/21/06	4.0062	-0.051	5
LEU	TIMS	V	3/21/06	4.0064	-0.046	5
LEU	TIMS	V	4/5/06	4.0099	0.042	5
LEU	TIMS	V	4/5/06	4.0078	-0.011	5
LEU	TIMS	V	4/5/06	4.0080	-0.006	5
LEU	TIMS	V	4/26/06	4.0088	0.014	5
LEU	TIMS	V	4/26/06	4.0100	0.044	5
LEU	TIMS	V	4/26/06	4.0091	0.022	5
LEU	TIMS	V	3/21/06	4.0110	0.069	5
LEU	TIMS	V	3/21/06	4.0082	-0.001	5
LEU	TIMS	V	3/23/06	4.0040	-0.106	5
LEU	TIMS	V	4/5/06	4.0096	0.034	5
LEU	TIMS	V	4/5/06	4.0025	-0.143	5
LEU	TIMS	V	4/7/06	4.0079	-0.008	5
LEU	TIMS	V	4/26/06	4.0096	0.034	5
LEU	TIMS	V	4/26/06	4.0093	0.027	5
LEU	TIMS	V	4/26/06	4.0100	0.044	5
LEU	TIMS	W	9/1/06	4.3933	0.040	Kuehn
LEU	TIMS	W	9/2/06	4.3926	0.024	Kuehn
LEU	TIMS	W	9/3/06	4.392	0.010	Kuehn
LEU	TIMS	W	9/4/06	4.3916	0.001	Kuehn
LEU	TIMS	W	9/1/06	4.392	0.010	Kuehn
LEU	TIMS	W	9/2/06	4.3908	-0.017	Kuehn
LEU	TIMS	W	9/4/06	4.3921	0.013	Kuehn

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	ICPMS	AF	6/5/06	3.995	-0.330	SR
LEU	ICPMS	AF	6/5/06	3.997	-0.280	SR
LEU	ICPMS	AF	6/5/06	3.982	-0.654	SR
LEU	ICPMS	AF	6/5/06	3.969	-0.979	SR
LEU	ICPMS	AF	6/6/06	3.991	-0.430	SR
LEU	ICPMS	AF	6/6/06	4.007	-0.031	SR
LEU	ICPMS	AF	6/6/06	4.003	-0.130	SR
LEU	ICPMS	AF	6/6/06	4.000	-0.205	SR
LEU	ICPMS	AF	6/5/06	4.005	-0.081	SR
LEU	ICPMS	AF	6/5/06	3.981	-0.679	SR
LEU	ICPMS	AF	6/5/06	3.984	-0.605	SR
LEU	ICPMS	AF	6/5/06	3.979	-0.729	SR
LEU	ICPMS	AF	6/6/06	3.985	-0.580	SR
LEU	ICPMS	AF	6/6/06	3.99	-0.455	SR
LEU	ICPMS	AF	6/6/06	4	-0.205	SR
LEU	ICPMS	AF	6/6/06	3.982	-0.654	SR
LEU	ICPMS	BE	4/7/06	4.067	1.466	MHK
LEU	ICPMS	BE	4/7/06	4.072	1.591	MHK
LEU	ICPMS	BE	4/7/06	4.069	1.516	MHK
LEU	ICPMS	BE	4/7/06	4.063	1.366	MHK
LEU	ICPMS	BE	4/7/06	4.059	1.267	MHK
LEU	ICPMS	BE	4/7/06	4.058	1.242	MHK
LEU	ICPMS	BE	4/20/06	4.030	0.543	MHK
LEU	ICPMS	BE	4/20/06	4.024	0.393	MHK
LEU	ICPMS	BE	4/20/06	4.030	0.543	MHK
LEU	ICPMS	BE	4/20/06	4.035	0.668	MHK
LEU	ICPMS	BE	4/20/06	4.029	0.518	MHK
LEU	ICPMS	BE	4/20/06	4.034	0.643	MHK
LEU	ICPMS	BE	12/12/07	4.372	-1.970	MHK
LEU	ICPMS	BE	12/12/07	4.373	-1.948	MHK
LEU	ICPMS	BE	12/12/07	4.386	-1.656	MHK
LEU	ICPMS	BE	12/12/07	4.373	-1.948	MHK
LEU	ICPMS	BE	12/12/07	4.377	-1.858	MHK
LEU	ICPMS	BE	12/12/07	4.378	-1.836	MHK
LEU	ICPMS	BE	12/12/07	4.377	-1.858	MHK
LEU	ICPMS	BE	12/12/07	4.365	-2.127	MHK
LEU	ICPMS	BE	10/25/07	4.334	-2.822	MHK
LEU	ICPMS	BE	10/25/07	4.34	-2.688	MHK
LEU	ICPMS	BE	10/25/07	4.328	-2.957	MHK
LEU	ICPMS	BE	10/25/07	4.342	-2.643	MHK
LEU	ICPMS	BE	10/25/07	4.382	-1.746	MHK

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
LEU	ICPMS	BE	10/25/07	4.387	-1.634	MHK
LEU	ICPMS	BE	10/25/07	4.361	-2.217	MHK
LEU	ICPMS	BE	10/25/07	4.358	-2.284	MHK
LEU	ICPMS	BE	10/3/06	3.989	-0.480	MHK
LEU	ICPMS	BE	10/3/06	3.987	-0.530	MHK
LEU	ICPMS	BE	10/3/06	3.979	-0.729	MHK
LEU	ICPMS	BE	10/3/06	3.983	-0.629	MHK
LEU	ICPMS	BE	10/3/06	3.993	-0.380	MHK
LEU	ICPMS	BE	10/3/06	3.984	-0.605	MHK
LEU	ICPMS	BE	10/3/06	3.98	-0.704	MHK
LEU	ICPMS	BE	10/3/06	3.99	-0.455	MHK
LEU	ICPMS	BE	10/4/06	4.021	0.319	MHK
LEU	ICPMS	BE	10/4/06	4	-0.205	MHK
LEU	ICPMS	BE	10/4/06	4.002	-0.155	MHK
LEU	ICPMS	BE	10/4/06	3.996	-0.305	MHK
LEU	ICPMS	BE	10/4/06	3.994	-0.355	MHK
LEU	ICPMS	BE	10/4/06	3.984	-0.605	MHK
LEU	ICPMS	BE	10/4/06	4.007	-0.031	MHK
LEU	ICPMS	BE	10/4/06	3.995	-0.330	MHK
LEU	ICPMS	BE	10/25/07	3.956	-1.303	MHK
LEU	ICPMS	BE	10/25/07	3.95	-1.453	MHK
LEU	ICPMS	BE	10/25/07	3.957	-1.278	MHK
LEU	ICPMS	BE	10/25/07	3.962	-1.153	MHK
LEU	ICPMS	BE	10/25/07	3.937	-1.777	MHK
LEU	ICPMS	BE	10/25/07	3.935	-1.827	MHK
LEU	ICPMS	BE	10/25/07	3.931	-1.927	MHK
LEU	ICPMS	BE	10/25/07	3.927	-2.027	MHK
LEU	ICPMS	BE	12/12/07	3.91	-2.451	MHK
LEU	ICPMS	BE	12/12/07	3.873	-3.374	MHK
LEU	ICPMS	BE	12/12/07	3.868	-3.499	MHK
LEU	ICPMS	BE	12/12/07	3.872	-3.399	MHK
LEU	ICPMS	BE	12/12/07	3.949	-1.478	MHK
LEU	ICPMS	BE	12/12/07	3.898	-2.750	MHK
LEU	ICPMS	BE	12/12/07	3.874	-3.349	MHK
LEU	ICPMS	BE	12/12/07	3.848	-3.998	MHK
LEU	ICPMS	BE	2/4/08	4.494	0.765	MHK
LEU	ICPMS	BE	2/4/08	4.456	-0.087	MHK
LEU	ICPMS	BE	2/4/08	4.481	0.474	MHK
LEU	ICPMS	BE	2/4/08	4.451	-0.199	MHK
LEU	ICPMS	BE	2/4/08	4.438	-0.490	MHK
LEU	ICPMS	BE	2/4/08	4.427	-0.737	MHK
LEU	ICPMS	BE	2/4/08	4.447	-0.288	MHK
LEU	ICPMS	BE	2/4/08	4.456	-0.087	MHK
LEU	ICPMS	BE	3/4/08	4.461	0.025	MHK
LEU	ICPMS	BE	3/4/08	4.465	0.115	MHK
LEU	ICPMS	BE	3/4/08	4.467	0.160	MHK
LEU	ICPMS	BE	3/4/08	4.495	0.788	MHK

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>	<u>% RD</u>	<u>Analyst</u>
LEU	ICPMS	BE	3/4/08	4.505	1.012	MHK
LEU	ICPMS	BE	3/4/08	4.48	0.452	MHK
LEU	ICPMS	BE	3/4/08	4.449	-0.244	MHK
LEU	ICPMS	BE	3/4/08	4.47	0.227	MHK
LEU	ICPMS	BE	2/4/08	3.985	-0.580	MHK
LEU	ICPMS	BE	2/4/08	3.991	-0.430	MHK
LEU	ICPMS	BE	2/4/08	3.981	-0.679	MHK
LEU	ICPMS	BE	2/4/08	4.003	-0.130	MHK
LEU	ICPMS	BE	3/4/08	3.982	-0.654	MHK
LEU	ICPMS	BE	3/4/08	3.963	-1.128	MHK
LEU	ICPMS	BE	3/4/08	3.994	-0.355	MHK
LEU	ICPMS	BE	3/4/08	4.015	0.169	MHK
LEU	ICPMS	BE	3/4/08	4.01	0.044	MHK
LEU	ICPMS	BE	3/4/08	4.002	-0.155	MHK
LEU	ICPMS	BE	3/4/08	4.032	0.593	MHK
LEU	ICPMS	BE	3/4/08	4.017	0.219	MHK
LEU	ICPMS	BE	3/4/08	4.007	-0.031	MHK
LEU	ICPMS	BE	3/4/08	4.005	-0.081	MHK
LEU	ICPMS	BE	3/4/08	3.975	-0.829	MHK
LEU	ICPMS	BE	3/4/08	4.027	0.468	MHK
LEU	ICPMS	EA	3/28/07	1.292	0.008	
LEU	ICPMS	EA	3/28/07	1.292	0.008	
LEU	ICPMS	EA	3/29/07	1.293	0.085	
LEU	ICPMS	EA	3/29/07	1.293	0.085	
LEU	ICPMS	EA	3/28/07	1.293	0.085	
LEU	ICPMS	EA	3/28/07	1.293	0.085	
LEU	ICPMS	EA	3/29/07	1.293	0.085	
LEU	ICPMS	EA	3/29/07	1.293	0.085	
LEU	ICPMS	EA	3/28/07	2.986	0.057	
LEU	ICPMS	EA	3/28/07	2.986	0.057	
LEU	ICPMS	EA	3/29/07	2.986	0.057	
LEU	ICPMS	EA	3/29/07	2.985	0.023	
LEU	ICPMS	EA	3/31/07	2.984	-0.010	
LEU	ICPMS	EA	3/31/07	2.984	-0.010	
LEU	ICPMS	EA	3/29/07	2.985	0.023	
LEU	ICPMS	EA	3/29/07	2.985	0.023	
LEU	ICPMS	EA	6/14/07	4.7874	-0.104	NP2
LEU	ICPMS	EA	6/14/07	4.788	-0.092	NP2
LEU	ICPMS	EA	6/21/07	4.7895	-0.061	NP1
LEU	ICPMS	EA	6/21/07	4.7893	-0.065	NP1

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	ICPMS	EA	6/14/07	4.7876	-0.100	NP2
LEU	ICPMS	EA	6/14/07	4.792	-0.008	NP2
LEU	ICPMS	EA	6/21/07	4.79	-0.050	NP1
LEU	ICPMS	EA	6/21/07	4.7892	-0.067	NP1
LEU	ICPMS	EA	3/31/07	4.39	-0.035	
LEU	ICPMS	EA	3/31/07	4.391	-0.012	
LEU	ICPMS	EA	3/29/07	4.392	0.010	
LEU	ICPMS	EA	3/29/07	4.392	0.010	
LEU	ICPMS	EA	3/31/07	4.39	-0.035	
LEU	ICPMS	EA	3/31/07	4.39	-0.035	
LEU	ICPMS	EA	3/29/07	4.392	0.010	
LEU	ICPMS	EA	3/29/07	4.392	0.010	

<u>Material</u>	<u>Method Type</u>	<u>Facility</u>	<u>Analysis</u>	<u>Reported</u>		
			<u>Date</u>	<u>Result</u>	<u>% RD</u>	<u>Analyst</u>
LEU	GSMS	BC	4/12/06	3.1885	-0.011	ET
LEU	GSMS	BC	4/12/06	3.1911	0.071	ET
LEU	GSMS	BC	4/12/06	3.1911	0.071	ET
LEU	GSMS	BC	4/12/06	3.1929	0.127	ET
LEU	GSMS	BC	4/13/06	3.1913	0.077	ET
LEU	GSMS	BC	4/13/06	3.1862	-0.083	ET
LEU	GSMS	BC	4/13/06	3.1943	0.171	ET
LEU	GSMS	BC	4/13/06	3.1991	0.322	ET
LEU	GSMS	BC	10/24/06	3.1988	0.312	ET
LEU	GSMS	BC	10/24/06	3.2008	0.375	ET
LEU	GSMS	BC	10/24/06	3.1883	-0.017	ET
LEU	GSMS	BC	10/27/06	3.1816	-0.227	ET
LEU	GSMS	BC	10/27/06	3.1849	-0.123	ET
LEU	GSMS	BC	10/27/06	3.1888	-0.001	ET
LEU	GSMS	BC	10/24/06	3.1984	0.300	ET
LEU	GSMS	BC	10/24/06	3.1978	0.281	ET
LEU	GSMS	BC	10/24/06	3.1886	-0.007	ET
LEU	GSMS	BC	10/27/06	3.1873	-0.048	ET
LEU	GSMS	BC	10/27/06	3.1807	-0.255	ET
LEU	GSMS	BC	10/27/06	3.1765	-0.387	ET

Appendix C: NBL Measurement Evaluation Program: New Database Development

Srinivasan, B¹, Soriano, Michael¹, Waggoner, Joseph W.¹, and Kuziel, R.²
 New Brunswick Laboratory, Argonne, Illinois 60439,¹ and
 CNI Information Technology LLC, Argonne, Illinois 60439.²

Abstract

The New Brunswick Laboratory has been tasked by the United States Department of Energy (DOE) to administer an “inter-laboratory” control program to provide independent verification of internal quality control in nuclear materials measurements. DOE facilities performing material control and accountability measurements for nuclear material safeguards participate in the program. This program is open to international laboratories also. The program provides periodic evaluations of results from quantitative measurements of uranium and plutonium bearing materials. Both destructive and non-destructive measurement results are evaluated. One major criterion used in this evaluation is a comparison of measurement bias and precision against the respective international target values (ITVs). Large deviations from ITVs are a matter of concern as they indicate the difficulty and/or inability to detect material loss should they occur in processing, or by theft or by diversion of material. NBL maintains a permanent record of the measurement evaluation results and reports. These archives are ideal sources for analyzing long term trends in analytical performance and also an aid in setting the international target values.

In the inter-laboratory program (also known as the measurement evaluation program), the test samples measurement results are evaluated against the respective characterized values using a database application program internally developed several years ago. This is being replaced by a new database program with certain enhancements and conforming to quality assurance requirements. The new program is a web-based system utilizing Java whose interface is platform independent. Data is stored in an SQL database server providing reliable retrieval, backup and restoration of data, and with multiple layers of security. The system stores data on assay and isotopic abundance measurements, performs outlier tests, compares the measurement results against the respective characterized values, calculates bias and precision of the measurements using statistical methods, and prepares evaluation reports and graphs. The new system when it becomes fully operational will require direct input of measurement results by the participants and will also permit direct electronic retrieval of reports. The main features of the new database program will be presented.

A. Introduction

The United States Department of Energy (DOE) has charged the New Brunswick Laboratory (NBL) with the responsibility to administer an “inter-laboratory” control program to provide independent verification of internal quality control in nuclear materials measurements. DOE facilities performing accountability measurements of uranium and plutonium bearing materials for nuclear material safeguards are required to participate in the program; non-DOE facilities (e.g., U.S. Nuclear Regulatory Commission facilities and international laboratories) also participate, but on a voluntary basis. The program is divided into two parts: Safeguards Measurement Evaluation (SME) Program and the Calorimetric Exchange Program (CALEX). In the SME program, NBL evaluates measurement results from destructive methods of analyses; in the CALEX program non-destructive analysis results are evaluated.

B. SME program

The Safeguards Measurement Evaluation program was established in 1986 as a successor to the Safeguards Analytical Laboratory Evaluation (SALE) and General Analytical Evaluation (GAE) programs. In the SME program, NBL sends uranium and plutonium test samples for elemental and isotopic-abundance analyses, and the participants analyze the samples by methods routinely used by them in accountability measurements (e.g., Davies-Gray titration, XRF, IDMS and TIMS). NBL evaluates the results with respect to “characterized” reference values. The accuracy and precision in the measurements are compared against international target values for bias and precision. Measurement evaluation reports are sent to the laboratories and to the oversight offices.

C. CALEX program

The CALEX program was originally administered by the Mound Laboratory, and it was transferred to NBL in the mid 1990's. In this program, using calorimetry and gamma ray spectrometry methods, participants measure power (i.e., heat output) and isotopic abundance measurements of two different working reference material standards, known as Calex I and Calex II. The standards contain well defined quantities of plutonium oxide of known isotopic compositions. NBL evaluates the measurement results for accuracy and precision. In addition, effective specific power and plutonium mass, quantities derived from the measurement results, are evaluated.

D. Microsoft FoxPro database and applications

Until recently, the measurement results submitted to the SME and CALEX programs were entered into a FoxPro database, and evaluated using statistical methods. The FoxPro based applications used in the evaluation were developed mainly through the efforts of a NBL physicist. The development work started with the evaluation of uranium assay and uranium isotope abundance results from destructive analysis using an application named as UEx (for uranium exchange). The apparent success of the UEx system led to the development of PuEx (plutonium exchange) for plutonium results from destructive analysis and to CALEX for calorimetry/gamma spectrometry results. The three applications are quite similar, but separate. The SME and CALEX programs relied on these three applications for generating statistical evaluation reports and also the annual reports. The physicist who developed them maintained the system, updated the applications, and in general provided satisfactory service. However, with his departure from NBL in 2003, the responsibility to maintain the three applications was transferred to a NBL statistician who possessed limited knowledge of the FoxPro database and the applications. In collaboration with contracted computer professionals under the auspices of the Office of Science - Chicago Office (SC-CH,) the FoxPro system was maintained in operational status, but several drawbacks in the program design were found; a lack of documentation the absence of quality assurance, subsequent modifications, and outdated and unsupported software exasperated the issues. In early 2006, NBL management recommended development of a new database and application program based on modern programming techniques to replace the outdated FoxPro based system because of extraordinary efforts required to maintain the latter. The new system (Safeguards Measurement Evaluation System or SMES) has been under development since then and will be maintained by computer professionals fully cognizant of the need to maintain system documentation and quality assurance measures. The SMES will take advantage of the advances in "web-technology" to facilitate user interface.

E. SMES: Design features

NBL personnel worked closely with computer professionals of CNI Information Technology LLC, to draw up system and design specifications for the new SMES. The new system is being developed with efficient and reliable data collection capabilities, and to provide timely and accurate analysis of measurement results and reports. Specific features are described below.

- **User self-service:** One of the main objectives of the SMES application is to provide access to participants to enter their own data, validate the entries, and retrieve the evaluation reports. Figures 1 and 2 below are examples of the user self-service report distribution and data entry screens of SMES, respectively. The FoxPro application does not allow direct data entry; NBL personnel entered data sent by the participants via e-mail/postal services, validated the data entry and mailed performance evaluation reports back to participants.
- **Security through the Web:** SMES will provide secure access through the Internet by using User ID/passwords, role-based access and encryption with considerations of confidentiality of data submitted and reports generated. The FoxPro application was only available on the NBL local area network and did not permit remote user access.

- **Quality assurance/documentation:** The SMES application is being developed and tested, and validated by computer professionals (SC-CH contractor) using modern development techniques. The DOE standards for software development, change control and quality assurance are being followed. The FoxPro application was not designed with these standards in mind.
- **Modular programming:** The three separate FoxPro applications (UEx, PuEx and CALEX) are combined into a single application in the SMES. The SMES employs modular programming techniques and reusable code, and will be easy to maintain. The design allows for future expansions of the measurement evaluation programs such as evaluation of measurement results from new techniques. The FoxPro applications were designed with significant code duplication. It meant corrections made in one location in an application required duplication of corrections in the other applications followed by checks and cross-checks to ensure proper working of all applications – a time consuming and inefficient process.

Reports								
	Fiscal Year	Lab	Report Name	Type	Status	Create Date	ID	
	1996	NBL	NBL59UNH	U ASSAY	report published		5	
	1996	NBL	NBL5CUO3	U ASSAY	report published		6	
	1996	NBL	nbl 12 uf6	U ASSAY	report published		10	
	1996	NBL	NBL Jan96	U ASSAY	report published		16	
	1996	NBL	NBL MAR96	U ASSAY	report published		22	
	1996	NBL	NBL APRIL 1996	U ASSAY	report published		27	
	1996	NBL	NBL May 96	U ASSAY	report published		35	
	1996	NBL	NBL JUNE96	U ASSAY	report published		47	
	1996	NBL	NBL JULY96	U ASSAY	report published		48	

Page 1 of 1

[Main Menu](#)

Figure 1

- **Modern technology** – The SMES employs modern web-based technologies and uses a well supported modern database system (SQL). The FoxPro database is no longer well supported and lacked the capability needed to migrate to a robust web-based application.

Uranium Exchange Assay Data Review

Report:	NBL Assay UNH Nov 05 DG		
Facility:	NBL		
Measurement Type:	U ASSAY	Material Type:	UNH Solution
Report Year:	2006	Quarter:	-Select Quarter-
Status:	Open	Report No.	718
Sample:	<input type="button" value="Edit Sample"/> <input type="button" value="Print"/>		
Measurement Method:	Davies-Gray Titration	Measured As:	g/g Percent
Notes:	<div style="border: 1px solid black; height: 40px; width: 100%;"></div>		
<input type="button" value="OK"/> <input type="button" value="Save"/> <input type="button" value="Cancel"/>			

Reported Values

	Analysis Date	Analyst	Sample Number	Reported Value	Out	D Out	ID
	11/22/2005	231	98NU0074 - 98NU0074-041	1.0043	F	F	5347
	11/22/2005	231	98NU0074 - 98NU0074-041	1.0049	F	F	5348
	11/22/2005	231	98NU0076 - 98NU0076-034	1.0012	F	F	5355
	11/22/2005	231	98NU0076 - 98NU0076-034	1.0026	F	F	5356
	11/21/2005	231	98NU0076 - 98NU0076-034	0.9999	F	F	5353
	11/21/2005	231	98NU0076 - 98NU0076-034	0.9998	F	F	5354
	11/21/2005	231	98NU0074 - 98NU0074-041	1.0043	F	F	5345
	11/21/2005	231	98NU0074 - 98NU0074-041	1.0044	F	F	5346
	11/18/2005	237	98NU0076 - 98NU0076-034	1.0000	F	F	5351
	11/18/2005	237	98NU0076 - 98NU0076-034	1.0000	F	F	5352
	11/18/2005	237	98NU0074 - 98NU0074-041	1.0041	F	F	5343
	11/18/2005	237	98NU0074 - 98NU0074-041	1.0040	F	F	5344
	11/17/2005	237	98NU0074 - 98NU0074-041	1.0032	F	F	5341
	11/17/2005	237	98NU0074 - 98NU0074-041	1.0034	F	F	5342
	11/17/2005	237	98NU0076 - 98NU0076-034	0.9999	F	F	5349
	11/17/2005	237	98NU0076 - 98NU0076-034	0.9995	F	F	5350

Figure 2

F. SMES Specifications

The SMES leverages the calculation techniques presently found in the FoxPro application. All data presently in the FoxPro application will be migrated into the new system so as to retain historical data. The specifications for SMES are as follows.

- **Java enterprise server architecture:** SMES is designed around the Java 2 Enterprise Edition (J2EE) architecture. The system uses a dedicated J2EE Application Server.
- **SQL database storage:** SMES data is securely stored on an SQL database server for quick retrieval and updating of data; data backups are automated.
- **Secure platform independent thin client:** Laboratories will be able to enter their own measurement results to SMES via the Internet, eliminating the need to mail the data to NBL. SMES will support the most popular current web browsers with Secure Socket Layers (SSL) and will require no special browser add-ons.

- **Role-based security:** The system provides a number of access roles including those required for data entry, data validation and published report retrieval. SMES will provide access to participating laboratories and oversight agencies (e.g., DOE area office). Note that participant laboratories will have access only to their own data and reports.
- **Historical Data:** All historical data contained in the FoxPro database will be migrated to SMES.
- **Calculation Techniques:** The time proven statistical analysis tools (e.g., outlier tests, calculation of mean and standard deviation of %RDs, tests to determine day-to-day and analyst-to-analyst variations, determination of 95% C.I. etc.) originally written for the FoxPro application will be retained.

G. SMES Development Timeline

The SMES is expected to become operational by the end of the third quarter of CY 2007 with web-based features becoming functional by the end of the year. Access to the SMES system by international laboratories will require additional security measures that are under discussion.

H. Conclusions

The new Safeguards Measurement Evaluation System (SMES) system replaces three FoxPro programs. It is developed and maintained by computer professionals, is web-based, and will provide users with access to the Internet to enter data and retrieve reports. SMES is designed with full system documentation, quality assurance, security and confidentiality in mind.

Acknowledgements

The authors acknowledge the help given by Richard Palczynski, IT Project Manager at SC-CH, and the computer professionals Kevin Atto, Joel Catausan, Miguelito Domingo and Mark Jilek (all of CNI Information Technology,LLC.), in developing the new SMES. And special thanks go to Jon Neuhoff, NBL Director, for financial support and leadership in making this new system possible.

**Appendix D: New Brunswick Laboratory Safeguards Measurement Evaluation Program:
Operational Features**

New Brunswick Laboratory Safeguards Measurement Evaluation Program: Operational Features

**B. Srinivasan, K. J. Mathew, P. Croatto, U. Narayanan &
J. Neuhoff**

*U. S. Department of Energy, New Brunswick Laboratory, Argonne, Illinois 60439,
USA*

Corresponding author: B.Srinivasan@ch.doe.gov

Abstract: The Material Control and Accountability plan of nuclear facilities must specify how nuclear material holdings will be accounted for and controlled. This plan includes elements that are designed to (i) deter and detect loss, theft, and diversion of nuclear materials and; (ii) ensure that nuclear material are in their authorized locations and are being used for their intended purposes. In addition, the U.S. Department of Energy facilities performing nuclear material measurements participate in the New Brunswick Laboratory Measurement Evaluation Program to monitor performance and to provide an independent check of the quality control measures used in accountability measurements. In this paper, the operational features of this Measurement Evaluation Program designed to support nuclear safeguards and nuclear non-proliferation programs are described briefly.

Keywords: Nuclear Materials; Material Control and Accountability Measurements; Nuclear Safeguards; Non-Proliferation.

Reference to this paper should be made as follows: Srinivasan, B. et al, (2007) 'Nuclear material sample measurement comparison and performance evaluation in support of non-proliferation programs', *Int. J. Nuclear Knowledge Management*, Vol. 1, Nos. 1/2/3, pp.43–54.

Biographical notes: B. Srinivasan received his PhD in Chemistry from the University of Missouri in 1971. He is currently coordinator of the Measurement Evaluation Program at the New Brunswick Laboratory, Argonne, IL, USA. His current research interest topics include analytical chemistry of actinide elements and measurement techniques development efforts at the New Brunswick Laboratory.

K. J. Mathew is a Physical Scientist in the Standards and Evaluation Division of New Brunswick Laboratory, Argonne, IL, USA. He received his PhD in Physics from Physical Research Laboratory, Ahmedabad, India in 1993. His current research interests include Thermal Ionization Mass Spectrometry, Isotopic composition of U and Pu, and actinide measurement method development.

P. Croatto is a Physical Scientist in the Standards and Evaluation Division of New Brunswick Laboratory, Argonne, IL, USA. He received his PhD in Chemistry from the University of Illinois at Chicago in 1990. His current research interests include uranium hexafluoride mass spectrometry, alpha spectrometry, and development of methods for actinide measurement.

U. Narayanan received her PhD in Physical Chemistry from the University of Illinois, Chicago in 1987. She is currently the Director of the Standards and Evaluation Division of

New Brunswick Laboratory, Argonne, IL, USA. Her current research interests include uranium, neptunium, plutonium, and americium measurement method development, certification of reference materials, and sample measurements for safeguards.

J. Neuhoff is the Director of New Brunswick Laboratory, Argonne, IL, USA. He received his M.S. Degree in Nuclear Engineering from University of Illinois at Urbana-Champaign. His research interests are nuclear material packaging and transportation, Safeguards, Non-Proliferation, and nuclear material measurement methods.

1 INTRODUCTION

The New Brunswick Laboratory (NBL) serves as the U.S. government's central authority for nuclear material measurements and measurement evaluation, and is the U.S. government's certifying authority for nuclear reference materials. The major mission of the NBL is to provide technical assistance to the Department of Energy (DOE) in the following areas: administering the measurement evaluation program, preparation of certified (nuclear) reference materials, measurement techniques development, and measurement services. In addition, NBL provides help to DOE in three other areas: conducting technical audits, resolving shipper/receiver differences in material transfers, and assisting in nuclear nonproliferation programs. This paper describes the operational aspects of the NBL measurement evaluation program designed to monitor performance in nuclear safeguards measurements. At the discretion of the DOE, this program is open to participation by non-DOE facilities in the U.S. and abroad.

2 MEASUREMENT EVALUATION PROGRAM

Accountability measurements of nuclear materials, mainly of uranium and plutonium, are done by either destructive or non-destructive methods. The methods must be capable of providing quantitative results within acceptable limits of accuracy and precision. Large bias and/or poor precision compromise the ability to detect material loss in processing or by theft or by diversion.

The measurements must be made compatible with the national and/or international measurement reference base. In other words, the measurement results remain traceable to higher-level standards. In the U.S., these (reference material) standards are made by NBL and/or by NIST (National Institute of Standards and Technology). In addition, the safeguards measurement laboratories are required to participate in the NBL Measurement Evaluation Program. The program provides an independent verification of the internal analytical quality control practices in nuclear material accountability measurements. Taken together, measurement results traceability and measurement results evaluation form the basis for maintaining a sound and reliable nuclear safeguards material accountability program. In particular, through participation in the performance evaluation program the participants receive timely and periodic evaluations to identify deficiencies in measurement techniques and practices and correct those deficiencies.

The NBL Measurement Evaluation Program has two parts: the Safeguards Measurement Evaluation (SME) program for destructive analyses measurement results and the Calorimetry Exchange (CALEX) program for non-destructive analyses measurement results. This paper describes the essential features of the SME program.

3 SME PROGRAM

The SME program is designed to evaluate the accuracy and precision achieved in routine nuclear material accountability measurements. The participating facilities receive well characterized test samples that are closely matched to their routine measurement needs. For example, laboratories involved with uranium fuel cycle material analyses will receive test samples of that type (e.g., UO_2 pellets with specific level of ^{235}U enrichment). The laboratories analyze the test samples as part of their routine measurements and submit the results to NBL for statistical evaluation. Usually, the laboratories participate in the program on a quarterly basis, performing an adequate number of test samples analyses to permit statistical evaluation of bias, precision, day-to-day variation and/or analyst-to-analyst variation in the measurement results. The bias and precision of each set of measurement results are compared against the method/material specific International Target Values (ITVs)¹. Results conforming to ITVs signify satisfactory level of performance; non-conformity requires review of laboratory practices, experimental techniques, technicians' training etc. The participant laboratories receive copies of the statistical evaluation reports along with a cover letter commenting upon the performance quality and any need for improvement.

3.1 TEST SAMPLES

The test samples used in the SME Program are made by NBL from Certified Reference Materials (CRMs) or Working Reference Materials (WRMs), or custom-made. The test samples are also characterized (for assay and isotopic abundance) at NBL before sending them to participants. Table I shows a list of test samples that are available at this time. New test samples will be added to this list to reflect changing needs.

MEASUREMENT EVALUATION: IN SUPPORT OF NON-PROLIFERATION

3

Table 1. Materials and Methods evaluated in SME Program

Material	Typical Assay Method	Typical Isotopic Method
UNH Solutions	D&G, HPT, XRFL, IDMS	TIMS, ICPMS
UO2	D&G	ICPMS, TIMS
UF6	D&G	GSMS
UO3	D&G, IDMS, XRFS	
LEU		TIMS, ICPMS, GSMS
HEU		TIMS
DU		TIMS
NU		TIMS
Plutonium Sulfate (Low burn-up)	IDMS	TIMS
Plutonium Sulfate (High burn-up)	IDMS	TIMS

UNH Uranyl Nitrate Solution

UO2 Uranium Dioxide Pellet

UF6 Uranium Hexafluoride

UO3 Uranium Trioxide Powder

HEU High Enriched Uranium

LEU Low Enriched Uranium

D&G Davies-Gray Titration

HPT High Precision Titration

IDMS Isotope Dilution Mass Spectrometry

XRFL X-Ray Fluorescence - Liquid

XRFS X-Ray Fluorescence - Solid

TIMS Thermal Ionization Mass Spectrometry

ICPMS Inductively Coupled Plasma Mass Spectrometry

GSMS Gas Source Mass Spectrometry

4 STATISTICAL EVALUATION METHODS AND EVALUATION REPORTS

The measurement results are evaluated using statistical methods. First, the percent relative difference (% RD) of each experimental result is calculated with respect to the corresponding reference value, the latter obtained from characterization measurements. The % RD is defined as follows:

$$\% \text{ RD} = 100 \times \frac{\{\text{observed value} - \text{reference value}\}}{\text{reference value}}$$

Next, each set of % RDs is examined for outliers using a number of statistical tests. A particular result is identified as a potential outlying value if at least two of the statistical tests show it to be an "outlier" at $\geq 99\%$ significance level. The data set, with outliers removed, is then tested to identify significant sources of variation (e.g., day-to-day and/or analyst-to-analyst) using standard one-factor analysis of variance (ANOVA). If the ANOVA results indicate no significant variation, then the standard uncertainty is the simple standard deviation (σ) of the results divided by the square root of n , where n is the number of measurements. The coverage factor is the Student's 95% "t" factor with $n-1$ degrees of freedom. For example, in a set of 8 results showing no day-to-day or analyst-to-analyst variation, the number of degrees of freedom is 7, and the coverage factor is 2.36.

If the ANOVA results indicate significant day-to-day and/or analyst-to-analyst variation ($\geq 95\%$), then the standard uncertainty in the mean % RD is estimated from a combination of the mean square for the "error" and the mean square for the "model" quantities from the ANOVA, with degrees of freedom determined from Satterthwaite's approximation. For measurements done on two different

days (or by two different analysts), the formula for estimating the standard uncertainty in the mean % RD is reduced to the square root of the mean square for the "model" quantity obtained from ANOVA results. In this case, the coverage factor is 12.71 (i.e., the Student's 95% "t" factor with one degree of freedom).

A typical example of the statistical report for uranium assay by Davies-Gray titration is shown in Table 2; the same results are also displayed in Fig. 1.


Table 2: Sample Data Evaluation Report (No statistically significant difference due to analysis day)

Day to Day ANOVA analysis					
Report for Laboratory: XX					
UO2 Pellet – U Concentration					
Davies-Gray Titration					
Date of Report:	November 30, 2003	Sample Number	Aliquant Number	Analysis Date	Reported %U
95EU0079-1	1	11/03/03		88.126	-0.0934 XXX
95EU0079-1	2	11/03/03		87.990	-0.1577 XXX
95EU0079-2	1	11/03/03		88.031	-0.1112 XXX
95EU0079-2	2	11/03/03		87.892	-0.2689 XXX
95EU0079-1	3	11/04/03		88.030	-0.1123 XXX
95EU0079-1	4	11/04/03		87.950	-0.2031 XXX
95EU0079-2	3	11/04/03		87.922	-0.2349 XXX
95EU0079-2	4	11/04/03		88.002	-0.1441 XXX

Number of Results Analyzed	8
Mean % Difference	-0.154
Mean Absolute % Difference	0.154
95% C.L. of Mean (df = 7)	0.070
Standard Deviation	0.083
Between-Day Standard Deviation (df = 1)	0.054
Within-Day Standard Deviation (df = 6)	0.087
Statistical Significance of Between-Day Standard Deviation	44.3%

International target value for bias in Davies-Gray Titration is 0.1%.
International target value for precision in Davies-Gray Titration is 0.1%.

Figure 1: Sample data Evaluation Report (Table 2 results are displayed).



The uncertainties shown in the Table 2 are the 95% confidence limit (C.L.) of means. In the figure, the 95% confidence interval (C.I.) of the mean is constructed from the C.L. Note that the C.I. represents the interval containing all values between the mean % RD minus the C.L. and the mean % RD plus the C.L. Thus, the 95% C.L. of the mean are just the two end points of the C.I.

A measurement is considered to be bias-free if the 95% C.I. included zero. The Table 2 results show that the measurements suffer from a negative bias (-0.154 ± 0.070). Also the result fails to meet the bias ITV of 0.1%.

The Table 2 results show that the standard deviation (σ) is 0.083 which represents the precision of the measurement results. This is within the precision ITV of 0.1.

5 SUMMARY AND CONCLUSIONS

The New Brunswick Laboratory Safeguards Measurement Evaluation Program provides performance evaluation service to the DOE laboratories making nuclear material accountability measurements. The program, at the discretion of the U.S. DOE, is open to non-DOE laboratories in the U.S. and abroad.

The participants derive the following benefits through participation in the SME program:

- a) Performance evaluation of nuclear safeguards accountability measurement results of uranium (e.g., fuel cycle materials) and plutonium (e.g., low and high burn-up material).
- b) Verification of internal quality control practices.
- c) Assistance to correct deficiencies in measurement practices and techniques
- d) Assistance in training analysts in measurement techniques
- e) Assistance in qualifying new methods of analyses, and in qualifying analysts
- f) Resolution of shipper-receiver differences in nuclear commerce
- g) Long-term maintenance of performance evaluation records
- h) Availability of a wide variety of test samples pertinent to nuclear safeguards and non-proliferation applications.

In addition to the above, NBL organizes an annual meeting of the Measurement Evaluation Program in conjunction with the annual meeting of the Institute of Nuclear Materials and Management (INMM). The participation at this meeting is open to all participants and potential participants in the SME program. The meeting provides opportunities to present papers on nuclear material measurement techniques. It also provides an open forum to discuss issues and concerns in the operation of nuclear safeguards and nuclear nonproliferation programs.

REFERENCES

-
- Aigner H., Binner R., Kuhn E. et al. International Target Values 2000 for Measurement Uncertainties in Safeguarding Nuclear Materials, International Atomic Energy Agency Report STR-327.

[This page intentionally left blank.]